



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 196454

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : 24x7 PHC Keezhmad	Call Registration Date : 09-02-2024
Address : Ernakulam	Caller ID : 95199
Ernakulam (dt)	Date of Visit : 12-02-2024
Ph : 09497324261	Asset No. : 0743851
	EQPT Name : Nebuliser
	Manufacture Lifecare Model : NA
	S. No. NA Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Nebuliser not working

Action Taken : Checked the machine and found motor and piston faulty.

Completed Date : 12-02-24 Time : 10:30AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Motor	1 No.		
2. Piston	1 No.		
3.			

Cyrix Engineer	Date	Start Time	End Time
Jibin Thomas George	12-02-24	9:30 AM	10:30 AM

Customer Remark Completed Pending

Service Engineer Name : Jibin Thomas
Signature : *Jibin Thomas*
Date : 12-02-24
Contact Number : 9047958051

Customer Name : *Janhana*
Signature : *Janhana*
Date : 12-02-24
Contact Number : *9047958051*
Designation : Medical Officer Incharge
Hospital Seal : Family Health Centre Keezhmad
Aluva - 685 442




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	24X7 PHC KEEZHMAAD
3	Name of Equipment with Make, Model and Serial Number	Equipment: NEBULISER Make: LIFECARE Model: NA Sr.no- NA
4	Equipment ID & Barcode	95199 & 0743851
5	Date of purchase / Year of manufacture /Installation Date	23-02-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	09-02-24 (TOLL-FREE)
9	Action taken	Checked the machine and found motor and piston faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	It is not recommend for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs.1353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found motor and piston faulty. The equipment outlived 6+ years and installed on 23-2-2018. Since the spares are not available in the local market, recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Devally
24/4/24
Signature of JC BM (NHM)

CONSULTANT
Bio-Medical Engg
National Health Mission (NHM)
Aroyyakeralam Ernakulam

Man Man



Superintendent / Medical Officer (i/c)

Signature of seal

Date

Medical Officer Incharge
Family Health Centre Keezhmad
Erumathala P.O.,
Aluva - 683 112

from old book
 56
 page no: 101


STOCK REGISTER

Date	Particulars	V. No. Bill. No.	RECEIPT		ISSUED		BALANCE	
			Quantity		Quantity			
23/2/18	Kd from store (Paediatric)							
12/6/19	(10) (10cm)		1				2	2
9/10/20	"		1				3	3
1/6/2021	Red from the store (ARDRAM)		1				4	4
1/6/21	Issued to C.R.L.T.C				2 Nos		5	5
16/10/21	Returned from Doc		4				7	7
22/1/24	Transferred to condemnation Register Volume 2 page no 109.				2		5	5



Handwritten signature

Medical Officer Incharge
 Family Health Centre Keezhmad
 Erumthala P.O.,
 Aluva - 683 112



Sham Sham

Medical Officer Incharge
Family Health Centre Keezhmad
Erumathala P.O.,
Aluva - 683 112

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0743851



NEBULIZER
ON COMPRESSOR
Operating Time: 30 min ON - 30 min OFF
TECHNOLOGY, GERMANY



Handwritten signature

Medical Officer Incharge
Family Health Centre Keezhmad
Erumathala P.O.,
Aluva - 683 112