



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 183230

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THQH	Call Registration Date : 15/2/2024
Address : Chertthala	Caller ID : 96289
Alappuzha	Date of Visit : 16/2/2024
Ph : 9846553583	Asset No. : 0420340
	EQPT Name : ECG machine
	Manufacture BPL Model : 610BT
	S. No. AVMPAB23194 Dept. ECG Dept

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : ECG machine is not getting ON.

Action Taken : checked the machine and found that ECG main board assembly is complaint. Enquired the OEM for quotation.

Completed Date : 6/3/24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Nasret		10:00am	12:00pm

Customer Remark : Not working.
 Completed Pending
 Nishe.K ECG Technician

Service Engineer Name : Nasret
 Signature : *Nasret*
 Date : 6/3/24
 Contact Number : 9526744524

Customer Name : *[Signature]*
 Signature : *[Signature]*
 Date : *[Signature]*
 Contact Number : *[Signature]*
 Designation : *[Signature]*
 Hospital Seal : *[Signature]*



SUPERINTENDENT
KALUK HEAD QUARTERS HOSPITAL
CHERTHALA

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ALAPPUZHA
2	Name of Hospital	THQH CHERTHALA
3	Name of Equipment with Make, Model and Serial Number	ECG MACHINE SINGLE CHANNEL MAKE :- BPL MODEL :- CARDIART 6108T SN :- AVMP4B23194
4	Equipment ID & Barcode	96289 & 0420340
5	Date of purchase / Year of manufacture /Installation Date	19/04/2014
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15/02/2024
9	Action taken	Checked the machine and found that ECG Mainboard assembly is complaint. Enquired the OEM for Quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not Recommending For Repair.
12	Cost of spares (specify parts and cost)	RS:- 17500/-

13	Asset Value	RS:- 24622.50/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Percentage Value Of The Cost Of Spare:- 71%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service Report & OEM Quotation
16	Reasons for recommending the equipment as BER	Checked the machine and found that ECG Mainboard assembly is complaint. Enquired the OEM for Quotation. The equipment installed on 19/04/2014, aged up to 9 years and 10 months. Percentage of service cost is high (71%) and economically not viable. Both criteria met as per the tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NASEEF K N <i>Naseef</i>

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER

[Signature] 11/08/24
JYOTHISH THOMAS
 Consultant (Biomedical Engineering)
 NHM ALAPPUZHA
 Signature of JC BM (NHM)



Date

Seal

For condemnation

[Signature] 6/3
 Signature of
 Superintendent / Medical Officer (i/c)

SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA

STOCK REGISTER

Name: BPL ECG MODEL 6108T (SINGLE CHANNEL ECG MACHINE) 13

Date	Received from or issued to	Inv. No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
22.09.13	Inv. No: EKM/FG/202/ 15-14 dt 23.08.13 Supplied by BPL Ltd, Cochin		22227	50	1		1	4
	KMSCL order No: 680/2013/37 dt 05.08.13	VAT	1113	88				
			23391	38				
22.09.13	Issued to ECG Dept (Page No: 29)					1	NIL	
27/3/14	Inventory: EKM/FG/40459 dt 22.3.14 BPL Pvt Ltd Cochin.		23450	7				
			1122	50				
			24622	50			1	
19/10/14	Issued to ECG Dept					1	NIL	
								PRADEEP.V.N ECG Dept

Received
Pradeep V.N
22/09/13





Received
Pradeep V.N
19/10/2014

for
Pradeep V.N

6/3



SUPERINTENDENT
BALUK HEAD QUARTERS HOSPITAL
CHERTHALA



For
6/3

SUPERINTENDING
HEALTH SERVICES
CHERTHALA



Abhilash A <abhilashacyrix@gmail.com>

PRICE ENQUIRY :-"0420340" - TICKET NO 96289 - ECG MACHINE

2 messages

Abhilash A <abhilashacyrix@gmail.com>

Sat, Mar 2, 2024 at 12:38 PM

To: BPL Medical Equipment Service <biomedixservice@gmail.com>, BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>

Dear Sir,

I would like to inform you that please share the price and availability of the Mainboard for the ECG machine model BPL Cardiart 6108 T(SN:-AVMP4B23194) and also update the expected date of delivery after issuing PO. I am expecting a favourable response from you soon.

With Regards,**Abhilash A****SENIOR BIOMEDICAL ENGINEER | BEMP-KERALA****Mob: +91 8157068890 , 7593847172****Email: abhilashacyrix@gmail.com | Website:- www.cyrixhealthcare.com****Cyrix Healthcare Pvt Ltd**

30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

*" Genius in Healthcare "***BIOMEDIX SOLUTIONS <biomedixsales@gmail.com>**

Sat, Mar 2, 2024 at 5:17 PM

To: Abhilash A <abhilashacyrix@gmail.com>

Dear sir,

Please find the attached document

Thanks & Regards

Biomedix Solutions

Kottayam

[Quoted text hidden]

**PROFORMA INVOICE 30 AJITH CYRIX.pdf**

313K

BISO/KTM/QTN/030/02042024

02-03-2024

QUOTATION

To,

Cyrix Health Care Pvt LTD
Ernakulam

Dear Sir,

Please find below our lowest quotation for the Following items

Sl no	Model description	Qty	Unit Price
01	MAINBOARD WITH MICON PCB BPL 6108T	1	17500+TAX

Terms and conditions

1. Payment- 100% advance
2. Delivery 15 days

Thanking you,

For Biomedix Solutions

Authorized Signatory

Mobile: 9846969564 | Email: biomedixservice@gmail.com



BIOMEDIX SOLUTIONS

No: 7/437- B, Mannarakayam P.O,
Kathilankalpadi, Kanjirapally, Kottayam - 686506,
Mob:9846969564,

Branches: Kozhikode, Kannur, Kottarakkara-Kollam.

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



for P.H.



Happier Living Everyday

6/3

**SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
CHERTHALA**