



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 200629

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>CHC</u></p> <p>Address : <u>CHC Vengoor</u> <u>Ernakulam, Kerala</u></p> <p>Ph : <u>9633090531</u></p>	<p>Call Registration Date : <u>20/2/2024</u></p> <p>Caller ID : <u>97158</u></p> <p>Date of Visit : <u>22/2/2024</u></p> <p>Asset No. : <u>0732604</u></p> <p>EQPT Name : <u>Spot light</u></p> <p>Manufacture : <u>Nil</u> Model : <u>Nil</u></p> <p>S. No. : <u>Nil</u> Dept. : <u>Minor OT</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken :
Checked and spot light circuit fully damaged identified. Base and Body fully rusted. wheels damaged. Reflector, illuminator broken. Lamp and wiring harness fully in damaged condition.

Completed Date : 22/2/24 Time : 10:30AM Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1. Lamp	1		
2. reflector	1		
3. base & wheel illuminator	1		

Cyrix Engineer	Date	Start Time	End Time
<u>Yeelhukalshna C S</u>	<u>22/2/24</u>	<u>10:00AM</u>	<u>10:30AM</u>

Customer Remark Completed Pending

Service Engineer Name : Yeelhukalshna C S
Signature : [Signature]
Date : 22/2/24
Contact Number : 8086147236

Customer Name : Mimmi / Saei Sro
Signature : [Signature]
Date : [Date]
Contact Number : 9633090531 22/2/24
Designation :
Hospital Seal :




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	CHC VENGOOR
3	Name of Equipment with Make, Model and Serial Number	Equipment : SPOT LIGHT Make : LOCAL Model : NA Sr.no-: NA
4	Equipment ID & Barcode	97158 & 0732604
5	Date of purchase / Year of manufacture /Installation Date	13-10-2008
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20/02/2024 (Toll-free)
9	Action taken	Checked the machine and found that lamp, reflector, illuminator, basement wheels and mechanical movements are faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	It is not recommend for repair

12	Cost of spares (specify parts and cost)	NA
13	Asset Value	RS.6,300
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found lamp, reflector, illuminator, basement wheels and mechanical movements are faulty. This equipment outlived 15+years and it is an outdated model. Quotation not submitted since spares are not available in the market. So recommending for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Sanathy
27/3/24

Signature of JC BM (NHM)

Superintendent / Medical Officer (i/c)


VKA

മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്
സാമൂഹിക ആരോഗ്യ കേന്ദ്രം
വേങ്ങൂർ

Signature of seal



Date





CYRD INDUSTRIAL EQUIPMENT MAINTENANCE
& REPAIR SERVICE

TOLL FREE NO: 1800 - 425-7669

BAR CODE: (5304-890615225)
0732604



