



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 171536

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Thgh</u>	Call Registration Date : <u>28/2/24</u>
Address : <u>Kothamangalam</u>	Caller ID : <u>98571</u>
<u>Ernakulam</u>	Date of Visit : <u>29/2/24</u>
Ph : <u>9446848367</u>	Asset No. : <u>0722454</u>
	EQPT Name : <u>Portable Suction</u>
	Manufacture <u>Easy care</u> Model : <u>Easy PSU</u>
	S. No. <u>15406223209</u> Dept. <u>paediatric</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : unit not working

Action Taken : check the unit, found issue with the motor and piston where not good, bottle, filter cap and tubing were damaged. Enquired the spare

Completed Date : 29/2/24 Time : 11:30AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.	NA		
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Akhil T. Lenin</u>	<u>29/2/24</u>	<u>10:00AM</u>	<u>11:30AM</u>

Customer Remark Completed Pending

Service Engineer Name : Akhil T. Lenin
Signature : [Signature]
Date : 29/2/24
Contact Number : 9603474326


Customer Name : [Signature]
Signature : [Signature]
Date : 29/2/2024
Contact Number : 9446848367
Designation : [Signature]
Hospital Seal : [Seal]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	THQH KOTHAMANGALAM
3	Name of Equipment with Make, Model and Serial Number	Equipment: PORTABLE SUCTION Make : EASYCARE Model : Easy psu-307 Sr.no:-15H06223207
4	Equipment ID & Barcode	98571 & 0722454
5	Date of purchase / Year of manufacture /Installation Date	15/03/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	28/02/2024 (Toll-free)
9	Action taken	Checked the machine and found that motor with piston, bottle, filter and tubing are faulty. Enquired the spare.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	It is not recommend for repair
12	Cost of spares (specify parts and cost)	Motor- Rs.5103 Bottle with lid- Rs.1917 Filter-Rs.271 Tubing -Rs.354 Total- Rs.7645

13	Asset Value	RS.8200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	93%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Spare estimation and Cyrix service report attached
16	Reasons for recommending the equipment as BER	The motor with piston, bottle with lid, filter and tubing are defective. The equipment outlived 5+ years (6/4/2019) and the spare cost for repairing the equipment is 93% of the actual asset value of the equipment. Hence recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Ravathy
16/11/24
Signature of JC BM (NHM)

CONSULTANT
Biomedical Engineer
National Health Mission (NHM)
Aranyakulam



SUPERINTENDENT
TALUK HEAD QUARTERS HOSPITAL
ANGALAM - 686 691

Superintendent / Medical Officer (i/c)

Signature of seal

Date



Phel gam Sact

15/3/17

R.F.S

R.F.S

2/5/17 Threat to Casualty
(St. Bk) Page no - 96

15/5/17 Issued to pt
Reg. no. 60/17
(Returned) @

11/9/17 Threat to Fmct
[St. Book. p. No - 58]

20/10/17 Issued to pt. Rajan. mu.
Reg no. 118/17
(Returned) @

19/11/2000 Sayer. K. G. 5540
Reg no. 125/20.
Kochigattai
Kothamangalam
Ph no 8113835218
(Returned)

11/1/2005 Suth 28455
Cherupurath
Puthupadi
ward - 2i
Ph: 9744 962513
(Returned Suth)


kin apparatus

Qd	Issued	Balance
4		4
Rs 7950		
	1	3
	1 (to pt)	2
	1	1
	1	0
		2
	1	1
		2
	1	1
		2

Rjn
K
ay
Raj
Raj
Raj
Raj
Raj
Raj
Raj
Raj

Sent



**CLOUD99**Office : 97444 73043
Service : 97444 73243

Sales and service of all kinds of medical equipments
Room No: 22, First Floor, CSI Golden Jubilee Complex,
Behind Shastri Road Waiting Shed, Kottayam-686001
Email: cloud99biotech@gmail.com

QUOTATION**Ref No:** C99B/0147/2023-24**Date:** 23/03/2024

To,

CHC KOTHAMANGALAM
ERNAKULAM

Dear Sir

We are pleased to submit a quotation for **CHC KOTHAMANGALAM** for Suction Apparatus. Bar code is **0722454**.

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	MOTOR	4325	18%	5103
2	BOTTLE WITH LID – 1 NOS	1625	18%	1917
3	FILTER	230	18%	271
4	TUBINGS	300	18%	354
	TOTAL			7645

Terms & Conditions

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully

FOR CLOUD99 BIOTECH**AUTHORIZED SIGNATORY**

We deal with sales and service of all kinds of hospital equipments, neonatal equipments, laboratory equipments, dental equipments, suction, hospital furniture, radiology equipments, physiotherapy equipments, all type of autoclaves. ICU equipments. etc.

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

DE - (8004-890615225)
0722454



0722454





EASYCARETM
If it's Health!

ECPSU-307
PORTABLE PHLEGM SUCTION UNIT

PALLIATIVE
THRCT.
Kothamangalam

Active Capacity

1000
800
600
400