



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 204141

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>phe</u></p> <p>Address <u>anamthunth</u> <u>Kottayam</u></p> <p>Ph : <u>9387108963</u></p>	<p>Call Registration Date : <u>23/5/2024</u></p> <p>Caller ID : <u>112849</u></p> <p>Date of Visit : <u>27/5/2024</u></p> <p>Asset No. : <u>0542132</u></p> <p>EQPT Name : <u>BP APPARATUS</u></p> <p>Manufacture <u>ANITA</u> Model : <u>ELICOMETER</u></p> <p>S. No. <u>68052111</u> Dept. <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Mercury leak

Action Taken :
Checked and found out one mercury tank leaked and spilled out one mercury. screw of hinges are broken


Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. /	/	/	/
2. /	/	/	/
3. /	/	/	/

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>27/5/2024</u>	<u>1:45pm</u>	<u>2:30pm</u>

Customer Remark Completed Pending
Equipment is not working



<p>Service Engineer Name : <u>AMAL BABU</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/5/2024</u></p> <p>Contact Number : <u>9567575983</u></p>	<p>Customer Name : <u>Sarangi's</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/5/24</u></p> <p>Contact Number : <u>[Number]</u></p> <p>Designation : <u>staff</u></p> <p>Hospital Seal : <u>9387108963</u></p>
--	---


Medical Officer in charge
Poonithura Health Centre, Poonithura
[Signature]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC ONAMTHURUTH
3	Name of Equipment with Make, Model and Serial Number	Equipment :BP APPARATUS Make: ANITA Model:Elkometer SN: 68052111
4	Equipment ID & Barcode	112849 & 0542132
5	Date of purchase / Year of manufacture /Installation Date	18/08/2008
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	23/05/2024 (Toll free)
9	Action taken	Checked and found out the mercury tank leaked and spilled out the mercury, Screw of hinges are broken
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	Rs. 850/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the mercury tank leaked and spilled out the mercury, Screw of hinges are broken. Machine installed in 18/08/2008 and covered up to 15 + years. Mercury not available in the market. So recommending to condemnation
17	Name & Signature of CYRIX Authority	AMAL BABAU 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - Rs 850
 Cost of spare - Mercurid as spare not available.
 Life period - Installation Date 18/8/2008
 15 years
 As per the RBER guidelines recommended as RBER (criteria satisfied)


 15/6/24
 Signature of JCBM (NHM)

CONSULTANT BIOMEDICAL
 NATIONAL HEALTH MISSION
 KOTTAYAM

Date _____ Seal _____




 Signature of Medical Officer in charge
 Superintendent / Medical Officer (i/c)
 Primary Health Centre Onamthuruth

(Ann
mea

B.P Apparatus, Stethoscope

bits

Date

Receipt JESKE Babu

18/8/09

B.P apparatus	850	1
Stethoscope	450	1
Tax	53	

[Signature]



Medical Officer in charge
Primary Health Centre Onamthuruth

FAC No: 1111
S-11111111



CYRIX
HEALTHCARE SERVICES
Cm : POCO 28
Dnc : Op apparatus
ID : 0542132
CEL. No : 21-2-23
CEL. No : 120-2-23
ISO 9001 : 2008 Certified Co. No: B7440251

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225)
0542132

dy Medical Officer
Primary Health Centre Onamthuruth
PIN: 686 602
ONAMTHURUTH