



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 204777

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : Phc	Call Registration Date : 21/5/2024
Address : Phc Trivikovilvattom	Caller ID : 112270
kollam	Date of Visit : 22/5/2024
Ph : 8943256599	Asset No. : 0250987
	EQPT Name : ECG Machine
	Manufacture BPL Model : 6108T
	S. No. AYMTIT29477 Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : ECG noise issue, battery & ECG cable defective

Action Taken : Checked and found that main pcb, battery & ecg cable defective
Enquire spare with OEM


Completed Date : 22/5/2024 Time : 11:40 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time

Customer Remark Completed Pending



Service Engineer Name : Rajeshlal A.S
Signature :
Date : 22/5/2024
Contact Number : 8593847128


Customer Name : Aswathy
Signature :
Date : 22/5/24
Contact Number : 94380 4263
Designation : Nursing officer
Hospital :

REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOLLAM
2	Name of Hospital	PHC TRIKOVILVATTOM
3	Name of Equipment with Make, Model and Serial Number	Equipment : ECG MACHINE Make: BPL Model: 6108T SN : AVMT7J29474
4	Equipment ID & Barcode	Ticket id: 112270 / Barcode : 0250987
5	Date of purchase / Year of manufacture / Installation Date	28-02-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	21-05-2024
9	Action taken	Checked and found that Main Pcb, battery & ecg cable defective, Enquire spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	ECG 6108T Main Pcb Assy = 14000/- NIMH Battery 9.6v 1.4ah = 2048/- 10 lead Ecg Cable for 6108T = 3136/- Total = 19184/-


13	Asset Value	Rs .21500
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	89.22%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report, Email communication & Quotation attached
16	Reasons for recommending the equipment as BER	Checked and found Main Pcb, battery & ecg cable defective, the machine installed on 28-02-2019 equipment outlived 5 years 2 month, Repairing cost of equipment is 89.22%. As per tender clause 5.3.14.1, both criteria met. so recommending to condemnation.
17	Name & Signature of CYRIX Authority	RAJESHLAL 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Consultant (Biomedical) NHM:

Checked quotation received from the company and found that repairing cost is very high. So recommended for beyond economic repair


 20/06/24
 Signature of J.C. BM (NHM)
 ABMO
 NHM, KOLLAM

Date 20/06/2024



 Seal


 Signature of
 Superintendent Medical Officer (i/c)
 Primary Health Centre
 Thiruvananthapuram
 Kollam-1800

BPL

CARDIART 6108T

Always use consumables supplied by BPL for best performance



Regal Offset-In-Chain
 Trinity Hallis Centre
 Karamana-78



BPL CARDIART 6108T
 VOLTAGE INBUILT RECHARGEABLE
 NiMH BATTERY 2.6V
 POWER CONSUMPTION 12W MAX
 SN AVMT7J29474
 BPL MEDICAL TECHNOLOGIES PRIVATE LIMITED
 BPL Works, Palakkad - 678 007 Kerala, India
 Ph: 1800-425 2355 e-mail: regulatory.healthcare@bpl.in

CN
 ID
 CALL
 CALL

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
CYRIX
 TOLL FREE NO: 1800-425-1861
 BAR CODE - (8004-890815225)
 0250987



174

ECG Machine

✓ Sale		Recd	Issd	Balance
28/2/19	Rtd from store	1	0	1

[Handwritten signature]



Medical Officer-in-Charge
Primary Health Centre
Parikkovilattom
Kottayam - 686 012

QTN/BISO/040624-017

04/06/2024

QUOTATION

To,
CYRIX HEALTH CARE PVT LTD
ERNAKULAM

Dear Sir,

Please find below our lowest quotation for the following items:-

SL NO.	ITEM DESCRIPTION	QUANTITY	UNIT PRICE	PRICE (incl of all taxes)
1	ECG 6108T MAIN PCB ASSY NEW (WITH LPC IC)	01	12500+12%TAX	14000/-
2	BM/OM NIMH BATTERY 9.6V 1.4 AH	01	1600+28%TAX	2048/-
3	10 LEAD ECG CABLE FOR 6108T	01	2800+12%TAX	3136/-


Terms and conditions:

1. Payment 100% in advance along with Purchase order
2. Delivery within 15 days

Thanking you and assuring you our best attention always.

For Biomedix Solutions




Authorized Signatory

Mobile: 9846969564

Branches: Kozhikode, Kannur, Kottayam-Kollam.

For enquiries contact: biomedixsales@gmail.com, biomedixservice@gmail.com



BIOMEDIX SOLUTIONS
No: 7/437- B, Mannarakayam P.O,
Kathilankalpadi, Kanjirapally, Kottayam - 686506,
Mob:9846969564.

Re: PRICE ENQUIRY - ECG MACHINE - PHC TRIKOVILVATTOM - 0250987

1 message

BPL Medical Equipment Service <biomedixservice@gmail.com>
To: "Abilal S [_DI Kollam_] " <abilalcyrix@gmail.com>

4 June 2024 at 11:56

Dear sir,
Please find the quotation attached herewith. Stock is available .

Regards
Biomedix Solutions
Kottayam

On Sat, 1 Jun 2024 at 22:40, Abilal S [_DI Kollam_] <abilalcyrix@gmail.com> wrote:

Dear Sir,

Kindly share the Price and Stock availability of Spares needed for ECG MACHINE.

Spares Needed :

6108T Main PCB
Battery
Ecg Cable

Hospital : PHC TRIKOVILVATTOM
Make : BPL
Model : Caidiart6108t
SI No : AVMT7J29474
Barcode. : 0250987

--

Thanks and Regards

ABILAL .S

District incharge

CYRIX HEALTHCARE Pvt. Ltd

Kollam - 691002

Ph:+91 7593847161

 **QTTN CYRIX HEALTH CARE 01720240604_11465165.pdf**
138K