



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 218511

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bomp.kl@cyrilx.in

Health Facility <u>SPHL</u>	Call Registration Date : <u>6/5/24</u>
Address <u>Thiruvananthapuram</u>	Caller ID : <u>109748</u>
<u>Kerala</u>	Date of Visit : <u>14/5/24</u>
Ph : <u>9400722854</u>	Asset No. : <u>0111446</u>
	EQPT Name : <u>Centrifuge</u>
	Manufacture <u>Rotek</u> Model : <u>Plus</u>
	S. No. <u>XX190</u> Dept. <u>Molecular Diagnostic Section</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked and found that motor defective. Need to replace the spare for further checking.

Completed Date : 14/5/24 Time : 11:15AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.			
3.			


Cyrilx Engineer	Date	Start Time	End Time
<u>Soorya Sudheep</u>	<u>14/5/24</u>	<u>11:00AM</u>	<u>11:15AM</u>

Customer Remark Completed Pending

Service Engineer Name : Soorya Sudheep Signature : [Signature] Date : 14/5/24 Contact Number : 7907400244

Customer Name : [Signature] Date : 18.6.24 Contact Number : [Signature] Designation : [Signature] Hospital Seal : [Signature]


JAICHAND. J
SCIENTIFIC OFFICER
State Public Health & Clinical Laboratory
Thiruvananthapuram



REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)
Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	STATE PUBLIC HEALTH LABORATORY
3	Name of Equipment with Make, Model and Serial Number	Equipment Name: CENTRIFUGE Manufacturer: ROTEK PLUS Model: RCH-8C SI: NA
4	Equipment ID/Barcode	109748/0111446
5	Date of purchase/ Year of manufacture/ Installation Date	22/07/2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/Tollfree)	06/05/2024 (Tollfree)
9	Action taken	Checked the found motor defective. Need to replace for further check: g.
10	Present status of the equipment (Fully damaged/partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Centrifuge Motor - 4600/-

13	AssetValue	Rs .5000/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/AssetValue	92%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and vendor quotation attached
16	Reasons for recommending the equipment RBER	Checked and found that the centrifuge motor defective . The unit is installed on 22/07/2020 and covered upto 3+ years. The repairing cost is 92 % . As per the tender clause 5.3.14.1, Both criteria met. Hence recommending for condemnation.
17	Name & Signature of CYRIX Authority	Kasyep PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

*Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor defective. Repair cost is 92% of asset as per the quotation received . . .


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


10-6-2024

Signature of JCBM(NHM)



Date:


18.6.24
Signature of Superintendent/Medical Officer (I/c)

DIRECTOR
State Public Health &
Clinical Laboratory
Thiruvananthapuram



9744473043 | 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastry Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KL-07--0030714

Estimate

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT
LTD KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-15

Date: 27/05/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	CENTRIFUGE MOTOR (MOTOR FOR CENTRIFUGE)	8421	1	Nos	₹ 3,898.31	₹ 701.69 (18%)	₹ 4,600.00
Total			1			₹ 701.69	₹ 4,600.00



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121



Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 3,898.31
SGST@9%	₹ 350.85
CGST@9%	₹ 350.85
Total	₹ 4,600.00
You Saved	₹ 0.01

Description

EQUIPMENT: CENTRIFUGE

MAKE: ROTEK

TICKET ID: 109748

BAR CODE: 0111446

HOSPITAL: STATE PUBLIC HEALTH


Estimate Amount In Words

Four Thousand Six Hundred Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



Repairing Quotation for Centrifuge Barcode 0111446

1 message

CLOUD 99 <cloud99biotech@gmail.com>
To: kasyepcyrix@gmail.com <kasyepcyrix@gmail.com>

Tue, 28 May 2024 at 11:21 am

SUSAN JOHN
Cloud99 Biotech
Medical Engineering Works
Room No: 22, CSI Golden Jubilee Complex
Kottayam 686 001
Kerala, India
Ph: 9744473043
email: cloud99biotech@gmail.com

