



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 178139

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kk@cyrix.in

Health Facility <u>PHC</u>	Call Registration Date : <u>8/4/24</u>
Address <u>Bakrampuzham</u>	Caller ID : <u>105297</u>
<u>Tiruvadur</u>	Date of Visit : <u>8/4/24</u>
Ph : <u>90747 21876</u>	Asset No. : <u>019 324 2</u>
	EQPT Name : <u>Centrifuge</u>
	Manufacture <u>Cole</u> Model <u>Lab</u>
	S. No. <u>302106</u> Dept. <u>Lab</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching on.

Action Taken : Checked the machine and found that motor controller are fully damaged. Need to replace the spare for further checking of the machine.

Completed Date : 8/4/24 Time : 5:30pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrilx Engineer	Date	Start Time	End Time
<u>Deenan</u>	<u>8/4/24</u>	<u>4:30pm</u>	<u>5:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Deenan</u>	Customer Name : <u>for</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>8/4/24</u>	Date : <u>8/4/24</u>
Contact Number : <u>90747 21876</u>	Contact Number : <u>90747 21876</u>
	Designation : <u>Lab Technician</u>
	Hospital Seal : <u>Fdc Balamapuralam</u>




[Signature]
MEDICAL OFFICER IN-CHARGE
FAMILY HEALTH CENTRE
BALAMAPURALAM

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Thiruvananthapuram
2	Name of Hospital	PHC Balaramapuram.
3	Name of Equipment with Make, Model and Serial Number	Equipment :Centrifuge Make :C-Poul0 Labtech Medico Pvt Ltd Model :80-2 SN : 308106
4	Equipment ID & Barcode	105277 & 0143242
5	Date of purchase / Year of manufacture /Installation Date	25-10-2014
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07-04-2024 (Toll free)
9	Action taken	Checked and found motor and speed control defective. Enquired the spare with vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	1)Centrifuge motor – Rs-4600 2)Speed control -Rs-1300 Total – Rs.5900,

13	Asset Value	Rs . 9651/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	61%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor quotation attached
16	Reasons for recommending the equipment as BER	Checked and found motor and speed control defective. Equipment installed on 25-10-2014 and aged up to 9 + years. Repairing cost of the equipment is 61% both criteria met, so recommended equipment for condemnation as per the tender clause 5.3.14.1.
17	Name & Signature of CYRIX Authority	Kasyep .PV 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Motor and speed control damaged. Machine is installed on 2014. So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


27-5-2024
Signature of JC BM (NHM)

Date




MEDICAL OFFICER IN-CHARGE
 SUPERVISOR / BALARAMAPURAM
 FAMILY HEALTH CENTRE
 Medical Officer (i/c)



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO: : UDYAM-KL-07--0030714

Estimate

Estimate For:

ABHIJITH

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT
LTD KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847141

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C99B-QTN-12

Date: 16/05/2024

Place of Supply: 32-Kerala

#	Item name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	CENTRIFUGE MOTOR (MOTOR FOR CENTRIFUGE)	8421	1	Nos	₹ 3,898.31	₹ 701.69 (18%)	₹ 4,600.00
2	SPEED CONTROL FOR CENTRIFUGE (SPEED CONTROL FOR CENTRIFUGE)	8479	1	Nos	₹ 1,101.69	₹ 198.31 (18%)	₹ 1,300.00
Total			2			₹ 900.00	₹ 5,900.00



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH



Sub Total	₹ 5,000.00
SGST@9%	₹ 450.00
CGST@9%	₹ 450.00
Total	₹ 5,900.00
You Saved	₹ 0.01

Description

EQUIPMENT: CENTRIFUGE
MAKE: C- POULO
MODEL: 80-2
SL NO: 308106

Estimate Amount In Words

Five Thousand Nine Hundred Rupees only

Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



Abijith A <abijithacyrix@gmail.com>

REGADDING QUOTATION - PHC BALARAMAPURAM - CENTRIFUGE @0143242

2 messages

Abijith A <abijithacyrix@gmail.com>

To: CLOUD 99 <cloud99biotech@gmail.com>

Thu, May 16, 2024 at 3:34 PM

Cc: Kasyep P V <kasyepcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Dear Sir,

Kindly share a quotation for Motor and speed control.

Equipment - Centrifuge

Make - C -Poulo. Labtech Medico P Ltd

Model - 80-2

Serial no - 308106

—
Thanks & Regards,**Abijith A****Divisional Incharge****Mob-7593847141**

abijithacyrix@gmail.com

www.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

CLOUD 99 <cloud99biotech@gmail.com>

To: Abijith A <abijithacyrix@gmail.com>

Thu, May 16, 2024 at 4:38 PM

Dear Sir

Kindly find the attachment for the quotation

SUSAN JOHN

Cloud99 Biotech

Medical Engineering Works

Room No: 22, CSI Golden Jubilee Complex

Kottayam 686 001




Kerala, India

Ph: 9744473043

email: cloud99biotech@gmail.com

[Quoted text hidden]

 **Estimate_C99B-QTN-12_16_05_24.pdf**
198K



[Handwritten signature]

വെഡിക്കൽ ഓഫീസർ ഇൻ-ചാർജ്ജ്
MEDICAL OFFICER IN-CHARGE
കുടുംബാരോഗ്യ കേന്ദ്രം
FAMILY HEALTH CENTRE
ബലരാമപുരം / BALARAMAPURAM