


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

505

Recommendations for Beyond Economic Repair (BER)


PROFORMA



Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	STATE PUBLIC HEALTH LABORATORY
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : PH Meter Make : Micropro Model: Gradimate SI : NA
4	Equipment ID/ Barcode	104954 / 0111323
5	Date of purchase/ Year of manufacture/ Installation Date	25/04/2008
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
38	Date of breakdown (Date of registration of complaint through email/ Toll free)	05/04/2024
9	Action taken	Checked and found mainboard, conductive electrode and reference electrode defective. Need to replace these spares for the further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair.
12	Cost of spares (specify parts and cost)	Conductive Electrode - 2360/- Reference Electrode - 2301/- Total - 4661/-

13	Asset Value	Rs . 6437/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	72.40. %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and OEM Quotation Attached
16	Reasons for recommending the equipment as BER	Checked and found mainboard,conductive electrode and reference electrode defective. The equipment was installed on 25/4/2008 and aged upto 16+ years. The repairing cost is 72.40 %. As per the tender clause 5.3.14.1,Both crieteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

<p>Remarks and Recommendations of Junior Consultant (Biomedical) NHM</p> <p>Main board damaged. Electrodes are also defective. So as per guidelines, this can be RBER.</p> <p style="text-align: center;">MANEESHA MOHAN Junior Consultant (Biomedical) National Health Mission Thiruvananthapuram</p>	<p style="text-align: right;"> 10.6.2024 Signature of JC BM (NHM)</p>
---	--

<p>Date:</p> <p style="text-align: center;"></p>	<p style="text-align: right;"> 18.6.24 Signature of Superintendent/Medical Officer (i/c) DIRECTOR</p>
---	---

**State Public Health &
Clinical Laboratory
Thiruvananthapuram**



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 182617

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>State Public Health Lab</u></p> <p>Address <u>Thiruvananthapuram</u></p> <p><u>Kerala</u></p> <p>Ph : <u>9961 29 2176</u></p>	<p>Call Registration Date : <u>05/04/24</u></p> <p>Caller ID : <u>104954</u></p> <p>Date of Visit : <u>05/04/24</u></p> <p>Asset No. : <u>0111323</u></p> <p>EQPT Name : <u>PH Meter</u></p> <p>Manufacture <u>Micropo</u> Model : <u>Leadmate</u></p> <p>S. No. <u>00452</u> Dept. <u>Medic</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not working.

Action Taken : Checked and found that the main lead and electrode detector. Need to replace these parts & check the further working condition of the machine

Completed Date : 05/04/24 Time : 03:00 pm Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>NR</u>	<u>NR</u>	<u>NR</u>	<u>NR</u>
3.				


Cyrux Engineer	Date	Start Time	End Time
<u>Mohammed Sijath M</u>	<u>05/04/24</u>	<u>02:30 pm</u>	<u>03:00 pm</u>

Customer Remark Pending

<p>Service Engineer Name : <u>Mohammed Sijath M</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>05/04/24</u></p> <p>Contact Number : <u>7293499177</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>05/04/24</u></p> <p>Contact Number : <u>[Blank]</u></p> <p>Designation : <u>Scientific Officer</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	--



JAICHAND. J
SCIENTIFIC OFFICER
State Public Health & Clinical Laboratory
Thiruvananthapuram
18/6/24



MEDPLUS BIOMEDICALS

Athulraj M
Kumaraswamy raja road
Chelannur (P.O)
Kozhikode

☎ 8848372192 ✉ medplusbiomedicals.clt@gmail.com
GSTIN 32CEYPA2140F1ZV

Quotation

To,
Cyrix Healthcare
Manager
Kochi

Quotation# Quote-23

Date: 29/05/2024

Dear Sir/Mam,

We are pleased to submit the purchase order as below.


#	DESCRIPTION	QTY	PRICE	GST	TOTAL
1	Conductive electrode	1	₹ 2,000.00	₹ 360.00 18.0%	₹ 2,360.00
2	Reference electrode	1	₹ 1,950.00	₹ 351.00 18.0%	₹ 2,301.00
SUB TOTAL					₹ 3,950.00
GST					₹ 711.00
GRAND TOTAL					₹ 4,661.00

Your prompt attention to this order is greatly appreciated, and we look forward to a successful transaction.

Terms & Conditions:

- Quotation validity : one month
- 100 % advance
- Delivery : 7 days after confirmation

For, MEDPLUS BIOMEDICALS



AUTHORIZED SIGNATURE

