



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 206461

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CHC</u>	Call Registration Date : <u>11/3/24</u>
Address : <u>Ramamangalam</u>	Caller ID : <u>100558</u>
<u>Ernakulam, Kerala</u>	Date of Visit : <u>16/3/24</u>
Ph : <u>9947728905</u>	Asset No. : <u>0730025</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture : <u>TRISCO</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>Casualty</u>

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Equipment motor was sully damaged

**Action Taken :** Checked the equipment and found that motors with piston, bottle, filter and tubing are defective. Enquired the spare.

Completed  Date : ..... Time : ..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Amal</u>	<u>16/3/24</u>	<u>11:00am</u>	<u>1:30pm</u>

Customer Remark  Completed  Pending

Service Engineer Name : Amal  
Signature : [Signature]  
Date : 16/3/24  
Contact Number : 8281015732




Customer Name : Suja ma  
Signature : [Signature]  
Date : 16/3/24  
Contact Number : 9447728905  
Designation :  
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	Ernakulam
2	Name of Hospital	CHC RAMAMANGALAM
3	Name of Equipment with Make, Model and Serial Number	Equipment: SUCTION APPARATUS Make : TRISCO Model : NA Sr.no-: NA
4	Equipment ID & Barcode	100558 & 0730025
5	Date of purchase / Year of manufacture /Installation Date	18/1/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	16/03/2024 (Toll-free)
9	Action taken	Checked the machine and found that motor with piston, bottle, filter and tubing are faulty. Enquired the spare.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	It is not recommend for repair
12	Cost of spares (specify parts and cost)	Motor- Rs.5487 Bottle with lid- Rs.3837 Filter-Rs.271 Tubing -Rs.354  Total- Rs.9947


13	Asset Value	RS.10,800
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	92%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Spare estimation and Cyrix service report attached
16	Reasons for recommending the equipment as BER	The motor with piston, bottle with lid, filter and tubing are defective. The equipment outlived 6+ years (18/1/2018) and the spare cost for repairing the equipment is 92% of the actual asset value of the equipment. Hence recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	Amal Sasi 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

  
 Signature of JC BM (NHM)



Superintendent / Medical Officer (i/c) Signature of seal Date



..... suction apparatus .....

Date	Received from or issued to	Inv. No. or V. No.	Amount		Receipts	Issues	Balance	Remarks
			Rs.	Ps.				
	RFS on 18/11/2018	1011	7800	00	1 CH	0730031		tip
	RFS on 15/10/2020 (Cochus Sappals Kulasa)	90	9115	00	1	0730025	2	tip
8/11	Taken out to unassemble articles page 25					1	1	Verified date 2/11

**CLOUD99**

Office : 97444 73043

Service : 97444 73243

*Sales and service of all kinds of medical equipments*  
Room No: 22, First Floor, CSI Golden Jubilee Complex,  
Behind Shastri Road Waiting Shed, Kottayam-686001  
Email: cloud99biotech@gmail.com

**QUOTATION****Ref No:** C99B/0146/2023-24**Date:** 23/03/2024

To,

CHC RAMAMNGALAM

ERNAKULAM

Dear Sir

We are pleased to submit a quotation for **CHC, RAMAMNGALAM** for Suction Apparatus. Bar code is **0730025**.

SL No	ITEM NAME	PRICE PER UNIT	GST	PRICE INCLUDING GST
1	MOTOR	4650	18%	5487
2	BOTTLE WITH LID - 2 NOS	3250	18%	3835
3	FILTER	230	18%	271
4	TUBINGS	300	18%	354
	TOTAL			9947

**Terms & Conditions**

1. Taxes: GST @ 18 % inclusive
2. Payment Terms: 100% advance payment
3. Quotation validity: 15 days from date of quotation

We hope you will find our rates are competitive and wait for your valued order.

Thanking you,

Yours Faithfully

FOR CLOUD99 BIOTECH

AUTHORIZED SIGNATORY

e deal with sales and service of all kinds of hospital equipments, neonatal equipments, laboratory equipments, dental equipments, suction, hospital furniture, radiology equipments, physiotherapy equipments, all type of autoclaves, ICU equipments, etc.

FILTER



POWER



SUCTION

ADT WORKING 9/21



Nebulising Room  
TRISCO



FUSE

FUSE

ON / OFF

V. CONTROL

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)



**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-890615225)  
0730025**





FILTER

ADD WATER 9/11



SUCTION

Nebulising Recycler  
TRISCO

ON/OFF

CONTROL

FUSE

FUSE