


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

528

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	WOMEN & CHILD HOSPITAL THYCAUD
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Make : Anita Industries Model : Elkometer Serial No : NA
4	Equipment ID/ Barcode	118743/0115C28
5	Date of purchase/ Year of manufacture/Installation Date	19/04/2007
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22/06/2024
9	Action taken	Checked and found mercury spillage. Identified mercury tank and glass rod defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1323/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not Available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found Mercury spillage, Identified mercury tank and glass rod defective. The unit is installed on 19/04/2007 and covered upto 17+ years. Quotation not attached since the mercury is not available in the market. So recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach: Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mercury spillage, tank and glass rod defective. So as per tender conditions can be recommended beyond economic repair.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


4-07-2023
Signature of JC BM (NHM)

Date: 5/7/2024




SUPERINTENDENT
Govt. W & C Hospital, Thiruvananthapuram - 14
Superintendent/Assistant Officer (i/c)

B.P. Apparatus
particulars

Date	Particulars	Receipt	Issue	Balance	Initials
	B.P. Apparatus (315)	1			Surolice
5/2/04	BP Apparatus	5	2	4	Lillys
	Transferred to G.W.D. - 2		2		
	Received from BP. outfit				J.P.
12/2/04	B.P. Apparatus transferred to Postnatal Ward		1	3	Lillys
	Received.				Sakishel
12/2/04	1 BP Apparatus transfer to N.N. Ward		1	2	Lillys
13/2/04	1 BP Apparatus transfer to N. Room		1	1	Lillys
					Received on 13/2/04 - D. ...
19/4/07	1 BP Apparatus A.C.S. page no. 18	1	1		
	Vol II Pro. 315				



Santhi

SUPPLY OFFICER
G.W.D. & C.I. Thycud
Mithamantapuram - 14



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 224166

528

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/04 1 D, Patta Junction, Poonthura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bamp.kl@cyrilx.in

Health Facility : <u>W.A.C</u>	Call Registration Date : <u>22-6-2024</u>
Address : <u>Thycaud, Thiruvananthapuram</u>	Caller ID : <u>118743</u>
Ph : <u>8921557053</u>	Date of Visit : <u>26-6-2024</u>
	Asset No. : <u>0115028</u>
	EQPT Name : <u>BP Apparatus</u>
	Manufacture <u>Eukometry</u> Model : <u>NA</u>
	S. No. <u>NA</u> Dept. <u>OT</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP not getting, mercury spillage

Action Taken : Checked and found mercury spillage. Identified BP glass rod and mercury tank defective. Need to replace these spares for further checking.

Completed Date : 26-6-2024 Time : 12:30 p.m Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. NA	NA	NA	NA
2. NA	NA	NA	NA
3. NA	NA	NA	NA

Cyrilx Engineer	Date	Start Time	End Time
<u>Rehna . R</u>	<u>26-6-2024</u>	<u>12:10 p.m</u>	<u>12:30 p.m</u>

Customer Remark : Completed Pending

Service Engineer Name : Rehna . R
Signature : Rehna
Date : 26-6-2024
Contact Number : 9947239956

Customer Name : Govt. W & C Hospit. Thycaud
Signature : [Signature]
Date : 26/6/24
Contact Number : 8547117208
Designation : 820
Hospital Seal : 26/6/24 820



SUPERINTENDENT
Govt. W & C Hospit. Thycaud
Thiruvananthapuram

