



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 220066

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC Kadaplattom</u></p> <p>Address <u>Rottayam</u></p> <p>Ph : <u>9497089076</u></p>	<p>Call Registration Date : <u>20/6/2024</u></p> <p>Caller ID : <u>118263</u></p> <p>Date of Visit : <u>24/6/2024</u></p> <p>Asset No. : <u>0531006</u></p> <p>EQPT Name : <u>Digital BP Apparatus</u></p> <p>Manufacture <u>Life Care</u> Model : <u>PW-215</u></p> <p>S. No. <u>NA</u> Dept. <u>SUB Centre</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked the machine NIBP pump and PCB Board not working and damaged. Spare are not available in the market so recommended the machine for condemnation.

Completed Date : 24/6/2024 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Josekutty George</u>	<u>24/6/2024</u>	<u>12 pm</u>	<u>12:30 pm</u>

Customer Remark Completed Pending



MEDICAL OFFICER IN CHARGE
CHC - KADAPLATTOM

Service Engineer Name : Josekutty George
Signature : [Signature]
Date : 24/6/2024
Contact Number : 8848230407


Customer Name : Sandhya K.V.
Signature : [Signature]
Date : 24/6/24
Contact Number : 9497089076
Designation : [Signature]
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	CHC KADAPLAMATTOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :BP APPARATUS Make: LIFE CARE Model :PW -215 SN: PW/DS/0818
4	Equipment ID & Barcode	118263 & 0531006 ✓
5	Date of purchase / Year of manufacture /Installation Date	21/09/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20/06/2024 (Toll free)✓
9	Action taken	Checked and found out the problem main PCB and NIBP pump
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	Rs. 1750/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and NIBP pump. Machine installed in 21/09/2019 and covered up to 4 + years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Josekutty George 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset Value - 1750
 Cost of spare - Spare not available
 report from Cyrix
 Life period - 4 years -
 As per the BER guideline recommended as BER

CONSULTANT BIOMEDICAL ENGINEER
 CONSULTANT BIOMEDICAL ENGINEER
 NATIDU


3/7/21
 Signature of JC BM (NHM)

Date




 Signature of
 Superintendent / Medical Officer (i/e)
MEDICAL OFFICER IN CHARGE
 CHC - KADAPLAMATTOM

Blood Pressure Monitor



SYS
mmHg

DIA
mmHg

PUL
/MIN

Blood Pressure Monitor
 Pw-233
 Fully Automatic Electronic
 Blood Pressure Monitor
 Batch No. : PW/DS/0816
 Battery : 6V (4* AA Batteries)



(414)68(254)05(21)4245

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0531006



MEDICAL OFFICER IN CHARGE
 CHC - KADAPLAMATTOM