



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 220707

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC</u></p> <p>Address : <u>PHC Meenadom</u> <u>Kottayam</u></p> <p>Ph : <u>8086366437</u></p>	<p>Call Registration Date : <u>18/06/2024</u></p> <p>Caller ID : <u>117718</u></p> <p>Date of Visit : <u>22/06/2024</u></p> <p>Asset No. : <u>0564211</u></p> <p>EQPT Name : <u>BP Apparatus (Digital)</u></p> <p>Manufacture : <u>Niscomed</u> Model : <u>PW-215</u></p> <p>S. No. <u>PW/DS/0820</u> Dept. <u>OP</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP Apparatus is not Working
Found problem with main PCB and NIBP Pump

Action Taken : Machine installed on 1/03/2018, covered 6+ years. Spare parts were not available in the market to purchase. The equipment is irreparable.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sanish Sivan</u>	<u>22/6/24</u>	<u>9:30 am</u>	<u>11:30 am</u>

Customer Remark Completed LPending

Equipment is not working. Needs replacement. Bindumol. T.S, Nig-office
Meenadom


<p>Service Engineer Name : <u>Sanish Sivan</u></p> <p>Signature : </p> <p>Date : <u>22/6/2024</u></p> <p>Contact Number : <u>9946646571</u></p>	<p>Customer Name : <u>Dr. Sujimon PP</u></p> <p>Signature : </p> <p>Date : <u>29/6/2024</u></p> <p>Designation : <u>Dr. Sujimon PP</u></p> <p>Hospital Seal : </p>
---	---

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC MEENADOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :BP APPARATUS Make: NISCOMED Model :PW -215 SN: PW/DS/0820
4	Equipment ID & Barcode	117718&0544211
5	Date of purchase / Year of manufacture /Installation Date	1/03/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	18/06/2024 (Toll free)
9	Action taken	Checked and found out the problem main PCB and NIBP pump
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	Rs. 1741/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main PCB and NIBP pump. Machine installed in 1/03/2018 and covered up to 6+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sanish sivan 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value 1741
 cost of spare - mentioned as spare not available.
 life period - 5 year
 AS per the RBER guidelines recommended AS RBER


 CONSULTANT BIOMEDICAL ENGINEER
 NATIONAL HEALTH ASSOCIATION
 Signature of JC BM (NHM)

Date



 Dr. Srinivas P.P.
 29/6/2024
 Signature of
 Superintendent Medical Officer



B. P. APPARATUS DIGITAL
Moresen - (9120)

Date	Receipt	Issue	Bal	Sign
01/03/18	From Saij Dental and Surgical, Thoppil Complex Near S.H. Lab, Kandhinagar P.O., KTM. Invoice no 336/17/18 dt. 01/03/18, Rate: 1741 Total bill amt Rs. 5810/-	3 nos.	3 nos.	Dr. [Signature]
01/03/18	Clo to page 35 bound to 111 of stock registers of equipments and instruments	3	nil	Dr. [Signature]
01/03/18	From Saij Dental and Surgical, Thoppil Complex, Near S.H. Lab, Kandhinagar P.O., KTM. Invoice no 353/17/18, dt. 01/03/18. Total bill amt Rs. 17051/- Rs. 1741.07	3 nos.	3 nos.	Dr. [Signature]
01/03/18	Clo to page no. 35 of stock registers for equipments & instruments.	3	nil	Dr. [Signature]
28/03/18	From Saij Dental & Surgical, Kandhinagar P.O., KTM. Invoice no: 366/17/18 dt. 22/03/18, Rs. 1741.07 Total bill amt Rs. 3635/-	1	1	Dr. [Signature]
28/03/18	Clo to page no. 35 of stock registers for equipments & instruments	1	nil	Dr. [Signature]
29/10/20	Inv: 20BC1919 Standard Surgical Syndicate Dt 29/10/20. Dista B.P App. Dr. Moresen B.P.02 Rate 1200.00	1	1	Dr. [Signature]
31/10/2020	Issued to S/W stock pg: 52	1	nil	Dr. [Signature]

**GENERAL OFFICE
PRIMARY HEALTH CENTRE
MADON - 505 51
MOTTAYAM**



original to office



Handwritten signature
 MEDICAL OFFICE
 PRIVATE HEALTH CENTRE
 MEERADON - 300 51
 30 ITA YAN



(4)

(5)