



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 203915

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report - 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC Karikkalhoor</u></p> <p>Address <u>Karikkalhoor,</u> <u>Kottayam</u></p> <p>Ph : <u>7560824694</u></p>	<p>Call Registration Date : <u>3/6/2024</u></p> <p>Caller ID : <u>114786</u></p> <p>Date of Visit : <u>4/6/2024</u></p> <p>Asset No. : <u>0541269</u></p> <p>EQPT Name : <u>Weighing Machine</u></p> <p>Manufacture <u>Samis</u> Model : <u>adult</u></p> <p>S. No. <u>NA</u> Dept. <u>Sub Centre</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Display not working. Digits not clear

Action Taken : Cleaned the display connection board & disassembled the board. The complaint is not resolved. Need to replace the main board of display in order to check the working condition of the equipment.

Completed Date : 4/6/2024 Time : 12:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sreyas Suresh</u>	<u>4/6/2024</u>	<u>10:40am</u>	<u>12:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : Sreyas Suresh
Signature : [Signature]
Date : 4/6/2024
Contact Number : 7902854749

Customer Name : [Signature]
Signature : [Signature]
Date : 04/06/24
Contact Number : 9497574718
Designation : Pharmacist
Hospital Seal :



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**


Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC KARIKKATTOOR
3	Name of Equipment with Make, Model and Serial Number	Equipment : Weighing Machine Make : SAMSO Model :150kg SN:200720052250
4	Equipment ID & Barcode	114786 & 0541269
5	Date of purchase / Year of manufacture /Installation Date	12/10/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registrationof complaint through email/ Toll free)	03/06/2024 (Toll free)
9	Action taken	Checked and found out the problem with Display and mainboard
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Nil




*Medical Officer
Primary Health Centre
Karikkattoor*

13	Asset Value	Rs . 1608/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Nil
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with Display and main board Machine installed in 12/10/2019 and covered up to 4+Year .Quotation not submitted since spare are not available in the market So recommending the machine for condemnation
17	Name & Signature of CYRIX Authority	Sreeju Sureesh 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
Asset value - 1608
cost of Spare - mentioned as spare not available..
life period - 4 year 8 months
As per the BER guidelines recommended as RBER


CONSULTANT BIOMEDICAL ENGINEER
NATIONAL HEALTH MISSION
Signature of JC BM (NHM)



Date

Seal


Signature of
Superintendent / Medical Officer (i/c)

Medical Officer
Primary Health Centre
Karikkattoor

WEIGHING MACHINE (ADULT - DIGITAL)

Sub
sheet

12/10/19 Supplied by Kaunya Medical Depot
Tissue (Subcentre kit) dt: 12/9/19.
Digital Weighing Machine
IN no: 2450 Rs: - 1608.04/- 1 1

10
11
12
13
14

12/10/19 Supplied by Kaunya Medical Depot
Tissue (Subcentre kit) dt: 12/9/19.
IN no: 2451 Rs: - 1608.04/- 1 2

12/10/19 Supplied by Kaunya Medical Depot
Tissue (Subcentre kit) dt: 12/9/19
IN no: 2452 Rs: - 1608.04/- 1 3

12/10/19 Issued. to LHI (Change JPHAN no) 3 NH

Sub sheet
Page no: 117

[Signature]
rel.



[Signature]
Medical Officer
Primary Health Centre
Karikkattoor



[Handwritten Signature]
Medical Officer
Primary Health Centre
Karikkattoor




Medical Officer
Primary Health Centre
Karikkattoor



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0541269




Medical Officer
Primary Health Centre
Karikkattor