


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA



Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	PHC KODANAD
3	Name of Equipment with Make, Model and Serial Number	Equipment: NEBULISER Make : DR.MOREPEN Model : CN-06 Sr.no : 18110769
4	Equipment ID & Barcode	114585 & 0740215
5	Date of purchase / Year of manufacture /Installation Date	28-03-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	01-06-24 (Toll Free)
9	Action taken	Checked the machine and found that compressor unit with piston are faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA



13	Asset Value	Rs.1353
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that compressor unit with piston are faulty. Equipment installed on 28-03-2019 and the machine outlived 5+years. Quotation not submitted since the spares are not available in the market..So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:
compressor and piston damage. unable to repair

 24/6/24
 Signature of JC BM (NHM)


Superintendent  Medical Officer (i/c) Signature of  Date

മെഡിക്കൽ ഓഫീസർ ഇൻചാർജ്
 എസ്.എച്ച്.സി. കോടമ്പാട്





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



Service Provider
No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 215303

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : Phc
Address : Phc Koclanad
Ezhaikulam, Kerala
Ph : 7306082160

Call Registration Date : 01/06/2024
Caller ID : 119585
Date of Visit : 03/06/2024
Asset No. : 0790215
EQPT Name : Nebulizer
Manufacture : Demoreper Model : CW06
S. No. 18110769 Dept. Inpatient room

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified :
nebuliser working but no fumes generated

Action Taken : checked Machine compressor working good checked tubing and connected compressor issue checked diaphragm (piston cup) faulty replaced. Not reparable. Identified spare not available. (Installation date: 23/03/2019)

Completed Date : 03/06/24 Time : 11:00 AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Compressor set for nebuliser	1		
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Yedhukrishna C	3/6/24	10:00 AM	11:00 AM

Customer Remark: Completed Pending

Service Engineer Name : Yedhukrishna C
Signature :
Date : 03/06/24
Contact Number : 9086147236

Customer Name :
Signature :
Date : 03/06/24
Contact Number : 6235367127
Designation : Nursing officer
Hospital Seal :

Nebulizer

29

9/8/16	Received from stone	1 No	1
	(Model: Readyset)		
17/2/16	Received from Stone (Truechall)	1 No	2-1 1
28/3/19	Received from Stone (Dr. Moxeped) (Model: CN-06)	1 No	2
7/1/2020	RFS	1	3

*June 19 1 item transferred to unseviceable
Register Page no: 27 Page 2*

