



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

**No. :** 184488

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>DH</u>	Call Registration Date : <u>03/06/2024</u>
Address <u>PERINTHALMANNA</u>	Caller ID : <u>114771</u>
<u>MALAPPURAM</u>	Date of Visit : <u>04/06/2024</u>
Ph : <u>9447928627</u>	Asset No. : <u>1010433</u>
	EQPT Name : <u>OPHTHALMOSCOPE</u>
	Manufacture <u>Heine</u> Model : <u>BETA 200</u>
	S. No. <u>NA</u> Dept: <u>SKOP</u>

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :**  
Head section not working

**Action Taken :** checked the machine, ophthalmoscope head defective. need to replace these spare for further checking.

Completed  Date : 4/6/24 Time : 12:00PM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

**Customer Remark**  
SUPERINTENDENT DISTRICT HOSPITAL PERINTHALMANNA-619322  
Dr. Consultant Biomedical Engineering Malappuram  
Completed  Pending


<b>Service Engineer Name:</b> <u>CIOKUL K. T</u> <b>Signature:</b> <u>[Signature]</u> <b>Date:</b> <u>4/6/24</u> <b>Contact Number:</b> <u>9809314280</u>	<b>Customer Name:</b> <u>Soumitra K. S</u> <b>Signature:</b> <u>[Signature]</u> <b>Date:</b> <u>4/6/24</u> <b>Contact Number:</b> <u>9447928627</u> <b>Designation:</b> <u>Senior optometrist</u> <b>Hospital Seal:</b> <u>[Seal]</u>
--	--

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	DH PERINTHALMANNA
3	Name of Equipment with Make, Model and Serial Number	EQUIPMENT: OPHTHALMOSCOPE Make: HEINE Model: BETA 200 Serial Number : NA
4	Equipment ID/ Barcode	#1010433 – 114771
5	Date of purchase/ Year of manufacture/Installation Date	23/03/1999
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of Purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03/06/2024
9	Action taken	Checked the machine Ophthalmoscope head defective. Need to replace These spare for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	NOT WORKING
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	Ophthalmoscope head:35840/-

13	Asset Value	26602/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	134%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report & OEM Quotation
16	Reasons for recommending the equipment as BER	Checked the machine and found Ophthalmoscope head defective. The equipment was installed on 23-03-1999 Aged up to 25 Years and 2 months. Repairing cost is 134% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	GOKUL KT 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Verified with the done complaint for the head part. Need to replace the same. Beyond economic repair



  
13/6/24

Signature of JC BM (NHM)

Jr. Consultant Biomedical Engineer  
NHM Malappuram

Date: 14/5/24



SUPERINTENDENT  
DISTRICT HOSPITAL  
PERINTHALMANNA-679 320  
MALAPPURAM DT.

Signature of   
Superintendent/Medical Officer (I/c)

# Vaansari Marketing Services

ISO 9001:2008 CERTIFIED CO



3-B, LAKSHMI STREET, THRUMANGALAM ROAD, VILLIVAKKAM, CHENNAI - 600049. INDIA

PHONE: 26181205, 26181324 FAX: 0091-44-26181419

E-MAIL: vaansarimarketing@gmail.com WEBSITE: www.vaansari.com

TO :		PROFORMA INVOICE/QUOTATION		
Sl No	Description	Qty.	Rate IN RS	Amount IN RS
M/s. CYRIX HEALTHCARE PVT LTD, MALAPURAM, KERALA. KIND ATTN: Mr. GOKUL - DISTRICT INCHARGE EMAIL: gokulcyrix@gmail.com MOBILE: +91 98093 14280		No.VMS/QTN/417/24-25/SI	DT: 10.06.2024	
		REF:RNK	GST NO. 33AADPN5542K1ZO	
1.	HEINE BETA 200 OPHTHALMOSCOPE HEAD ONLY  ADD:GST@12% HSNCODE:90189042  (RUPEES THIRTY FIVE THOUSAND EIGHT HUNDRED FORTY ONLY)	1 NO	32,000/-	32,000.00 <hr/> 32,000.00 <hr/> 3,840.00 <hr/> 35,840.00

### GENERAL TERMS

PAYMENTS: 100% IN ADVANCE TO BE IN OUR FAVOUR

### BANK DETAILS

ACCOUNT NAME	:	VAANSARI MARKETING SERVICES
BANK	:	INDIAN OVERSEAS BANK
BRANCH	:	KAMBAR COLONY BRANCH,
ACCOUNT NO	:	281102000000289
ADDRESS	:	65, THIRUMANGALAM ROAD, VILLIVAKKAM, CHENNAI - 600 049.
RTGS/NEFT CODE	:	IOBA0002811

DELIVERY: GENERALLY ITEMS ARE READY STOCK. IN CASE NOT READY 4 TO 8 WEEKS FROM THE DATE OF ORDER.

WARRANTEE: 1 YEAR AGAINST ANY MANUFACTURING DEFECT (BULB, OPTICAL SYSTEMS, PLASTIC COMPONENTS, KNOB AND INSTRUMENTS HAS NO GUARANTEE/ WARRANTEE)

VALIDITY: 30 DAYS

FOR VAANSARI MARKETING SERVICES



*B. M. D. C.*  
SUPERINTENDENT  
DISTRICT HOSPITAL  
PERINTHALMANNA-679 322  
MALAPPURAM DT.

*Dr. R. Naresh Kumar*  
(DR.R.NARESH KUMAR)



Jr. Consultant Biomedical Engineer  
NHM Malappuram

# Govt: Hospital, Perinthalmanna

## STOCK REGISTER OF \_\_\_\_\_

Name of Articles @Phthalmoscope

Date	No and date of Voucher or Invoice	From whom received of whom issued	Receipt	Issued	Balance	Initials of receiver	Remarks
------	-----------------------------------	-----------------------------------	---------	--------	---------	----------------------	---------

23-3-99

Govt Store 1

1

Receipt No 35

25.3.99

Issued to O.P

N/C

Not working


*L. V*  
*M. M. M. C*



**SUPERINTENDENT**  
 DISTRICT HOSPITAL  
 PERINTHALMANNA-679 322  
 MALAPPURAM DT.

*[Signature]*  
 13/02/24

**Dr. Consultant Biomedical Engineer**  
 NHM Malappuram





**CYRIX HEALTHCARE PVT LTD, MALAPPURAM, KERALA - BETA 200 OPHTHALMOSCOPE HEAD - 10.06.2024 - 417**

1 message

Mon, 10 Jun, 2024 at 12:39 pm

Vaansari <vaansarimarketing@gmail.com>  
To: Gokul K T <gokulcyrix@gmail.com>  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <Zm2.klbemp@cyrix.in>

Dear Sir/Madam,

Thank you for your enquiry .

Please find the enclosed quotation.

Awaiting for your valuable order .

Thanks & Regards

By

Vaansari

**From:** Gokul K T [mailto:gokulcyrix@gmail.com]  
**Sent:** 09 June 2024 18:05  
**To:** vaansarimarketing@gmail.com  
**Cc:** Sarang KM; Blesson Jose  
**Subject:** 1010433 OPTHAMALSCOPE Need Spare quotation

Dear Sir, Madam

Can you send the Price quotation for this Head, Lamp, Body & battey For Ophthalmoscope, make Heine, model BETA200  
Spare image attached below

GOMUL KT  
CYRIX HEALTHCARE PVT Ltd  
DISTRICT INCHARGE  
9809314280  
MALAPPURAM



*[Handwritten signature]*  
Jr. Consultant Biomedical Engineer  
NHM Malappuram

*[Handwritten signature]*  
SUPERINTENDENT  
DISTRICT HOSPITAL  
PERINTHALMANNA-679 322  
MALAPPURAM DT.



*Handwritten signature in green ink.*

**SUPERINTENDENT  
DISTRICT HOSPITAL  
PERINTHALMANNA-679 322  
MALAPPURAM DT.**



**Jr. Consultant Biomedical Engineer  
NHM Malappuram**