


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	DH NILAMBUR
3	Name of Equipment with Make, Model and Serial Number	Make: EURO DIAGNOSTIC SYSTEMS Model: ELECTRA Serial Number :A-7150 EQPT NAME - HAEMO MIXER
4	Equipment ID/ Barcode	#108888 - 1011305
5	Date of purchase/ Year of manufacture/Installation Date	10/07/2018
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of Purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	30/04/2024
9	Action taken	Checked the machine motor defective. Need to replace These spare for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	NOT WORKING
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	7620/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report & OEM mail communication attached
16	Reasons for recommending the equipment as BER	Checked the machine and found Motor defective. The equipment was installed on 10-7-2018 Aged up to 5 Years and 9 months. Oem informed through mail that company stopped the production and distribution of mixer and its spares, so spares are not available. As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	GOKUL KT 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Worked with the damage of Board and as per OBM, they stopped production of the same. Recommended for Beyond economic repair.


27/05/24
Signature of Consultant Biomedical Engineer
NHM Malappuram

Date: 29/5/24




Signature of Superintendent
SUPERINTENDENT
Govt. District Hospital
Nilamour - 679329
Malappuram Dist



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 221587

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility :	DH	Call Registration Date :	30/04/2024
Address :	NILAMBUR	Caller ID :	108888
.....	MALAPPURAM	Date of Visit :	01/05/2024
Ph :	9496467552	Asset No. :	1011305
		EQPT Name :	Haemo mixer
		Manufacture Euro Diagnostic Model :	Electra
		S. No.	A-7150 Dept. LAB

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

Machine not working.

Action Taken :

On checking found that machine motor is defective. So need to replace these spare for further checking.

Completed Date : 01/05/2024 Time : 10:10AM... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.	NA	NA	
3.			

Customer Remark:

Completed Pending

SUPERINTENDENT
Govt. District Hospital
Nilambur
Jr. Consultant Biomedical Engineer
Malappuram Dist

Dorshur

Service Engineer Name :	Customer Name :
Signature :	Signature :
Date : 01/05/2024	Date :
Contact Number : 7593847152	Contact Number : 9496467552
	Designation : Senior grade CT
	Hospital Seal :





REQUIRED SPARE QUOTATION #108888- 1011305,#110663- 1031479 BLOOD MIXER

2 messages

[Handwritten Signature]
SUPERINTENDENT
Govt. District Hospital
Nilambur - 679329
Malappuram Dist

Sarang KM <sarangcyrix@gmail.com>
To: anandhu@peerlessbio.com
Cc: Gokul K T <gokulcyrix@gmail.com>, BLESSEN JOSE <zm2.klbemp@cyrix.in>

Sat, 18 May, 2024 at 4:27 pm

Dear Sir,
There is a complaint reported from DH NILAMBUR, CHC Chungathara , Regarding the complaint with BLOOD MIXER.
Required quotation for motor.

EQPT NAME- BLOOD MIXER (HAEMO MIXER)
MAKE- EURO Diagnostics systems
MODEL- N/A
SL NO- A7150
COMPLAINT- Motor complaint



Thanks & Regards

Sarang.k.m
Divisional Incharge|North Zone
Bemp-Kerala
Mobile : +91 7593847150
Email : sarangcyrix@gmail.com

[Handwritten Signature]
27/5/24

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

Jr. Consultant Biomedical Engineer
NHM Malappuram

P Save Environment, Please do not print this e-mail unless it's essential

anandhu <anandhu@peerlessbio.com>
To: sarangcyrix@gmail.com
Cc: gokulcyrix@gmail.com, zm2.klbemp@cyrix.in

Sat, 18 May, 2024 at 4:27 pm

Sir,
SL no A7150 blood Mixer is an old version,the company stopped the production and distribution of the mixer and it's spare, so the defected spare(motor) is currently unavailable.

Regards,
Anandhu pv
Peerless biotech Pvt Ltd
Kannur
Service engineer

— On Sat, 18 May 2024 16:23:13 +0530 sarangcyrix@gmail.com wrote —
[Quoted text hidden]



Govt. District Hospital
Nilambur - 679329
Malappuram Dist.


Diagnostic Systems


 KTPL Kirtleskar Group	BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BMMP)	 KMSCL
	Toll FREE- 1800-2700-690	
KTPL BARCODE - (8907579) - 80102699		

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE (8004-890615225) 1011305


SAMPLE MIXER(BLOOD MIXER)
DH/NBR/LAB/EQUIPMENT Reg No:3/72/02

Jr. Consultant Biomedical Engineer
NHM Malappuram





SUPERINTENDENT
Govt. District Hospital
Nilambur - 679329
Malappuram Dist




27/5/24

Jr. Consultant Biomedical Engineer
MHM Malappuram

Installed on 11/07/2018



M/C SR. NO.
A -
7150

BLOOD MIXER

€LECTRA

Marketed by
Euro Diagnostic Systems Pvt. Ltd.
36/37 Millennium Pocher, M. K. Srinivasan Nagar Main Road
Rajiv Gandhi Salai, Perungudi, Chennai-600 096, Tamil Nadu
India. Phone: +91 44-2496 0541, Fax: +91 44-2496 2620
e-mail: eurodiag@vsnl.net, web: www.eurodiagnostics.in

(Signature)
Jr. Consultant Biomedical Engineer
NHM Malappuram



(Signature)
SUPERINTENDING
Officer
Govt. District Hospital
Nilambur - 679329
Malappuram Dist