


Order pending

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Palakkad
2	Name of Hospital	DH Palakkad
3	Name of Equipment with Make, Model and Serial Number	Haematology Analyser Sysmex , XS-800i SL. 62880
4	Equipment ID & Barcode	0911197
5	Date of purchase / Year of manufacture /Installation Date	20.08.2009
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No Amc / CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	7.11.2023
9	Action taken	checked and found that equipment is 14 year's above ageing and it is not technical repair.
10	Present status of the equipment (Fully damaged / partially damaged)	fully damaged
11	Recommendations for repair (required service details)	Not Re commending
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	95,000
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	nil
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix and Transasia report attached.
16	Reasons for recommending the equipment as BER	checked and found that equipment is aged up to 14 year's above comparing life span of equipment used. the repair of the same is not cost effective and is beyond economic repair. So we recommended - condemnation
17	Name & Signature of CYRIX Authority	Dileep. 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Recommended for BER.






Signature of JC BM (NHM)



Date

Seal


Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad

Signature of
Superintendent / Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 165180

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>DH Palakkad</u>	Call Registration Date : <u>7.11.2023</u>
Address	Caller ID : <u>78931</u>
Ph : <u>9946373803</u>	Date of Visit : <u>07.11.2023</u>
	Asset No. : <u>091197</u>
	EQPT Name : <u>Haematology Analyser</u>
	Manufacture <u>Sys.mex</u> Model : <u>XS-800i</u>
	S. No. <u>62880</u> Dept. <u>Lab</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found that equipment is installed on 20.08.2009 and aged up to 14 years above. Comparing life span of equipment used. The repair of the same is not cost effective and is beyond economic repair. So we recommended for condemnation.

Completed Date : 28.11.23 Time : 2.06pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>nil</u>			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Dileep A.D</u>	<u>28.11.23</u>	<u>1.00pm</u>	<u>2.06pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Dileep A.D</u>	Customer Name :
Signature : <u>[Signature]</u>	Signature :
Date : <u>28.11.2023</u>	Date :
Contact Number : <u>9048751999</u>	Contact Number :
	Designation :
	Hospital Seal :


Vayalasee
 Deputy Director of Health Services
 & Superintendent,
 District Hospital, Palakkad



Date: 18-11-2023

To

The Superintendent
District Hospital
RDC Lab
Palakkad

Sub: - CONDEMNATION CERTIFICATE

Dear Sir,

We hereby certify that following supplied instrument has consumed its shelf life. The repair of the same is not cost effective and is beyond economic repair.

S.NO	Instrument Model	Sr. No.	Installation date
1	XS 800i Hematology Analyzer	62880	20-08-2009

We therefore, strongly recommend the Condemnation of the same and propose for it's replacement.

Thanking you.

Yours truly,
For Transasia Bio-Medicals Ltd.


S SURESH M S
ZONAL MANAGER





EQUIRED SERVICE SUPPORT 0911197 # 78931 HEMATOLOGY ANALYZER @ DH PALAKKAD

Messages

Wed, 26 Jun, 2024 at 11:18 a

chilesh P <akhileshcyrix@gmail.com>
): SANDEEP DAS M <s.das@transasia.co.in>
): Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir

There is a complaint reported from DH PALAKKAD ,palakkad district we are requesting you to do one time visit for rectification.please attend the complaint and revert back with service report after rectification

Hospital name: DH PALAKKAD

District: palakkad

Equipment name: HEMATOLOGY ANALYZER

Make : SYSMAX

MODEL : XS 800I

Serial no: 62880

Barcode: 0911197

Complaint I'd: 68931

Complaint reported: booting issue

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

chilesh P <akhileshcyrix@gmail.com>
): s.das@transasia.co.in
): Zm2.klbemp@cyrix.in, dm1.klbemp@cyrix.in

Tue, 9 Jul, 2024 at 3:39 p

Dear sir

Kindly update the status

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Tue, 9 Jul, 2024 at 3:55 p

SANDEEP DAS M <s.das@transasia.co.in>

>: Akhilesh P <akhileshcyrix@gmail.com>

>: Zm2.klbemp@cyrix.in <Zm2.klbemp@cyrix.in>, dm1.klbemp@cyrix.in <dm1.klbemp@cyrix.in>

Dear Akhilesh ,

Please find the attachment and do the needfull .

Get Outlook for Android

From: Akhilesh P <akhileshcyrix@gmail.com>

Sent: Tuesday, July 9, 2024 3:39:57 PM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: Zm2.klbemp@cyrix.in <Zm2.klbemp@cyrix.in>; dm1.klbemp@cyrix.in <dm1.klbemp@cyrix.in>

Subject: Re: REQUIRED SERVICE SUPPORT 0911197 # 78931 HEMATOLOGY ANALYZER @ DH PALAKKAD

[Quoted text hidden]

5PART HEMATOLOGY ANALYSER

Arinistee
Palakkad

26.8.09	Store Sydt. D.H. P.K.D	1 UNIT	1 NO	
1.	X.S - 8001 AUTOMATED 5PART HEMATOLOGY ANALYSER			
	SYSMEX - 62880			
2.	BLOOD MIXTURE (1307 Trans)	1 NO	1 NO	Attach to Page No-3
3	PC AD-360V OPTIPLEX + LANCARD	1 NO	1 NO	
4	PRINTER (14 HP)	1 NO	1 NO	
	<u>Free Pack</u>			
5.	CELL PACK (X9018, 01/10 20LTR Transasia)	1 NO	1 NO	
6	STROMATOLYSER, 4DS (42 ML)	1 NO	1 NO	
7.	STROMATOLYSER 4DL (X 9009 Transasia) 5LTR	1 NO	1 NO	
8 -	SULPHOLTSEK (Transasia) (500ml)	1 NO	1 NO	
	Rs. 12,01,460.00 + TAX 48058.40 + 480.58 Total Rs. 12,49,998.98.			
1.9.09	Used in Lab	1	NIL	

JEW



Vayasree
Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad

Erinjilal
Page No: 278