


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	CHC VALAYAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : Colorimeter Make: AIMIL LIMITED Model: PHOTOCHEM SN: 19497
4	Equipment ID & Barcode	113076 & 1130857
5	Date of purchase / Year of manufacture /Installation Date	06-01-2013 (Actual installation date is 27-08-2012. stock book attached for reference)
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24-05-2024
9	Action taken	Checked and found Main board, photo diode, filter assembly and display defective. Enquired spares with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	6472/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached and vendor EOL Letter attached
16	Reasons for recommending the equipment as BER	Checked and found Main board, photo diode, filter assembly and display defective. In software installation date is 06-01-2013, As per the stock book machine is installed on 27-08-2012 Equipment out lived 11 years 9 month. Oem informed that this model is obsolete Spare not available .As per tender clause 5.3.14.2 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHIL SURESH S 

\*Not mandatory

*#Based on the period of life and value as per the BER guidelines*

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that main board, photo diode, filter assembly and display defective Machine is installed on 27/08/2012. OEM informed that this model is obsolete & spares are not available. So recommended for condemnation

Consultant: Bio Medical Engineer  
National Health Mission  
Kozhikode



Recommended,

  
MEDICAL OFFICER  
CHC VALAYAM  
VALAYAM (P.O), KOZHIKODE - 673 517

Date 25/6/24

Seal

Superintendent / Medical Officer (i/c)



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. : 192571**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility ..... <u>CHC Valayam</u></p> <p>Address ..... <u>Valayam</u> <u>Kozhikode</u></p> <p>Ph : ..... <u>9633944718</u></p>	<p>Call Registration Date : ..... <u>24/5/24</u></p> <p>Caller ID : ..... <u>113076</u></p> <p>Date of Visit : ..... <u>27/5/24</u></p> <p>Asset No. : ..... <u>1130857</u></p> <p>EQPT Name : ..... <u>Colorimeter</u></p> <p>Manufacture ..... <u>Aimil</u> Model : <u>Photochem</u></p> <p>S. No. ..... <u>19497</u> Dept. ..... <u>Laboratory</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** ..... Not getting on

**Action Taken :** ..... Checked and found mainboard, photodiode, filter assembly and display defective need to replace those spares for further checking of machine working condition

Completed  Date : ..... 27/5/24 Time : ..... 4pm Spare Required


Spare Replaced  Requested

No.	Description	Qty.	Part Number	PR Number
1.	<u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
2.	<u>na</u>	<u>na</u>	<u>na</u>	<u>na</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Abhil Suresh S</u>	<u>27/5/24</u>	<u>3pm</u>	<u>4pm</u>

**Customer Remark** .....  Completed  Pending

<p>Service Engineer Name : <u>Abhil Suresh S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/5/24</u></p> <p>Contact Number : <u>9995706213</u></p>	<p>Customer Name : <u>Rajita</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>27/5/24</u></p> <p>Contact Number : <u>9539624378</u></p> <p>Designation : <u>LT</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	---



**MEDICAL OFFICER**  
**CHC VALAYAM**  
 VALAYAM (P.O.) KOZHIKODE  
[Signature]  
**MEDICAL OFFICER**  
**CHC VALAYAM**  
 VALAYAM (P.O.) KOZHIKODE



**Required Spare Quotation -1130857- 113076-Colorimeter -CHC VALAYAM- KOZHIKODE,Kerala**

2 messages

Thu, Jun 6, 2024 at 1:30 PM

Jineesh P P <jineeshcyrix@gmail.com>  
To: deepaksharma@aimil.com  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir

Please share price and spare availability of Main board,photo diode, filter assembly and display

Colorimeter  
Make : AIMIL LIMITED  
Model: PHOTO CHEM  
SN: 19497

Here I am attaching photos for your reference

Thanks & Regards,

Jineesh PP  
District Incharge - Kozhikode  
Mob-7593847163  
jineeshcyrix@gmail.com  
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Thu, Jun 6, 2024 at 1:39 PM

Deepak Sharma <deepaksharma@aimil.com>  
To: Jineesh P P <jineeshcyrix@gmail.com>  
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir,

Its an obsolete model so spares not available now.

Suggest customer to but lates unit- catalogue arttached.



**Deepak Sharma**  
Deputy Manager – PEI Group



**Sales & Service**  
off: +91 11 61310210

mob: +91 9560889051

A-8 Mohan Co-Operative Industrial Estate  
New Delhi 110044

web: [www.aimil.com](http://www.aimil.com)

**From:** Jineesh P P <[jineeshcyrix@gmail.com](mailto:jineeshcyrix@gmail.com)>  
**Sent:** Thursday, June 6, 2024 1:31 PM  
**To:** [deepaksharma@aimil.com](mailto:deepaksharma@aimil.com)  
**Cc:** Sarang KM <[sarangcyrix@gmail.com](mailto:sarangcyrix@gmail.com)>; Blesson Jose <[zm2.klbemp@cyrix.in](mailto:zm2.klbemp@cyrix.in)>  
**Subject:** [SUSPICIOUS:] Required Spare Quotation -1130857- 113076-Colorimeter -CHC VALAYAM-KOZHIKODE,Kerala


**CAUTION:** This email is from an external source. Exercise caution when opening attachments or clicking links.


[Quoted text hidden]




**FOR : AIMIL LIMITED, NEW DELHI**  
**BY : PHOTO ELECTRIC INSTRU. PVT. LTD.**  
 F-138, M.I.A., 2nd Phase, BASNI, JODHPUR

**CAT.NO.**      **SR. NO.**  
018      19497


**CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)**  
**TOLL FREE NO: 1800 - 425-7669**  
**BAR CODE - (8004-890615225) 1130857**





 PIN: 873 517  
 OFFICE OF THE CHC VALAYAM • WAYALAM (P.O.)



Photo copy returned  
**MEDICAL OFFICER**  
**CHC VALAYAM**  
 VALAYAM (P.O.), KOZHIKODE - 673 517  
 25/6/24