


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	DH Mananthavady
3	Name of Equipment with Make, Model and Serial Number	Equipment-3 kva UPS Make-Consul neowatt Model- FML11X0003336E18 Sn-C1202017141498
4	Equipment ID/ Barcode	1250152//TICKET 109612
5	Date of purchase/ Year of manufacture/Installation Date	30/09/2020
6	Warranty details (Yes/No)	NO warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	04/05/2024-TOLL FREE
9	Action taken	Checked and found Main board and battery(8nos) defective, Need to replace these spare for further checking. Enquired to supplier for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	Cosul neowatt 3KVA/96V board-20000/- 7Ah battery(8nos)-9600/- Total-29600/-

13	Asset Value	4500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	657%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from local vendor (Hybrid solar kalpetta) Cyrix service report
16	Reasons for recommending the equipment as BER	Checked and found main board (Cosul neowatt 3KVA/96V board) and 8 battery (12v7Ah) defective. Equipment installed on 30.09.2020 aged up to 3 years 8 months. Repairing cost is 657% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 


*Not mandatory

*Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Main board and battery are defective - Repair cost more than 50% of actual cost - Recommended condemnation


Signature of JC BM (NHM)

Date:




6/7/24
Signature of Superintendent
SUPERINTENDENT
DISTRICT HOSPITAL
WAYANAD MANANTHAVADY
Officer (I/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 210714

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility <u>DH</u></p> <p>Address <u>DH Mananthavady</u> <u>Wayanad</u></p> <p>Ph : <u>9947918269</u></p>	<p>Call Registration Date : <u>04/05/24</u></p> <p>Caller ID : <u>109612</u></p> <p>Date of Visit : <u>06/05/24</u></p> <p>Asset No. : <u>1250152</u></p> <p>EQPT Name : <u>UPS 3KVA</u></p> <p>Manufacture/Consult/Neowall/Model : <u>FML11x00</u> <u>3336E18</u></p> <p>S. No. <u>C12020174</u> Dept. <u>Radiology</u> <u>141498</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : No output

Action Taken : Checked and found that the Main board and batteries are defective. Needs to replace the same.

Completed Date : 06/05/24 Time : 1:00 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrrix Engineer

Date

Start Time

End Time

Ameer.C.H

06/05/24

9:30AM

1:00PM

Customer Remark

Completed

Pending



**SUPERINTENDENT
DISTRICT HOSPITAL
WAYANAD, MANANTHAVADY**

Service Engineer Name : Ameer.C.H

Signature : Ameer.C.H

Date : 06/05/24

Contact Number : 7994374680.

Customer Name : Wayanad Kerala - OT

Signature : [Signature]

Date : 6/5/24

Contact Number : 9947918269

Designation : radiology

Hospital Seal :

**HYBRID
SOLAR**

Near KFWW Board, Kalpetta Wayanad
Mob: 8089386389, 9745772211

HS/147/23-24
20/06/2024

TO
THE CYRIX HEALTH CARE
KOCHI

Attn: Sir,
Ref :

SUB: -

Ref: - As Per Your Enquiry Notice No:

We are pleased to quote our UPS & Battery Service Price as below

SL.No	Description	Rate	Qty	Amount
1	CONSUL NEOWATT 3KVA/96V BOARD REPLACEMENT	20000.00	1	20000.00
2	7AH BATTERY REPLACEMENT	1200.00	8	9600.00
	INCLUDING SERVICE CHARGE			
TOTAL				29600.00

Yours truly
For Hybrid Solar
Kalpetta
Wayanad

E-mail: hybridsolar8911@gmail.com





request for service quotation 3 KVA ups at dh mananthavady,Wayanad (Barcode-1250152,ticket-109612)

2 messages

Niyas Hadi AP <niyascyrix@gmail.com>

Sat, 15 Jun 2024 at 12:59 pm

To: hybridsolar8911@gmail.com <hybridsolar8911@gmail.com>

dear sir
please send the quotation for the service of 3kva ups in dh mananthavady,wayanad
equipment -3kva ups
make- consul neowatt

Thanks & regards

NIYAS HADI
District incharge
BEMP project
Wayanad
kerala
mob -7593847171

email -niyascyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Pooirithura | Cochin -38 | Kerala | India

Anoop PK <hybridsolar8911@gmail.com>

Thu, 20 Jun 2024 at 10:56 am

To: niyascyrix@gmail.com

Dear sir please find attached quotation
Thanks Hybrid solar Kalpetta

Ph. 9745772211

[Quoted text hidden]


HOSPITAL DEVELOPMENT COMMITTEE DISTRICT HOSPITAL MANANTHAVADY STOCK REGISTER

104

Name of Article: *Computed Radiography System with Accessories Model 8*

Invoice No Date	Received From / Issued To	Received	Issued	Balance
	<i>system Received from the store with Accessories</i>			
	<u>ACCESSORIES</u>			
1)	DICOM 2.0 Work Station Software	1		
2)	2 kW Digital online UPS system with AVR	1		
3	Fuji Imaging plate 10-675x442	10		
4	10 Cassette FP CR COMPASS 442 <small>CR 110</small>	10		
5	Medical Dry film 20 x 25 150SH	4		
6	Medical Dry film 21-44 35x43 100SH	2		
7	FCR Capsule XL 2, CR 1234RU	1		
8	Medical dry film 25x30 150SH	2		
9	DRY PIX EDGE 2T684G	1		
10	600 800V5 LT SC 2019 MS	1		
11	CONSOLE ADVANCE 8000 <small>KIT 2000</small>	1		
12	DLINK 8 PORT Switch	1		
13	EATG patch CO2 3TS	1		
14	Barcode Scanner OD4130	1		
15	14 P 600V5 Prodesk SFF	1		
16	14 10" INCH Screen	1		
	<i>Make - Fuji film Model - Capsule XL II DRY PIX EDGE Serial NO - 0630043 0632483</i>			
	<i>Purchase order NO - B1-7664/2020 CR01 Purchase date - 7-08-2020 Warranty - 29/9/2020 to 29/9/2023 PS 1049942</i>			

Tony...
...



Superintendent
SUPERINTENDENT
DISTRICT HOSPITAL
WAYANAD, MANANTHAVADY



SUPERINTENDENT
 DISTRICT HOSPITAL
 WAYANAD, MANANTHAVADY

RA
 Remb 07

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

CYRIX
REGULATED BY THE
TOLL FREE NO: 1800 - 425-1889

BAR CODE - (8004-8906*5225)
 1250152



SERIAL NO : C1202017141498

ITEM CODE : FMI11N0003336E18

DESC: UPS 3KVA, FIN, 96V, 7AH, CON, SMF, 8, IN BUILT

Fuji Electric Consul Newwatt Pvt. Ltd
 No : 119, 120, 120 A, Electrical & Electronics Ind Estate,
 Perungudi, Chennai 96.



CAUTION:

- For operation read user manual including safety warnings first!
- This unit may be opened by authorized technicians only!
- Even when switched off there is a hazardous voltage on the battery side!
- Lead acid battery in the inside of the enclosure!



Handwritten signature
 SUPERINTENDENT
 DISTRICT HOSPITAL
 WAYANAD, MANANTHAVADY

Handwritten signature
 Remle P.T