



**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	PHC KOLLENGODE
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS MAKE : OMRON MODEL : HEM-8712 SN : 20170813225V1
4	Equipment ID/ Barcode	0940312 # 116485
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil (As per stock book installation date - 22/10/2019, stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	11/06/2024 (toll free)
9	Action taken	Checked the machine found not getting on. Problem with board and Display . Need to replace board and Display for further checking.
10	Present status of the equipment (Fullydamaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA




13	Asset Value	RS.2200/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine found not getting on. Problem with main board and display. Equipment installation date as per stock book 22/10/2019. aged upto 4 year 7 months. Quotation not submitted since spare parts are not available on market. So we recommending the equipment for condemnation .
17	Name & Signature of CYRIX Authority	AKHILESH P 

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Main board and display faulty. Machine aged up to 4 years. Quotation not submitted.


 Signature of Junior Consultant (Biomedical) NHM (Aranyakal)

Date: _____

02/07/2024. 
 Signature of
 Superintendent/Medical Officer (Uc)



Medical Officer
 Family Health Centre
 Kollengode, Palakkad



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 207361

ISO 13485 : 2012 & ISO 9001:2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC KOLLENGODE</u></p> <p>Address <u>KOLLENGODE</u> <u>PALAKKAD</u></p> <p>Ph : <u>9745210676</u></p>	<p>Call Registration Date : <u>11/06/2024</u></p> <p>Caller ID : <u>116485</u></p> <p>Date of Visit : <u>13/06/2024</u></p> <p>Assot No. : <u>0940312</u></p> <p>EQPT Name : <u>BP APPARATUS</u></p> <p>Manufacture <u>OMRON</u> Model : <u>HEM-8712</u></p> <p>S. No. <u>2017081322SV/Dept.</u> <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching ON

Action Taken : Checked the machine. Found that machine not switching ON. Problem with board and display. Need to replace board and display for further checking

Completed Date : 13/06/24 Time : 12:30PM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>DWINIS</u>	<u>13/06/24</u>	<u>11:30AM</u>	<u>12:30PM</u>

Customer Remark Completed Pending

Service Engineer Name : DWINIS **Customer Name :** Dr. BOBY MARIM
Signature : [Signature] **Signature :** [Signature]
Date : 13/06/24 **Date :** 13/06/2024
Contact Number : 7593847155 **Contact Number :** [Number]



Medical Officer
Family Health Centre
Kollengode, Palakkad



B

Medical Officer
Family Health Centre
Kollengode, Palakkad

Brought forwarded from Nurse page: 128
 18 B.P Apparatus LCD. (Spend or (20.6.20))

29/12/18
 30/12/18

Received From Store
 Issued to ward
 Pre-check up Room

MS No	Qty	Issued	Balance	Page
MS No	2	2	2	128
MS No	2	2	2	128
MS No	1	1	3	128
MS No	1	1	3	128

BP Apparatus (Digital)
 Rtd from Store (Nurse page: 142)
 Issued to ward
 Received from Store
 Issued to ward

Brought forwarded from Nurse page: 131

MS No	2	2	2	131
MS No	2	2	2	131
MS No	1	1	3	131

Precheck (1)
 Injection room (1)
 Steel W/room (1)
 (for out reach)



Medical Officer
 Family Health Centre
 Kollengode, Palakkad