


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	THQH KOOTHUPARAMBA
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make: DISHA Model: DIAMOND SN: NA
4	Equipment ID/ Barcode	1321613 #104836
5	Date of purchase/ Year of manufacture/Installation Date	27-07-2013 (Actual installation date is 20.12.2018 Stock Book Copy attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 04-04-2024
9	Action taken	Checked the Equipment found Equipment is not Working, found that Mercury Empty and Mercury Glass tube also damaged . Enquired for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	MERCURY 5ML- 1,398.31/- GLASS TUBE- 84.75/- GST(18%)- 266.95/- TOTAL- 1,750/-



13	Asset Value	1323/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	132.275%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not Working, found that Mercury Empty and Mercury Glass tube also damaged . Enquired for quotation. In software Equipment installed on 27-07-2013.As per the stock book installation date is 20.12.2018 aged up to 5 years 6 months. Repairing cost is 132.275% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that the machine not working due to mercury glass tube damaged . Aged up to 5 years above . its service cost economically not feasible . So recommended for BER ~~per~~ BER Protocol.

Signature of JC BM (NHM)

  
27/07/2024  
GOVT. TALUK HEADQUARTERS MEDICAL  
NATIONAL HEALTH MISSION  
KANNUR - 670002

Date: 22/05/2024



Signature of   
Superintendent/Medical Officer (i/c)

**SUPERINTENDENT**  
**GOVT. TALUK HEADQUARTERS**  
**HOSPITAL**  
**KUTHUPARAMBA, PIN - 670 643**





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. :** 217002

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p><b>Health Facility</b> THQH KOTHUPARAMBA</p> <p><b>Address</b> KOTHUPARAMBA KANNUR</p> <p><b>Ph :</b> 09562847869</p>	<p><b>Call Registration Date :</b> 04/04/2024</p> <p><b>Caller ID :</b> 104836</p> <p><b>Date of Visit :</b> 06/04/2024</p> <p><b>Asset No. :</b> 1321613</p> <p><b>EQPT Name :</b> BP APPARATUS</p> <p><b>Manufacture</b> DISHA <b>Model :</b> DIAMOND</p> <p><b>S. No.</b> NA <b>Dept.</b> CASUALTY</p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Equipment is not switching ON.

**Action Taken :** Checked the equipment found equipment is not switching ON. Found that Mercury empty and Mercury glass tube also damaged. Enquired for quotation.

**Completed**  **Date :** 06/04/2024 **Time :** 1:30 pm **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Shyamkrishnan	06/04/2024	1:00 pm	1:30 pm

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> Shyamkrishnan</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 06/04/2024</p> <p><b>Contact Number :</b> 8281188514</p>	<p><b>Customer Name :</b> </p> <p><b>Signature :</b> </p> <p><b>Date :</b> 6/04/24</p> <p><b>Contact Number :</b> </p> <p><b>Designation :</b> Senior Asst. Officer</p> <p><b>Hospital Seal :</b> </p>
--	--

## AR MEDICARE

ROOM NO: 865, 2<sup>ND</sup> FLOOR  
AL-RAHA ARCADE, THAZHE CHOVVA  
KANNUR-670018  
EMAIL: [armedicare07@gmail.com](mailto:armedicare07@gmail.com)  
MOB: 9446057350, 7994654108,



### QUOTATION

REF: AR/QTN-25/24-25

Date: 26-04-2024

To,  
Cyrix Health Care Private Limited  
30/641B, Vandipetta Rd, near SBI Bank, Poonithura, Maradu  
Ernakulam, - 682038  
Kerala 682038

No.	Description	Qty	Unit Price(INR)	Total(INR)
1.	BP APPARATUS ACCESSORIES MERCURY- 5ML	1	1,398.31	1,398.31
2	GLASS TUBE	1	84.75	84.75
			<b>Grand Total</b>	<b>1,483.06 INR</b>
			<b>GST(18%)</b>	<b>266.95 INR</b>
			<b>Net Total</b>	<b>1,750.00 INR</b>
<b>INR One Thousand Seven Hundred And Fifty Only</b>				

### Terms and Conditions for sale :

- Validity:** One month from the date of quotation.
- Payment:** Within 30 days from the date of supply.
- Delivery:** Within 7 days after receipt of the signed purchase order.
- If any change in Govt. statutory charges like Customs duty, Excise duty, GST etc. will be applicable at the time of invoicing.
- Bank Details:  
Account holder's name : AR MEDICARE  
Bank Name : State Bank Of India, Thazhechovva  
Account No. : 39648455024  
IFSC code : SBIN0016861



For AR MEDICARE,  
KANNUR





Sarang KM <arangcyrix@gmail.com>

**REQUIRED QUOTATION BP APPARATUS PARTS 1321613 #104836 @ THQH KOOHUPARAMBA ,  
KANNUR**

2 messages

Sarang KM <arangcyrix@gmail.com>  
To: ar medicare <armedicare07@gmail.com>  
Cc: avinash t <avinashcyrix@gmail.com>

Tue, Jun 18, 2024 at 10:56 AM

Dear sir,  
Please share the price and stock availability of

5 ML - MERCURY  
GLASS TUBE

Kindly do the needful asp

—  
Thanks & Regards

**Sarang.k.m**  
Divisional Incharge|North Zone  
Bemp-Kerala  
Mobile : +91 7593847150  
Email : arangcyrix@gmail.com

**CYRIX HEALTHCARE (P) LTD**  
30/641B | Petta | Punithura | Kochi | Kerala- 682038  
www.cyrixhealthcare.com



Save Environment, Please do not print this e-mail unless it's essential

ar medicare <armedicare07@gmail.com>  
To: Sarang KM <arangcyrix@gmail.com>

Tue, Jun 18, 2024 at 1:29 PM

Sir,  
PFA




**AR MEDICARE**

Thanks & Regards  
**AR MEDICARE**  
Room No:865,2nd floor  
AL-RAHA Arcade  
Thazhe Chowva, Kannur  
Mob: 9446057350, 7994654108, 8138091108  
Email: armedicare07@gmail.com  
GSTIN :32BDYPV5030P1ZP

[Quoted text hidden]

**ARQ25. CYRIX BP APPARATUS ACCESSORIES (1) (1).pdf**  
536K





SUPERINTENDENT  
GOVT. TALUK HEADQUARTERS  
HOSPITAL  
KUTHUPARAMBA, PIN - 670 643


*Low  
Selfy 01*



B.P. APPARATUS

STOCK REGISTER

Date	Particulars	Received from Store	Issued	Balance	Signature
11/3/2020	Tr. in from old book vol (1) Page No. 5 67131921 (1321613)			01	<i>[Signature]</i>
20/12/18	Rd from the store (LP) Rs 205/- put in use		1		<i>[Signature]</i>
25/12/2023	Tr. out to Unserviceable Registers		1		<i>[Signature]</i>



*[Handwritten signature]*  
July 20 0 00

**SUPERINTENDENT**  
**HOSPITAL**  
SOVT. TALUK HEADQUARTERS  
KUTHUPARAMBA, PIN - 670 60