


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC ALAPAD
3	Name of Equipment with Make, Model and Serial Number	Refrigerator Make - LG Model - NA S/N - NA
4	Equipment ID/ Barcode	0831273
5	Date of purchase/ Year of manufacture/Installation Date	28/05/2010
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	18/1/24
9	Action taken	Checked the equipment refrigerator not cooling and found refrigerator door damaged, Compressor not working, Enquired for Spare quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	Refrigerator door -7,450/- Compressor - 4,500 Timer - 850 Thermostat - 850 TOTAL - 13,650/-

13	Asset Value	19,900/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and Quotation attached
16	Reasons for recommending the equipment as BER	Checked equipment found Compressor and Door, Thermostat, timer damaged, equipment installed on 28-05-2010 and aged up to 13+ years, Spare cost is 68.5% of asset value and both criteria met for RBER. As per tender clause 5.3.14.1 we are recommending equipment for condemnation.
17	Signature of CYRIX Authority	Absil Tony 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spare cost very high


Consultant (Biomedical)
NHM, Thrissur


Signature of JC BM (NHM)



Date


SUPERINTENDENT
Community Health Centre
Alappad, Thrissur
Superintendent/Medical Officer (I/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 194816

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>CHC</u>	Call Registration Date : <u>18/01/2024</u>
Address <u>Alappad</u>	Caller ID : <u>91478</u>
<u>Thrissur</u>	Date of Visit : <u>20/01/2024</u>
Ph : <u>9142909466</u>	Asset No. : <u>0831273</u>
	EQPT Name : <u>Refrigerator</u>
	Manufacture <u>L.G</u> Model : <u>LG195</u>
	S. No. <u>005NLH042</u> Dept. <u>ICTC Lab</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Refrigerator not working

Action Taken : checked the equipment Refrigerator not cooling and found refrigerator door damaged. compressor not working required for spare part

Completed Date : 20/01/2024 Time : 3:pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Absil Tony</u>	<u>20-01-24</u>	<u>2:pm</u>	<u>3:pm</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Absil Tony</u> Signature : Date : <u>20/01/2024</u> Contact Number : <u>9795 328470</u>	Customer Name : Signature : Date : Contact Number : Community Health Centre Designation : Alapad, Thrissur Hospital Seal :
---	---



Cool House

Deto
Opposite SBI BANK, Kodakara Thrissur
☎ 9895171425 ✉ coolhouse@gmail.com

Quotation

To,
Cyrix Healthcare Pvt Ltd
Kochi, Ernaakulam
☎ 9847299500

Quotation# Quote-256

Date: 04-03-2024

Dear Sir/Mam,

Thank you for your valuable inquiry. We are pleased to quote as below:

#	DESCRIPTION	QTY	PRICE	TOTAL
1	Refrigerator Door	1 Nos	₹ 7,450.00	₹ 7,450.00
2	Compressor	1 Nos	₹ 4,500.00	₹ 4,500.00
3	Thermostat	1 Nos	₹ 850.00	₹ 850.00
4	Timer	1 Nos	₹ 850.00	₹ 850.00
GRAND TOTAL				₹ 13,650.00

We hope you find our offer to be in line with your requirement.

Terms & Conditions:

- 50% Advance payment


For, COOL HOUSE



AUTHORIZED SIGNATURE



LG



**5 MORE STARS
MORE SAVINGS**

**POWER SAVINGS
GUIDE**

ELECTRICITY CONSUMPTION
244*
UNITS PER YEAR

Appliance	: Refrigerator
Brand	: LG
Model/Year	: GL-195NH5/2010
Type	: Direct Cool
Gross volume	: 185 Liter
Storage volume	: 175 Liter

**ENERGY IS LIFE
CONSERVE IT**

*Under test conditions, when tested in accordance with relevant standards. Actual electricity consumption will depend on how the appliance is used.



No.1
in ICE MAKING
Patented Ice Tray for Fastest Ice**
No extra load on compressor.

**Works Without
Stabilizer**





CYRIX
[Handwritten text and QR code]

COMMUNITY HEALTH CENTRE
ALAPPAD
PIN - 680 641
Date.....
THIRSSUR DISTRICT



OL - 195NH5.CPRZEBN
S. No 1005NLHV042705

CYRIX
SOLUTIONS
C/1 CHCA 063
Phase 1 of 2 - 1st floor
101, Q Road, 11/12
Cochin 682014

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO. 1800 - 425-7669
BAR CODE (8004-890615225)
0831273

