


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH CHAVAKAD
3	Name of Equipment with Make, Model and Serial Number	Ocular Ultrasound Make - sonomed Model - 300A+ S/N - 0350-0110-0218
4	Equipment ID/ Barcode	#100096 - 0820799
5	Date of purchase/ Year of manufacture/Installation Date	As per ticket master installation date is 13-12-2010 (Actual installation date is 27-01-2010 ,stock book attached for reference)
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	07-03-2024
9	Action taken	Checked the equipment and found mother board and probe are defective. need to replace the main board and probe for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Main board for A Scan Model 300A+ - 1,65,200/- A Scan probe - 1,60,160/- TOTAL - 325,360/-

13	Asset Value	1,51525/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	214.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report and OEM quotation Attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that probe and mother board are defective. The machine installed on 27-01-2010 and aged up to 14 years 1 month. The repair cost is 214.5% and both criteria met for RBER. As per tender clause 5.3.14.1 we are recommending equipment for condemnation,
17	Name & Signature of CYRIX Authority	BIJO T JOY 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Quotation cost is very high



  
Signature of JC BM (NHM)

Date:



  
Signature of  
Superintendent/ Medical Officer (i/c)  
SUPERINTENDENT  
TALUK HEAD QUARTERS HOSPITAL  
CHAVAKKAD



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 194853

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Thgh</u>	Call Registration Date : <u>7/3/24</u>
Address : <u>Chavakkad</u>	Caller ID : <u>100096</u>
<u>Thiruvananthapuram</u>	Date of Visit : <u>9/3/24</u>
Ph : <u>9961564465</u>	Asset No. : <u>0820799</u>
	EQPT Name : <u>Ocular ultrasound</u>
	Manufacture <u>Sonomed</u> Model : <u>300 AT</u>
	S. No. <u>0350-0110-0218</u> Dept. <u>EYE OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Equipment no going on

Action Taken : checked the equipment and found mother board and probe are defective

Completed  Date : 9/3/24 Time : 11:00am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy Joshy</u>	<u>9/3/24</u>	<u>10:00am</u>	<u>11:00am</u>

Customer Remark  Completed  Pending

Service Engineer Name : Christy Joshy

Signature : [Signature]

Date : 9/3/24

Contact Number : 6238844347

Customer Name : [Signature]

Date : 9/3/24

Contact Number : 9961564465

Hospital Seal : JC-OPTHALMOLOGY

**SUPERINTENDENT**  
**TALUK HEAD QUARTERS HOSPITAL**  
**CHAVAKKAD**



# LUMEN

MEDIWORKS (P) LTD

GST: 32AAECL4197J1ZB

CIN: U52335TN2020PTC139377

## QUOTATION

<b>Customer</b>	CYRIX HEALTH CARE	<b>QTN. No</b>	LMW-1/587/24
<b>Address</b>	30/641B, PETTA, PUNITHURA, KOCHI – 682038.	<b>Date</b>	26.03.2024
		<b>Reference</b>	EMAIL DT. 26.03.2024
		<b>Currency</b>	RUPEES
		<b>Country of origin</b>	USA
<b>S. No</b>	<b>Product Description</b>	<b>Qty</b>	<b>Price Incl. GST</b>
1	Main board for A Scan Model 300A+	1	1,65,200/-
2	A Scan Probe	1	1,60,160/-

### TERMS & CONDITIONS:-

**Payment** : 100% advance along Purchase Order.

**Supply Period** : 3 – 4 weeks on receipt of confirmation of order with advance.

**Validity** : 20 days.

For Lumen Mediworks Pvt Ltd.

Authorized Signatory.

**REQUIRED QUOTATION 0820799 #100096 FOR A SCAN MACHINE PARTS SONOMED  
PacScan 300A+ - THQH CHAVAKAD**

4 messages

Sarang KM <arangcyrix@gmail.com>  
To: info@lumenmediworks.com  
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Tue, Mar 26, 2024 at 12:31 PM

Dear sir,


There is a complaint reported from THQH CHAVAKAD, Thrissur, Kerala, Regarding the complaint with SONOMED PacScan 300A+ Ascan machine ,Please share the price and availability of the **Main Board , Direct contact A-scan Probe**, as soon as possible.

Make : SONOMED Inc  
Model: PacScan 300A+  
Serial No: 0350-0110-0218

—  
**Thanks & Regards**

**Sarang.k.m**  
Divisional Incharge|North Zone  
Bemp-Kerala  
Mobile : +91 7593847150  
Email : arangcyrix@gmail.com

**CYRIX HEALTHCARE (P) LTD**  
30/641B | Petta | Punithura | Kochi | Kerala- 682038  
www.cyrixhealthcare.com



Lumen Mediworks <info@lumenmediworks.com>  
Reply-To: info@lumenmediworks.com  
To: Sarang KM <arangcyrix@gmail.com>  
Cc: BLESSEN JOSE <zm2.klbemp@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Tue, Mar 26, 2024 at 5:11 PM

Dear Sir,

In continuation to the email, with regard to the service of Sonomed, USA A Scan Model 300A+ at THQH, Chavakkad, you may find below the cost for Main board and A Scan Probe.

S. No.	Description	Qty	Amount incl GST
1	Main board for A Scan Model 300A+	1	1,65,200/-
2	A Scan Probe	1	1,60,160/-

**TERMS & CONDITIONS:-**

- Payment : 100% advance along with Purchase Order.
- Supply Period : 3 - 4 weeks on receipt of confirmation of order with advance.

Validity : 20 days.

Kindly go through it and feel free to contact us for more details.

Thanking you and looking forward to your consideration & support.

Best Regards,



Ph. 0484 3551448, +91 6238826591

Email id – info@lumenmediworks.com

Website – www.lumenmediworks.com

[Quoted text hidden]

---

**Sarang KM** <sarangcyrix@gmail.com>  
To: info@lumenmediworks.com  
Cc: BLESSEN JOSE <zm2.kibemp@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Tue, Mar 26, 2024 at 5:28 PM

Dear sir,

Thank you for your fast response

Kindly provide the quotation in a letter head.

[Quoted text hidden]

---

**Lumen Mediworks** <info@lumenmediworks.com>  
Reply-To: info@lumenmediworks.com  
To: Sarang KM <sarangcyrix@gmail.com>  
Cc: BLESSEN JOSE <zm2.kibemp@cyrix.in>, Bijo T Joy <bijocyrix@gmail.com>

Tue, Mar 26, 2024 at 6:31 PM


Dear Sir,


We are attaching quotation for Main board and A Scan Probe for Sonomed Ultrasound A Scan Model 300A+.

Thanking you,

[Quoted text hidden]

---

 **img20240326\_18285749.pdf**  
276K



SONOMED (USA) PASCAN 2 printer


Name of the equipment	Qty	Date of Purchase	Cost Rs	Category	No. & date of supplier	Name & address of the supplier	Type of equipment	Make	Model No	Date of issue	Manufacturer No
Sonomed make PASCAN MODEL-300A With Printer.	1	27/1/2010	Rs 168,000 + 6720 Rs 174,720/-	ST	No. 902259 27/01/2010	Ascorm Medical Imports - 41, Minto's St, 114 No. 50/39, Seemthell Avenue, Boma Road, Ekka Thangal, CHENNAI 600 097	A SCAN		300 A Printer	28/01/2010	0350-0110-0218
BIS 600 VP-12V. (UPS for PASCAN)	1										

Issued to OT.

Issued to OT.  
28/1/10  
PSK

Recd  
28/1/10  
BTH

Issued to OT



5<sup>2</sup>/<sub>10</sub> 10027227 PM.