


REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC VATANAPPALLY
3	Name of Equipment with Make, Model and Serial Number	Nebuliser MAKE: Life Care MODEL:CN-03 SN: NA
4	Equipment ID/ Barcode	0830991- #95299
5	Date of purchase/ Year of manufacture/Installation Date	08-02-2020 (attached stock book for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	10-02-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. need to replace the piston and motor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	1875/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Equipment installed on 08-02-2020 aged up to 4 years.. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation
17	Signature of CYRIX Authority	BIJO T JOY 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Quotation is not received and spares not available


 Signature of JC BM (NHM)



Superintendent/Medical Officer (Seal) Signature of seal Date



ASSISTANT SURGEON
 BLOCK FAMILY HEALTH CENTRE
 VATAINAPPALLY



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 222793

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>CHC</u></p> <p>Address <u>VATANAPALLY</u> <u>THRISSUR</u></p> <p>Ph : <u>9605043418</u></p>	<p>Call Registration Date : <u>10/12/2024</u></p> <p>Caller ID : <u>95299</u></p> <p>Date of Visit : <u>12/10/2024</u></p> <p>Asset No. : <u>0830991</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>lifecare</u> Model : <u>CM-03</u></p> <p>S. No. <u>N/A</u> Dept. <u>Op.</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : machine not switching on

Action Taken : Checked the machine and found that machine is not switching on. Found motor and piston are defective. need to replace the piston and motor for further checking

Completed Date : 12/10/24 Time : 11:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>12/10/24</u>	<u>10:00am</u>	<u>11:00</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Christy</u></p> <p>Signature : <u>Christy</u></p> <p>Date : <u>12/10/2024</u></p> <p>Contact Number : <u>6238844347</u></p>	<p>Customer Name : <u>DS4US.k.m</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>12/10/2024</u></p> <p>Contact Number : <u>9605043418</u></p> <p>Designation : <u>Senior Nursing Officer</u></p> <p>Hospital Seal :</p>
---	---



[Signature]
ASSISTANT SURGEON
BLOCK FAMILY HEALTH CENTRE
VATANAPALLY

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
0830991



Life care



Slide To Open


SUPERINTENDENT
BLOCK FAMILY HEALTH CENTRE
VATANAPPALLY, P O THRITHALLUR
PIN: 680 619



CYRIX™ BIOMEDICAL EQUIPMENT MAINTENANCE
HEALTHCARE SYSTEMS LTD PROGRAMME (BEMP)


TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)

0830991



LifeCare






Slide To Open


**SUPERINTENDENT CENTRE
BLOCK FAMILY HEALTH CENTRE
VATANAPPALLY, PO THRITHALLUR
PIN: 680 619**

PERFORMED BY:
 NEXT PMS DUE: 08/11/22
 PMS DONE ON: 08/08/22
 PREVENTIVE MAINTENANCE SERVICE
 BEMP SERVICE PROVIDER
 CYRIX HEALTH CARE PVT.LTD

Compressor Nebulizer

Model : CN-03
 Input : 230V / 50 Hz
 Power : 80 VA
 Lot No. : 082019
 Mfg Date : Aug 2019
 Shelf Life : Not Applicable
 Made in India

 Technology from The frontiers of medical science Kirloskar Group	BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BMMP)	 KMSCL in service to public health
Toll FREE- 1800-2700-690		QR/2D MATRIX 
KTPL BARCODE - (8907579)- 80080535		


 ID/ SI No: 80080535
 Cal. Date: 19.10.2021
 Due Date: 19.10.2022
 Medical Engineering & Services
 Thrissur - 21. Ph: 0487-2420367



SUPERINTENDENT
 KIRLOSKAR HEALTH CENTRE
 THRISSUR

Page no.

90

Nebulizer - Rd Issue

(3)

ON WITH NULL 81
 85
 83
 87
 89
 91
 93
 BELT 95
 TAVAC 97
 99
 101
 103
 105
 107
 PS 109
 111
 113
 117
 119
 121
 123
 125
 127
 WHEEL 129
 131
 133
 135

28/4/17 Round from the Store
 (purchase from Home account)
 (NEC28 0m301)
 1 April
 [Signature]

28/04/17 Issued to ward NIL
 [Signature]

8/2/2020 PO in from AIT 4 5
 Sub centre Stock

kept in ~~store~~ ~~pharmacy~~ page no: 197 MC Register (2)

1/7/2020 Issued to ward 2
 not working kept in wooden alarm clock store room

10/9/2020 Issued to ward 1 2 (25)

11/9/2020 Transferred to Unserviceable
 registers (P-7) nos-1

8/12/2020 Issued to OP 1

23/10/2022 Rd from Store 2
 KLO0892/LS/00012/22-23

25/10/2022 Issued to OP 2



[Signature]
 SUPERINTENDENT
 BLOCK FAMILY HEALTH CENTRE
 VATANAPPALLY, P O THRITHALLUR
 PIN: 680 619