


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC NENMANIKKARA
3	Name of Equipment with Make, Model and Serial Number	Digital Bp Apparatus Make- OMRON Model -HEM-7120 Sr.no -20170417415VGI
4	Equipment ID/ Barcode	0840563 #94622
5	Date of purchase/ Year of manufacture/Installation Date	01-10-2018
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	06/02/2024
9	Action taken	Checked the equipment and found board and pump are defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA


13	Asset Value	2,200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked equipment found board and pump damaged, equipment installed on 01-10-2018 and covered up to 5 year and 3 month years, Quotation not submitted since spares are not available in the market.so we are recommend equipment for condemnation.
17	Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Board and pump defective.


 Jr. Consultant (Biomedical)
 NHM, Thrissur

Signature of JC BM (NHM)

Date




 Signature of
 Superintendent/Medical Officer (i/c)

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്
 കുടുംബാരോഗ്യ കേന്ദ്രം
 നെമ്മണിക്കൽ



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 222836

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC NENMANIKKARA</u>	Call Registration Date : <u>6/02/2024</u>
Address <u>THRISSUR</u>	Caller ID : <u>94622</u>
Ph : <u>9074995925</u>	Date of Visit : <u>7/02/2024</u>
	Asset No. : <u>0840563</u>
	EQPT Name : <u>Digital BP Apparatus</u>
	Manufacture <u>Omnion</u> Model : <u>HEM7120</u>
	S. No. <u>2017041741</u> Dept. <u>5791</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : BP Apparatus is not working.

Action Taken : checked the equipment and found Board and pump are defective

Completed Date : 7/02/2024 Time : 10.am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ABSIL TONY</u>	<u>7/02/2024</u>	<u>9.am</u>	<u>10.am</u>

Customer Remark

Completed

Pending

Service Engineer Name : ABSIL TONY

Signature : [Signature]

Date : 07/02/2024

Contact Number : 9795328470

Customer Name :


Signature : [Signature]

Date :

Contact Number :

Designation :

Hospital Seal :



Handwritten Malayalam text: ബി.ടി.സി. സെന്റർ ഇൻ ചാർജ്
ആരോഗ്യ സേവന കേന്ദ്രം
നേമനിക്കര

AUTOMATIC BP APPARATUS

9

		Receipt	Issue	Balance
	<p>BEORE</p> <p>Received from JPHN</p> <p>Naseema</p> <p>MODEL - HEM - 7120</p> <p>MRP - 2340/-</p>	1		1
1/10/18	<p>Hand over the charge to</p> <p>Smt. Sai Sw in this</p> <p>institution</p>			
1/10/18	<p>VAdun</p> <p>Received from JPHN</p> <p>Naseema</p>	1		1
1/10/18	<p>Issued to OP</p> <p>Blocked by window</p>	10/6	1	Nil



2ms

13/2/24

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്

കുടുംബാരോഗ്യ കേന്ദ്രം

നെമ്മനിക്കര

