


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	UPHC V.R PURAM
3	Name of Equipment with Make, Model and Serial Number	Weighing Machine Make : S.CURE Model : NA Sr.No- NA
4	Equipment ID & Barcode	#112948 - 0845045
5	Date of purchase / Year of manufacture /Installation Date	21-11-2020
6	Warranty details (Yes/No)	Nil
7	*AMC/ CAMC Period agreed at the time of purchase	Nil
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24-05-2024
9	Action taken	Checked the machine and found board and sensor are faulty. Need to replace these spares for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending
12	Cost of spares (specify parts and cost)	Nil

13	Asset Value	850/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked equipment found board and sensor damaged, equipment installed on 21-11-2020 and aged up to 3 years 6 months, Quotation not submitted since spares are not available in the market. and recommend equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 



*Not mandatory


#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Board and sensor all not get as spare



 Signature of JC BM (NHM)




ഡോ. എൽബർട്ട് റാഫേൽ
M.B.B.S.
റബി. നമ്പർ : 54101
ഇഡിക്കൽ ഓഫീസർ

Superintendent / Medical Officer (i/c) Signature of seal Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 186783

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>UPHC V.R. puzam</u></p> <p>Address : <u>Thrissur</u></p> <p>Ph : <u>7025330194</u></p>	<p>Call Registration Date : <u>24/5/24</u></p> <p>Caller ID : <u>112948</u></p> <p>Date of Visit : <u>25/5/24</u></p> <p>Asset No. : <u>0845045</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture : <u>S. line</u> Model : <u>na</u></p> <p>S. No. : <u>na</u> Dept. : <u>DP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : weighing machine not working

Action Taken : checked the machine and found Board and sensor are faulty. Need to replace these spares for further checking

Completed Date : 25/5/24 Time : 11:55pm Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Sahas</u>	<u>25/5/24</u>	<u>10:55pm</u>	<u>11:55pm</u>

Customer Remark Completed Pending

ഡോ. ഗിരീബാർട്ട് നാപ്പോൾ
 M.B.B.S.
 റജി. നമ്പർ : 54101
 മെഡിക്കൽ ഓഫീസർ

<p>Service Engineer Name : <u>Sahas</u></p> <p>Signature : <u>SL</u></p> <p>Date : <u>25/5/24</u></p> <p>Contact Number : <u>7593847135</u></p>	<p>Customer Name : <u>Nisha P.A</u></p> <p>Signature : <u>Nisha</u></p> <p>Date : <u>25-5-24</u></p> <p>Contact Number : <u>25-5-24</u></p> <p>Designation : <u>Staff Nurse</u></p> <p>Hospital Seal :</p>
---	--



STOCK REGISTER

Name of Articles WEIGHING MACHINE Cost Rate

Date	PARTICULARS	V.No. or B.No.	RECEIPT		ISSUED		BALANCE	
			Quantity		Quantity		Quantity	
26/3/18	Received from Thakur hospital (Shool)		1		1		1	
27/3/18	Issued to JPHN				1			—
	Not working condition (5/19)							
21/12/20	Received from Smart Surgical		1				1	
21/12/20	Demanded SN				1			—
18/9/21	Received from Kalyuga medicine Depot		1				1	
	Issued to JPHN				1			—

B.No: C/531
B-11 Date 21/12/20
Rs-1200/-

B.No-115
Exp. 12/28
Rs 1602.04/-

STOCK REGISTER


22

Name of Articles WEIGHING MACHINE Cost Rate

Date	PARTICULARS	VNo. or Bill No.	RECEIPT		ISSUED		BALANCE	
			Quantity		Quantity		Quantity	
26/3/18	Received from Dinku hospital (Shol)		1		1		1	
27/3/18	Issued to JPHN				1		—	
	not working condition (5/19)							
21/12/20	Received from Smart Surgical		1				1	
21/12/20	Issued to SN				1		—	
12/9/21	Received from Karnya Medical Dept		1				1	
	Issued to JPHN				1		—	

B.No: e/531
B-11 Date 21/11/20
Rs-1200/-

B.No-115
Exp. 12/28
Rs 1608.04/-



MAX: 180Kg D=50g

 s.cure®



CYRIX
HEALTH CARE SYSTEMS

**BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)**

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)

0845045




*Technology from The
frontiers of medical science*
Kirloskar Group

**BIOMEDICAL
MAINTENANCE PROGRAMME
(BMMP)**



QR/2D MATRIX



0845045

0845045

CYRIX
HEALTHCARE PVT LTD
C/11/PHCV02A
PRO: Digital wash machine
NO: 0845045
Calom: 2019123
ad ch: 14/12/24
ISO 9001:2008 Certified Co. Ph: 9744555073

S. Cure
S. Cure
Bullington St. Kottayam
MAX: 11005011-3000
22AAA
CEC
2015

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0845045

KTPL
Technology from The frontiers of medical science
Kirloskar Group
BIOMEDICAL MAINTENANCE PROGRAMME (BEMP)
KMSCL
in service to public health
Toll FREE- 1800-2700-690
KTPL BARCODE - (8907579)- 80082799
QR/2D MATRIX

ID/Sl No: 80082799
Cal. Date: 29-12-21
MES Due Date: 28-12-22
Medical Engineering & Services
Thissur - 21. Ph: 0487-2420367
MES CALIBRATION CENTRE

URBAN FAMILY HEALTH CENTRE
Pin - 680722
V.R. PURAM