


964

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	PHC MUDAKKUZHA
3	Name of Equipment with Make, Model and Serial Number	Equipment: BP APPARATUS Make : ELKO Model : NA Sr.no : NA
4	Equipment ID & Barcode	111504 & 0744574
5	Date of purchase / Year of manufacture /Installation Date	09-10-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	16-05-24 (Toll Free)
9	Action taken	Checked the machine and found that main board and pressure sensor are faulty.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs.2200
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that main board and pressure sensor are faulty. Equipment installed on 09-10-2020 and the machine outlived 4+years. Quotation not submitted since the spares are not available in the market..So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Unable to repair reported by Cyrix. main board and pressure sensor faulty.



24/6/24

Signature of JC BM (NHM)

Superintendent / Medical Officer (I/C)



Signature of seal


29/06/24
Dr. RAJIKA KUTTAPPAN
(Assistant Surgeon)
MEDICAL OFFICER
F.H.C. MUDAKKUZHA

Date

**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



CYRIX
HEALTHCARE PVT LTD

No. : 200720

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : Phc	Call Registration Date : 16/15/2024
Address : Phc mudakuzha Ernakulam, Kerala	Caller ID : 111504
Ph : 9605954537	Date of Visit : 22/05/2024
	Asset No. : 0794574
	EQPT Name : BP apparatus
	Manufacture : elico Model : n9
	S. No. : na Dept. : OP

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : checked Machine and found Board Issue checked Battery, powersupply Display working but Not sensing pressure Board Issue Confirmed. Board For BP apparatus Not available. (Installed on 2/10/2022)

Completed Date : 22/5/24 Time : 10:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1. Mainboard with pressure	1		
2. sensor			
3.			

Cyrix Engineer	Date	Start Time	End Time
Yechu Krishna	22/5/24	9:30AM	10:00AM

Customer Remark : Completed Pending

Service Engineer Name : Yechu Krishna
Signature :
Date : 22/10/2024
Contact Number : 8086147236

Customer Name : F.H. MUBARAKUZZHA
Signature :
Date : 22/5/24
Contact Number : 828174537
Designation : N/O 80L
Hospital Seal :

BP- Apparatus Digital - 1

3

		Rd	Bestm	
17/09/18	R/d from the store - 1		1	✓
09/10/18	Rd from the store (LED) Page no. 14 (Nonchem)	2	3	+
24/2/21	Rd from HI (Digital)	1	4	+
15/10/21	Rd from the store (Digital)	1	5	+
20/1/22	Rd from the store (Digital)	1	6	+
10/1/22	Rd from the store (Digital BP 02)	1	7	+
20/1/23	To out to unserviceable Resistor Vol 1 p. no 3 LED BP apparatus		6	Binet
11/3/24	Rd from store over plan and circuit Digital apparatus	1	7	+

