

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**


**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	PHC Kurukkanmoola
3	Name of Equipment with Make, Model and Serial Number	Equipment-Weighing machine Make-Samso Model- N/A Sn- N/A
4	Equipment ID/ Barcode	1240010//Ticket 110768
5	Date of purchase/ Year of manufacture/Installation Date	18-03-2019
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	13/05/2024 - Toll free
9	Action taken	Checked machine, found load sensor and display board defective Enquired spare from local vendor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Load sensor for samso weighing scale -1416/- Display board -448/- <b>Total=INR 1864/-</b>



*Handwritten notes in Malayalam script, partially illegible.*

13	Asset Value	RS 1600/-
14	*Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	116.5%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Quotation from local vendor Cyrux service report
16	Reasons for recommending the equipment as BER	Checked machine, found load sensor board and display board defective Equipment installed on 18.03.2019 aged up to 5 years. Repairing cost is 116.5% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	NIYAS HADI AP 

Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*Sensor and display board of the machine are defective. Repairing cost is high. Recommend condemnation.*

  
24  
CONSULTANT  
Bio Medical Engineer  
Signature of JC, BM (NHM)  
National Health Mission  
(Arogyakeralam)  
Wayanad-673122

Date:

*8/6/2024*



Signature of  
In-charge/Resident/Medical Officer



മുഖ്യ ഓഫീസർ  
ആരോഗ്യ സേവന വിഭാഗം  
വേണാട് ജില്ലാ ആരോഗ്യ കേന്ദ്രം  
ഫോൺ - 670 646



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 220812

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC</u></p> <p>Address <u>PHC Kurukkamoolu,</u> <u>Wayanad.</u></p> <p>Ph : <u>9496138685</u></p>	<p>Call Registration Date : <u>13/5/24</u></p> <p>Caller ID : <u>110768</u></p> <p>Date of Visit : <u>15/5/24</u></p> <p>Asset No. : <u>1240010</u></p> <p>EQPT Name : <u>Weighing Machine</u></p> <p>Manufacture <u>Samsa</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not getting ON

**Action Taken :** Checked and found load cell and display board defective. Needs to service or replace the same

Completed  Date : 15/5/24 Time : 12:00 PM Spare Required

Spare Replaced  Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Ameer. C. H</u>	<u>15/5/24</u>	<u>9:30 AM</u>	<u>12:00 PM</u>

Customer Remark \_\_\_\_\_ Completed  Pending

Service Engineer Name : Ameer. Customer Name : Saranya. M V  
 Signature : Ameer Signature : NA  
 Date : 15/5/24 Date : 15.05.24  
 Contact Number : 7994374680. Contact Number : 8078336306  
 Designation : Nursing officer  
 Hospital Seal : \_\_\_\_\_



<b>Ospio Healthcare</b> Kannadikal, calicut, Kerala, GSTIN: 32AZLPT7342N1ZP Mobile: 9447338356	Quotation No. 136	Quotation Date 26/03/2024	Expiry Date 25/05/2024
	BILL TO <b>CYRIX HEALTHCARE PVT LTD</b> Address: Poonithura, Emakulam, GSTIN: 32AAFCC2499H2ZM State: Kerala		

SHIP TO <b>CYRIX HEALTHCARE PVT LTD</b> Address: Poonithura, Emakulam,
--

S.NO.	ITEMS	QTY.	RATE	TAX	AMOUNT
1	Load sensor for samsco digital scale	1 PCS	1,200	216 (18%)	1,416
2	Display for weigh scale	1 PCS	400	48 (12%)	448
<b>TOTAL</b>		<b>2</b>		<b>₹ 264</b>	<b>₹ 1,864</b>

HSN/SAC	Taxable Value	CGST		SGST		Total Tax Amount
		Rate	Amount	Rate	Amount	
-	400	6%	24	6%	24	₹ 48
-	1,200	9%	108	9%	108	₹ 216

**Total Amount (in words)**  
 One Thousand Eight Hundred Sixty Four Rupees

**Terms and Conditions**  
 1. Goods once sold will not be taken back or exchanged  
 2. All disputes are subject to [ENTER\_YOUR\_CITY\_NAME] jurisdiction only





---

## request for spare quotation

messages

**yas Hadi AP** <niyascyrix@gmail.com>  
>: ospiohealthcare@gmail.com

Tue, 26 Mar 2024 at 11:55 a

Dear Sir  
Please send the quotation for following spare of Samsco digital weight machine  
1-load sensor and board  
2-dispay of weighing machine

Thanks & regards

NIYAS HADI  
District incharge  
BEMP project  
Wayanad  
kerala  
mob -7593847171

email -niyascyrix@gmail.com

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

---

**ospio healthcare** <ospiohealthcare@gmail.com>  
>: Niyas Hadi AP <niyascyrix@gmail.com>

Tue, 26 Mar 2024 at 12:10 p

I will send it soon.  
[Quoted text hidden]

---

**ospio healthcare** <ospiohealthcare@gmail.com>  
>: Niyas Hadi AP <niyascyrix@gmail.com>

Mon, 1 Apr 2024 at 3:25 p

Dear Sir/Madam,

Please refer to the attached DOC.

Thanks & Regards,

**OSPIO HEALTHCARE**  
11/502, KRISHNA KRIBA BUILDING,  
PARAMBIL KADAVU, KANNADIKKAL,  
VENGERI (P.O), CALICUT -673010  
MOB : 919447338356

[Quoted text hidden]

18/3/2019


1. B.P. Apparatus		2	
2. Weighing machine		2	(1) carried over to Substock Register PAGE NO: 23 Volume 3C
3. Thermometer		2	
4. Examination tunic		2	
5. Stethoscope		1	

21/10/19

1. Stethoscope	2	2	no 13 B P 161
2. Galvanometer (on call plus 1)	1	1	11 p 185
3. Galvanometer strip (on call plus 1)	100	100	11 p 63
4. Nebulizer (like can)	1	1	Substock 3 B p 195
5. Refrigerator (190 litre)	1	1	Returned to store no (519) 22
6. Stabilizer (v. covered)	1	1	
7. Register for comp. sink movement	1	1	by <del>...</del>

copy kit

Nebulizer	1		
O2 mask adult	10.		v.w. Pg no
adult (nasal O2 set)	10.		74
Finger tip pulse oxymeter.	2.		

Deeya P K Nursing Officer CA  




മെഡിക്കൽ ഓഫീസർ  
 പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം  
 കുറുക്കുന്ന് മൂല വയനാട്  
 പിൻ-670 646 98



