


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	DH PALAKKAD
3	Name of Equipment with Make, Model and Serial Number	PHARMACY REFRIGERATOR MAKE : ELECTROLUX MODEL : ER190D SN : 60200858
4	Equipment ID/ Barcode	0911233 # 90341
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - Nil As per stock book installation date 16/02/2006, stock book attached for reference
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	12/01/2024 (toll free)
9	Action taken	Checked the machine found not cooling. Problem with compressor, condenser, cooling coil , required gas charging. Enquire local technician for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	190 L Fridge Compressor = 3800/- Cooling coil = 1750/- Gas refilling charge = 3500/- Inner condenser = 850/- Total = 9900/-

13	Asset Value	RS.26494/- (Actual stock book value of the equipment is 9600/-)
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	103.1%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked and found not cooling. Problem with compressor, inner condenser, cooling coil, required gas charging . Equipment installed on 16/02/2006. aged upto 17 year 11 months. Repairing cost is 103.1 % with respect to book value of the equipment. both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


As per stock book installation date is 16/02/2006 and aged up to 17 years.

Consultant (Biomedical Engg)
NHM (Arogyakeralam), Palakkad



7/07/24


Signature of JC BM (NHM)

As per stock book installation date is 16/02/2006


4/7/24

Date: _____




Signature of
Superintendent/Medical Officer (i/c)
Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad.



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 175247

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <i>District Hospital</i></p> <p>Address <i>Palakkad</i> <i>Kerala</i></p> <p>Ph : <i>9447889799</i></p>	<p>Call Registration Date : <i>12/01/2024</i></p> <p>Caller ID : <i>90241</i></p> <p>Date of Visit : <i>13/01/2024</i></p> <p>Asset No. : <i>0911233</i></p> <p>EQPT Name : <i>Pharmacy Refrigerator</i></p> <p>Manufacture <i>Electrolux</i> Model : <i>ER 190D</i></p> <p>S. No. <i>Na</i> Dept. <i>Laboratory</i></p>
--	--

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : *Not getting cool and*
..... *compressor damaged.*

Action Taken : *Checked the equipment and found refrigerator*
..... *not getting cool and compressor damaged. Enquired spare with*
..... *vendor.*

Completed **Date :** *13/01/2024* **Time :** *12:20 pm* **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Muhammed lybal</i>	<i>12/01/24</i>	<i>11:50 am</i>	<i>12:20 pm</i>

Customer Remark *Kindly requesting 200 RBRER*
..... *Ans Kerala H.O*

Completed **Pending**



Service Engineer Name : *Muhammed lybal*

Signature : *[Signature]*

Date : *13/01/2024*

Contact Number : *7025548890*

Customer Name : *MD N T S J R*

Signature : *[Signature]*

Date : *13/01/2024*

Contact Number : *9447889799*

Designation : *Deputy Director of Health Services & Superintendent, Palakkad District Hospital*

Hospital Seal :

BHARATH

REFRIGERATION

DOOR NO. 18, MINI INDUSTRIES,
KALMANDAPAM, PALAKKAD - 678 001

Repairing & Servicing Of :-
REFRIGERATORS
AIR CONDITIONERS
DEEP FREEZERS
WASHING MACHINES
CHILLERS, COOLER

Ref:

Quotation for fridge Repairing

Date : 30/1/24

The cyrus Health Care
Cochin.

1	190 Lt Fridge Compressor replacement charge	} = - 3800.00
2	Cooling Coil replacement	} = - 1750.00
3	Gas refilling charge with driver, working charge	} = - 3500.00
4	Insurer Condenser replacement charge	} = - 850.00
5	to and fro charge	400.00
		<u>10300.00</u>

(Prices Tentative from holiday)

For Bharath Refrigeration



Proprietor

REFRIGERATOR.



31.1.06	Store Supdt Dist Hospital Palakkad. Stock Book No. 129 Old refrigerator warranty period Expired, Not working. Capacity 300 lit. Godrej Ultra. Reported. S.No. RCAMP of 2006.	1		1 F	
16.2.06	S.S. DH. PKD				
	Electrolux Refrigeration 180L. Rs. 800/- 5 years warranty V. Guard Sterilizer 500VA Rs 1000/-	1 F		2 F	
	Fridge Stand Rs 200/-	1*			
	Total 9600/-				
17.2.06	Used in late		2 FRS	NIL	2 FRS
29.5.06	Electrolux Refrigerator	4 F		6 F	
	Store Supdt - Dist Hospital PKD Capacity Rs 9672/- per one. Warranty V. Guard Sterilizer Rs 924/44 per one.				
30.5.06	Used in late		4 F + 4 S	NIL	6 F + 6 S
	1 - Photo				
	1 - Photo				
	1 - CP				
	1 - Enfo				
	1 - copy (computer)				
	1 - Store				

Vanishan
Deputy Director of Health Services
& Superintendent,
District Hospital, Palakkad



Vijayar
Deputy Director of Health Service
& Superintendent,
District Hospital, Palakkad
Dept. of Health
GOVT. OF KERALA