


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	PALAKKAD
2	Name of Hospital	WOMEN AND CHILD HOSPITAL PALAKKAD
3	Name of Equipment with Make, Model and Serial Number	FULLY AUTOMATIC ANALYZER Make : ERBA Model: EM360 SN: 60421
4	Equipment ID/ Barcode	0912654 # 114615
5	Date of purchase/ Year of manufacture/Installation Date	30-05-2014
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	01-06-2024 (toll free)
9	Action taken	Checked the machine found assembly of photometer with packing, internal tubing kit with connector, spt assembly, vacume pump assembly, big knf pump, 100ul syringe, rgt cooling assy, 500ul syringe damaged enquire oem for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Assembly of photometer with packing= 282500/- Total internal tubing kit with connector= 65800/- Assembly spt = 63000/- Double head vacuum pump assy = 73600/- Big knf pump = 45400/- Assembly 100ul syring pump = 36000/- Rgt cooling assy with 6mm plate = 82900/- Rct base assy without dowel pin = 37800/- Assembly of stirrer with packing = 52500/- Assembly 500ul syringe pump = 35800/- Gst 18% = 139554/- Total = 914854/-

13	Asset Value	RS. 1082525 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	84.51%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached Quotation attached
16	Reasons for recommending the equipment as BER	Checked the equipment found assembly of photometer with packing, internal tubing kit with connector, spt assembly, vaccume pump assembly, big knf pump, 100ul syringe, rgt cooling assy, 500ul syringe damaged enquire oem for quotation. Equipment installed on 30/05/2014 and aged upto 10 years. Repairing cost is 84.51% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine aged up to 10 years and service cost ~~is~~ is 84.51% of asset value. Hence recommended for BER.


 Signature of **Consultant (Biomedical Engg) NHM (Arakkulam), Palakkad**



Date: 19/6/2024


SUPERINTENDENT
Govt. W&C Hospital
Palakkad
 Signature of
 Superintendent/Medical Officer (i/c)



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 207202

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>Women and Child</i>	Call Registration Date : <i>01/06/2024</i>
Address <i>Hospital Palakkad</i>	Caller ID : <i># 114615</i>
<i>Palakkad</i>	Date of Visit : <i>03/06/2024</i>
Ph : <i>8075267851</i>	Asset No. : <i>0912654</i>
	EQPT Name : <i>Fully automated Random Access Clinical Chemistry Analyser</i>
	Manufacture <i>Monsista</i> Model : <i>Er-360</i>
	S. No. <i>60421</i> Dept. <i>Lab</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Blank control and Result variation.*

Action Taken : *Checked the equipment Panel photometer assembly internal tubing kit, spt assembly, vacuum pump assembly, Big knr pump, 100ml and 500ml Syringe pump, RBS cooling assy damaged. Enquire OEM for quotation.*

Completed Date : *03/06/24* Time : *10:00 am* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			



Cyrix Engineer	Date	Start Time	End Time
<i>Akhilesb.p</i>	<i>03/06/24</i>	<i>10:00 am</i>	<i>12:30 pm</i>

Customer Remark Completed Pending

Service Engineer Name : *Akhilesb.p*
Signature : *[Signature]*
Date : *03/06/2024*
Contact Number : *759384 7154*

Customer Name : *Sheeja.P.V.*
Signature : *[Signature]*
Date : *3/6/24*
Contact Number : *8129593255*
Designation : *Lab*
Hospital Seal : *SUPERINTENDENT Govt W&C Hospital Palakkad*

Palakkad

QUOTATION

Bill To: CYRIX HEALTHCARE PVT LTD
 First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala,
 Pincode 682038 , Palakkad district , Kerala 678001

Ref. No : 2425QUOTE001200
Date : Jun 7, 2024 1:01 PM
Validity : 30 days
TBM GSTIN : 27AAACTP2038C1ZT

Customer GSTIN No : 32AAFCC2499H2ZM

Ship to, Government Women & Child Hospital,Palakkad
 Govt. District Hospital (Pregnancy Ward), Stadium Bypass Road, Top in
 Town, Mankavu, Palakkad, Kerala, 678001, India , Palakkad district ,
 Kerala 678001

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: EM360 SR.NO.: 60421

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same is / are given below, for approval:

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount
1	103093	90278990	ASSEMBLY OF PHOTOMETER WITH PACKING	1	282500		18	282500
2	104951	90278990	TOTAL INTERNAL TUBING KIT WITH CONNECTOR (XL600/300/V2/A4/EM360)	1	65800		18	65800
3	105315	90278990	ASSEMBLY SPT (SPARE) FOR XL-600/300-V2	1	63000		18	63000
4	114795	90330000	ASSEMBLY PUMP VACUUM DOUBLE HEADED GD - 6EA WITH NOZZLES_XLV2/ XL-640/2R_S	1	73600		18	73600
5	182496	90278990	Big Knf Pump Assy.	1	45400		18	45400
6	116910	90278990	ASSEMBLY 100 ul Syringe Pump - KEYTO WITH BRACKET_S_XLV2	1	36000		18	36000
7	114595	90278990	RGT COOLING ASSY WITH 6MM PLATE & SPACER_XLV2	1	82900		18	82900
8	111845	90278990	RCT BASE ASSY WITHOUT DOWEL PIN_XLV2	1	37800		18	37800
9	103091	90278990	ASSEMBLY OF STIRRER WITH PACKING.	1	52500		18	52500
10	116908	90278990	ASSEMBLY 500 ul Syringe Pump KEYTO WITH BRACKET_S_XLV2	1	35800		18	35800

Merchant Name :TRANSASIA BIO-MEDICALS LTD

SCAN AND PAY WITH UPI QR
 Merchant Name: TRANSASIA BIO MEDICALS LTD



Total (Rs) 775300

Discount (Rs.) (-)



	GST (Rs.)	139554
Rupees in words: Nine lakhs Fourteen thousand Eight hundred Fifty Four rupees only	Grand Total (Rs.)	914854

TERMS AND CONDITIONS :

1. **Payment :** Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of "**Transasia Bio-Medicals Ltd**" to **Bank A/C no. : 00600330002538 IFSC code: HDFC0000060. UPI ID : transasiabiomedicalsLtd@citibank.** You can also pay us through **UPI apps like BHIM/ Google Pay/ Paytm/PhonePe** etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
2. **GST :** As applicable
3. **Other Cost / Local duties :** Customer's responsibility
4. **Delivery :** Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You
Your faithfully,

For **TRANSASIA BIO-MEDICALS LTD**

TRANSASIA BIO-MEDICALS LTD.

Head Office: Transasia House, 8, Chandivall Studio Road, Andheri (E), Mumbai - 400 072. Tel.: (022) 4030 9000 Fax: (022) 2657 3030 Mumbai : Toll Free No. 1800 103 8226, Whatsapp No.: 8451048434, SMS No.: 9212433444

www.transasia.co.in

DOC NO : SC00-405 / ISS - 7



RECEIVED & ACCEPTED

CUSTOMER SIGNATURE & STAMP:

CUSTOMER NAME: _____

CONTACT NO: _____

ISO 13485 & ISO 9001 Certified Company





**REQUIRED QUOTATION FOR FULLY AUTO BIOCHEMISTRY ANALYZER 0912654 # 105363 @
WOMEN AND CHILD HOSPITAL PALAKKAD, PALAKKAD**

4 messages

Akhilesh P <akhileshcyrix@gmail.com>
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Wed, 3 Jul, 2024 at 9:29 am

Dear sir
Please share the spare quotation for fully automatic analyser @ women and child hospital palakkad. Details are furnished below.

Hospital Name : WOMEN AND CHILD HOSPITAL PALAKKAD

Equipment Name : FULLY AUTO BIOCHEMISTRY ANALYZER

Make : TRANSASIA

Model : EM360

SN : 60421

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Akhilesh P <akhileshcyrix@gmail.com>
To: s.das@transasia.co.in
Cc: Zm2.klbemp@cyrix.in, dm1.klbemp@cyrix.in

Tue, 9 Jul, 2024 at 4:07 pm

Dear sir
Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Akhilesh P <akhileshcyrix@gmail.com>
To: s.das@transasia.co.in

Mon, 15 Jul, 2024 at 12:59 pm

Dear sir

Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Mon, 15 Jul, 2024 at 3:13 pm

SANDEEP DAS M <s.das@transasia.co.in>
To: Akhilesh P <akhileshcyrix@gmail.com>
Cc: Zm2.klbemp@cyrix.in <Zm2.klbemp@cyrix.in>, dm1.klbemp@cyrix.in <dm1.klbemp@cyrix.in>

Dear Akhilesh ,

Please find the attachment of quote and do the needfull

From: Akhilesh P <akhileshcyrix@gmail.com>
Sent: Monday, July 15, 2024 12:59 PM
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: Zm2.klbemp@cyrix.in <Zm2.klbemp@cyrix.in>; dm1.klbemp@cyrix.in <dm1.klbemp@cyrix.in>
Subject: Re: REQUIRED QUOTATION FOR FULLY AUTO BIOCHEMISTRY ANALYZER 0912654 # 105363 @ WOMEN AND CHILD HOSPITAL PALAKKAD, PALAKKAD

[Quoted text hidden]

48 Fully Auto Analyser model A (NR4m)
 Receipt Issued Balance Qty

30/5/14

As per the order Suppl. W & C Hospital palakkad
 Received - the item From GISH Kottayam
 As per the order DPM plcd.
 Order no: DPMSCU - plcd / 2157/A 21 / 2013 / DPMSCU
 Dt 8/11/14 Shifted - the fully automatic
 Biochemistry Analyser From GISH Kottayam to
 W & C Hospital palakkad.
 Supply order no: 788 / 2013 / 49 A
 Equipment code
 49A

Name of the Equipment:
 Fully Auto Analyser model A
 Model - GM 360
 make = Transasia Bio medicals Ltd.

quantity: 01

30/5/14

used - 10 Lab

01 01



19/6/2014

SUPERINTENDENT
 Govt. W & C Hospital
 Palakkad

CE
TRANSASIA

**AUTOMATED RANDOM ACCESS
CLINICAL CHEMISTRY ANALYSER**

MODEL : EM 360

SERIAL NUMBER : 60421

**POWER SOURCE : 220 V AC, 50 Hz
750 VA**

TRANSASIA BIO-MEDICALS LTD.

MADE IN INDIA

FOR IN VITRO DIAGNOSTICS

**CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE
PROGRAMME (BEMP)**

TOLL FREE NO: 1800 - 425-7669

**BAR CODE : (8004-890615225)
0912654**



19/01/2024
**SUPERINTENDENT
Govt. W&C Hospital
Palakkad**