



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 206011

SIB

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u></p> <p>Ph : <u>7012896198</u></p>	<p>Call Registration Date : <u>13-04-2024</u></p> <p>Caller ID : <u>106177</u></p> <p>Date of Visit : <u>14-04-2024</u></p> <p>Asset No. : <u>0110692</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>BPL</u> Model : <u>120/80 B9</u></p> <p>S. No. <u>MMTA.O.H.2913</u> Dept. <u>Fema/Surgical Unit</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Cult not inflating

Action Taken : Checked and found pump, control board and solenoid valve damage. Need to replace these parts for further checking of equipment.


Completed Date : 14-04-24 Time : 11:00 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>ADITHYA.T.S</u>	<u>14/04/24</u>	<u>10:30am</u>	<u>11:00am</u>

Customer Remark Completed Pending



<p>Service Engineer Name : <u>ADITHYA.T.S</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>14/04/24</u></p> <p>Contact Number : <u>9745054344</u></p>	<p>Customer Name : <u>Saeetha.V</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>14/4/24</u></p> <p>Contact Number : <u>7012896198</u></p> <p>Designation : <u>Nursing Officer</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
---	---

27/6/24

[Signature]


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

516

PROFORMA

Sl No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Manufacturer : BPL Model : 120/80 B9 Sl: MMTA0H12913
4	Equipment ID/ Barcode	106177/ 01 i 0692
5	Date of purchase/ Year of manufacture/Installation Date	31/3/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	14/04/2024 (Toll free)
9	Action taken	Checked and found pump, control board and solenoid valve defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 2200/-
14	Percentage value of the cost of spares with respect to Cost of Purchase / Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found pump, control board and solenoid valve defective. The apparatus was installed on 31/3/2021 and aged upto 3+ years. Quotation not attached as the spares are not available in the market. Hence recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Pump, control board and Solenoid valve is defective. ∴ as per the conditions, this can be categorised to as Beyond economic repair.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

 18-06-2024
Signature of JC BM (NHM)

Date:




Signature of
Superintendent/Medical Officer (i/c)
General Hospital
Thiruvananthapuram



70 Digital BP Apparatus

Date	Particulars	R/d	Issued	Balance	Remarks
22/11/19	R/d from Store (True check)	1		1	
22/2/21	R/d from store [Elko] (Exp Date 28/12/2030)	1	✓	2	B/W H/W
31/3/21	(LTD) Received from Store This is from Local purchase Register page no: 26 (Rs 1201.46) model Bx 575 Lot no S15C1B119 Batch no BH 19 - model B5 Bar code: 1) 0115893 2) 0115816 3) 0110982 4) 0110542	1		3	L/W F/W
20/10/2021	Physically found BPL 120/80 B9			4	B/W S/W


Lifecase

Expiry date

30/6/2030

Batch no

①
B/W
S/W



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800-425-7669



BAR CODE - 0110692

