

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	GH IRINJALAKUDA
3	Name of Equipment with Make, Model and Serial Number	Fumigator Make : Aerojet Model: Steriz SN:10j4821
4	Equipment ID/ Barcode	0812768 #107760
5	Date of purchase/ Year of manufacture/Installation Date	21-03-2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 23-04-2024
9	Action taken	Checked the machine and Found motor, bottle, filter and timer defective. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	motor - 8260/- bottle - 3245/- filter - 944/- Timer - 2124/- Total - 14573/-

13	Asset Value	20833 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	69.95%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report and quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine and Found motor, filter, bottle and timer are defective. Equipment installed on 21-03-2017 aged up to 7 years 1 month. Quotation submitted. Repairing cost is 69.95% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spare cost is very high

  
 Signature of JC BM (NHM)  
 Sr. Consultant (Biomedical)  
 Thiruvananthapuram

Date:

30/9/2024



Signature of Superintendent/Medical Officer (i/c)

സൂപ്രണ്ട്  
 മനറൽ ആശുപത്രി  
 ഇരിഞ്ഞാലക്കുട  
 പിൻ - 680 121





Quote photo

# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 193511

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

Health Facility : <u>GH</u>	Call Registration Date : <u>23/04/2024</u>
Address : <u>IRINJALAKUDA</u>	Caller ID : <u>107760</u>
<u>THRISSUR</u>	Date of Visit : <u>23/04/2024</u>
Ph : <u>8547108387</u>	Asset No. : <u>0812768</u>
	EQPT Name : <u>FUMIGATOR</u>
	Manufacture : <u>AEROJET</u> Model : <u>STER12</u>
	S. No. : <u>1074821</u> Dept. : <u>OT</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : MACHINE NOT SWITCHING ON

Action Taken : CHECKED THE MACHINE & FOUND MOTOR, BOTTLE, FILTER & THERMIST DEFECTIVE. NEED TO REPLACE FOR FURTHER CHECKING

Completed  Date : 23/04/24 Time : 12:30pm Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrrix Engineer	Date	Start Time	End Time
<u>AMBADY VS</u>	<u>23/04/24</u>	<u>10:00AM</u>	<u>12:30pm</u>

Customer Remark Completed  Pending

Service Engineer Name : <u>AMBADY VS</u>	Customer Name :
Signature : <u>[Signature]</u>	Signature :
Date : <u>23/04/24</u>	Date :
Contact Number : <u>7901996500</u>	Contact Number :
	Designation :
	Hospital Seal :



[Signature]  
മധുപ്രഭാ  
ഇൻ്റർ അസിസ്റ്റന്റ്  
ഇൻ്റീണാലക്കൂട്ട  
ഫീൽഡ് - 680 121

**BCS Healthcare Pvt.Ltd**  
3136,11nd door,Mathewsons square  
MetroPiller No:464 NearChangapuzha park  
Cochin, Kerala – 682024,  
Email : info@bcshealthcare.in  
web: www.bcshealthcare.in



GST NO: 32AAKCB2766M1Z

DATE: 16/05/2024

Ref: BCS/30/24-25

**QUOTATION**

TO,

**CYRIX HEALTHCARE PVT LTD**

SL.NO	PRODUCT NAME	QTY	RATE	TAX 18%	TOTAL
01	Fumigator Machine Make –Aerojet Model-Steriz				
	Spare:-				
	Motor	1	7,000.00	1,260.00	8,260.00
	Bottle	1	2,750.00	495.00	3,245.00
	Filter	1	800.00	144.00	944.00
	Timer	1	1,800.00	324.00	2,124.00
<b>GRAND TOTAL</b>					<b><u>14,573.00</u></b>

**Terms & Conditions**

- Delivery within 8 Days After confirmation of Purchase order
- Tax are included
- Payment 50% advance and balance 50% after delivery time

**For BCS HEALTH CARE PRIVATE LIMITED**





barcode - 0812768 ticket id - 107760-

Mon, 15 Jul, 2024 at 11:30 am

Bijo T Joy <bijocyrix@gmail.com>  
To: nfo@bcshealthcare.in  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

dear,  
kindly provide quotation for fumigator  
hospital - thqh irinjalakuda  
equipment - fumigator  
quotation for - motor,bottle,filter and timer  
barcode - 0812768  
ticket id - 107760

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Mail Delivery Subsystem <mailer-daemon@googlemail.com>

Mon, 15 Jul, 2024 at 11:30 am

To: bijocyrix@gmail.com



### Address not found

Your message wasn't delivered to **nfo@bcshealthcare.in** because the address couldn't be found or is unable to receive email.

[LEARN MORE](#)

The response was:

550 5.1.1 The email account that you tried to reach does not exist. Please try double-checking the recipient's email address for typos or unnecessary spaces. For more information, go to <https://support.google.com/mail/?p=NoSuchUser> 5614622812f47-3dab3dc5522sor1431646b6e.7 - gsmtip

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>  
To: info@bcshealthcare.in

Mon, 15 Jul, 2024 at 12:01 pm

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEATHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

----- Forwarded message -----

From: Bijo T Joy <bijocyrix@gmail.com>



---

**Quotation**

message

---

**CS Healthcare** <info@bcshhealthcare.in>  
To: Bijo T Joy <bijocyrix@gmail.com>

Wed, 26 Jun, 2024 at 10:04 a

Respected Sir/Madam  
Please find the attached file

Regards,  
BCS HEALTH CARE PRIVATE LIMITED





