


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	MALAPPURAM
2	Name of Hospital	PHC THUVVUR
3	Name of Equipment with Make, Model and Serial Number	Equipment : DIGITAL BP APPARATUS Make: DR MOREPEN Model: BP-15 Serial Number :2020091203564084877
4	Equipment ID/ Barcode	TICKET ID :121301 BARCODE NO : 1051128
5	Date of purchase/ Year of manufacture/Installation Date	30/11/2020
6	Warranty details (Yes/No)	NO WARRANTY
7	*AMC/ CAMC Period agreed at the time of Purchase	NO AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	05/07/2024
9	Action taken	Checked the machine and found problem with main board & display. Need to replace These spare for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	NOT WORKING
11	Recommendations for repair (required service details)	No recommendation
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	N/A
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX Service Report
16	Reasons for recommending the equipment as BER	Checked the machine and found problem with main board & display. The equipment was installed on 30/11/2020 and Aged up to 3 Years and 6 months. Quotation not submitted since spares not available in market. So we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	GOKUL KT 


*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

verified with damage for main board & display
 B. BER economic repair.





 15/11/20
 Jr. Consultant Biomedical Engineer
 NHM Malappuram

Signature of JC BM (NHM)

Date: 17/7/24




 Signature of
 Superintendent Medical Officer (i/c)
 PRIMARY HEALTH CENTRE
 TUVVUR 679 327



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 205576

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>P.H.C. THUVVUR</u></p> <p>Address : <u>MALAPPURAM</u></p> <p>Ph : <u>9495091942</u></p>	<p>Call Registration Date : <u>05/07/2024</u></p> <p>Caller ID : <u>121301</u></p> <p>Date of Visit : <u>06/07/2024</u></p> <p>Asset No. : <u>1051128</u></p> <p>EQPT Name : <u>Digital bp apparatus</u></p> <p>Manufacture Det. make/Model : <u>BP-15</u></p> <p>S. No. <u>20200912035</u> Dept. <u>OP</u> <u>64084877</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Digital bp apparatus not working.

Action Taken : checked the machine and found problems with main board & display. need to replace these spare for further checking.

Completed Date : 06/07/2024 Time : 10:30 AM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				



Engineer <u>Shibabudheen</u>	Date <u>06/07/2024</u>	Start Time <u>10 AM</u>	End Time <u>10:30 AM</u>
------------------------------	------------------------	-------------------------	--------------------------

Customer Remark : Completed

Signature : [Signature]

Designation : Biomedical Engineer

Address : Malappuram



MEDICAL OFFICER
PRIMARY HEALTH CENTRE
THUVVUR - 679 327

Service Engineer Name : Shibabudheen

Signature : [Signature]

Date : 06/07/2024

Contact Number : 9072522046

Customer Name : Fide-k

Signature : [Signature]

Date : 6/7/2024

Contact Number : 9747906885

Designation : Nursing officer

Hospital Seal :

FHC TUVVUR.

ELECTRONIC BP APPARATUS

Date	Particulars	Received	Issued	Balance	Remarks
11/5/2017	B/E P No - 41 Sub stock register (Came in stock)			2	
KIR	(1) - Received on 26/03/2002 - for life period				
KIR	(2) Received on 26/11/2016 (Rs 1466/- NCD fund)				
24/11/16	in use				
30/1/20	Rd from store	2 ✓		4	
10/2/20	in use in NCD clinic		2	4	
22/3/23	Received from store type core digital BP Apparatus 1			5	
22/3/23	Issued in NCD clinic	1 ✓		4	






[Signature]
15/7/24


Jr. Consultant Biomedical Engineer
NHM M. P.



[Signature]
MEDICAL OFFICER
PRIMARY HEALTH CENTRE
TUVVUR-679-327






 S. Jayaram
 15/11/2019
 Senior Biomedical Engineer
 NHM Malappuram



S. Jayaram
 MEDICAL OFFICER
 PRIMARY HEALTH CENTRE
 TUVVUR- 679 327