


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC KADAPURAM
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make :True check Model: NA SN: NA
4	Equipment ID/ Barcode	0830059#96183
5	Date of purchase/ Year of manufacture/Installation Date	19-12-2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registrationof complaint through email/ Toll free)	Toll free- 15-02-2024
9	Action taken	Checked the equipment and found board and pump are defective
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked equipment found board and pump damaged, equipment installed on 19-12-2020 and covered up to 3year, Quotation not submitted since spares are not available in the market.so we are recommend equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spare not available

Jr. Consultant (Biomedical)
NHM, Thrissur



Signature of JC BM (NHM)

Date:





Signature of
Superintendent/Medical Officer (i/c)

SUPERINTENDENT
COMMUNITY HEALTH CENTRE
KADAPPURAM
P.O. ANCHANGADI-680 514



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 185136

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.ki@cyrix.in

<p>Health Facility <u>CHC</u></p> <p>Address <u>Kadappuram</u> <u>Thirissur</u></p> <p>Ph : <u>9847214773</u></p>	<p>Call Registration Date : <u>15/02/2024</u></p> <p>Caller ID : <u>96183</u></p> <p>Date of Visit : <u>16/02/2024</u></p> <p>Asset No. : <u>0830059</u></p> <p>EQPT Name : <u>B.P. Apparatus</u></p> <p>Manufacture <u>Truecheck</u> Model : <u>N/A</u></p> <p>S. No. <u>N/A</u> Dept.</p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Bp apparatus not inflating

Action Taken : checked the equipment and found board and pump are defective. Need to replace these spares for further checking


Completed Date : 16/02/24 Time : 11:00am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>16/02/24</u>	<u>10:00</u>	<u>11:00pm</u>

Customer Remark Completed Pending



Service Engineer Name : Christy

Signature : [Signature]

Date : 16/02/24

Contact Number : 6238844342

Customer Name : Febm Samuel

Signature : [Signature]

Date : 16/2/2024

Contact Number : 9847214773

Designation : Superintendent

Hospital Seal : COMMUNITY HEALTH CENTRE, KADAPPURAM, P.O. ANCHANGADI-680 514



OFFICE
Date:.....
KADAPPURAM, P.O. ANCHANGADI-680 514

D: Date	Name of Item	Received	Issued	Balance
6/12/2020	Received from pharmacy store	1		
7/2/2020	NRM Register no: 36 issued to home care kit		1	nil
21				
14/12/2020	Received from pharmacy	1		
31	store			
13/12/2020	issued to palliative op		1	Nil
30				

