


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	THQH CHAVAKKAD
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: N/A Model: NY-10ZW SN: 2106243421090
4	Equipment ID/ Barcode	0820986 #85815
5	Date of purchase/ Year of manufacture/Installation Date	27-07-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 15-12-2023
9	Action taken	Checked the machine and found that sieve bed, compressor motor and main board defective , need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	26500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, and main board are faulty. These spare parts need replacement to assess the equipment's operational status. Unfortunately, the supplier and installation stock information is unavailable from end user as the equipment was donated without manufacturing details. Due to the lack of spare part information, we advise condemning the equipment as a special case (5.3.14.2).
17	Name & Signature of CYRIX Authority	Byo 


\*Not mandatory

#Based on the period of life and value as per the BER guidelines


\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

spare parts not available.

  
 Signature of JC BM (NHM)

Date: \_\_\_\_\_



  
 Superintendent  
 TALUK HEAD QUARTERS HOSPITAL  
 CHAVAKKAD



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. : 195062**

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>THH chavakkad</u></p> <p>Address <u>Thaissur</u></p> <p>Ph : <u>8943103098</u></p>	<p>Call Registration Date : <u>15/12/2023</u></p> <p>Caller ID : <u>85815</u></p> <p>Date of Visit : <u>16/12/2023</u></p> <p>Asset No. : <u>0820986</u></p> <p>EQPT Name : <u>Oxy gen Concentrator</u></p> <p>Manufacture <u>NA</u> Model : <u>NY-10ZW</u></p> <p>S. No <u>2106243421090</u> Dept. <u>Pain &amp; Palliative</u></p>
--	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Machine is not working.

**Action Taken :** Checked the machine and found that sieve bed, Compressor motor and main board defective. need to replace these spares for further checking.

Completed  Date : 16/12/2023 Time : 3:00 pm Spare Required


Spare Replaced  Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.				
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin-AR</u>	<u>16/12/2023</u>	<u>1.00 pm</u>	<u>3.00 pm</u>

**Customer Remark** Completed  Pending

<p>Service Engineer Name : <u>Aswin-AR</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/12/2023</u></p> <p>Contact Number : <u>8589965265</u></p>	<p>Customer Name : <u>SARITHA</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>16/12/2023</u></p> <p>Contact Number : <u>8943103098</u></p> <p>Designation : <u>Palliative</u></p> <p>Hospital Seal : <u>[Seal]</u></p>
--	---



**LAY SECRETARY & TREASURER**  
**TALUK HQ HOSPITAL**  
**CHAVAKKAD**



## THQH chavakad -0820986-85815- for supply details

3 messages

Thu, 11 Jul, 2024 at 1:52 pm

**Bijo T Joy** <bijocyrix@gmail.com>  
To: bmetsr2020@gmail.com, KMSCL <kmsclbmecentral1@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

dear mam,  
kindly provide the supply detail of below mentioned oxygen concentrator  
equipment - oxygen concentrator(donation)  
hospital -THQH CHAVAKAD  
barcode - 0820986  
caller id - 85815  
make - nil  
model - ZY-10ZW  
With Regards,

Thu, 11 Jul, 2024 at 4:23 pm

**KMSCL** <kmsclbmecentral1@gmail.com>  
To: Bijo T Joy <bijocyrix@gmail.com>

The supply details for the above mentioned equipment were not available with us. Please refer to the hospital stock data for additional details.  
[Quoted text hidden]


Regards,  
SALU U S  
DISTRICT BIOMEDICAL ENGINEER  
KMSCL  
Contact No: +91 91889 10013

Fri, 19 Jul, 2024 at 10:42 pm

**Bijo T Joy** <bijocyrix@gmail.com>  
To: KMSCL <kmsclbmecentral1@gmail.com>  
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, bmetsr2020@gmail.com

dear  
it is inform you that we have inquired about the supply details at the hospital. But we cannot find the supplier details  
With Regards,  
BIJO T JOY  
DISTRICT INCHARGE THRISSUR,  
**CYRIX HEATHCARE PRIVATE LIMITED**  
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia  
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]



# OXYGEN CONCENTRATOR

031

Date	Invoice no & particulars	Receipt	Issue	Balance	Remarks
19.5.2021	From District Drug Warehouse, KMSCL High road, opposite police officers quarters Thiruvananthapuram - 1 as per invoice No. KK008/ED/ 00083/21-22 Dated 19.5.2021 (MNO: K059051, B.No. 210400109 Mfg: 1-6-2021)	2	-	2	
22.5.21	Issued to Isolation ward		2	NIL	
25/5/2021	Received from Taluk head Quarters Hospital Chalakudy Issue note. NO. 5/ 25/5/2021				
	① ex: Yuwell - 9 f.5 AW	10		10	37674/1 Each
	② OLIVE OLV (10)	15		25	44706/1 Each
30.7.21	Issued to labour ward		1	24	
27.7.21	From District drug warehouse, High road opposite police officers quarters, Thiruvananthapuram as per inv. no. KK008/ED/00402/21-22 Dated 27.7.21 (Mfg: TATA AUTOCOMP)	5	-	29	
3.8.21	Issued to palliative care		5	24	
21.8.21	Issued to Isolation ward		5	19	
20.5.22	Issued to palliative care		1	18	
9-6-22	Issued to palliative care.		2	16	
28-6-22	Issued to palliative care		2	14	
8.11.22	Issued to palliative care		3	11	
9-2-23	Issued to Palliative Care		2	9	

