


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	PALAKKAD
2	Name of Hospital	PHC PATTITHARA
3	Name of Equipment with Make, Model and Serial Number	BIOCHEMISTRY ANALYZER MAKE : ROBONIK MODEL : PRIESTEST TOUCH SN : KT0540315RBK
4	Equipment ID/ Barcode	0943347 # 117999
5	Date of purchase/ Year of manufacture/Installation Date	26/02/2016
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	19/06/2024 (toll free)
9	Action taken	Checked the machine found not switching on. Problem with main Board and flow cell . Requested oem for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	RS.64760/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached OEM EOL letter attached
16	Reasons for recommending the equipment as BER	Checked the machine found not switching on. Problem main board and flow cell. Requested oem for quotation. . Equipment installed on 26/02/2016. aged upto 8 years 4 months . OEM informed that this model is obsolete. End of life letter issued As per tender clause 5.3.14.2 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AKHILESH P 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Equipment installed on 26/02/2016 and aged upto 8 years and OEM informed that this model is obsolete hence recommended for BER.

Consultant (Biomedical Engg)  
NHM (Arogyakeralam), Palakkad

  
4/07/24.  
Signature of JC BM (NHM)

Date:



  
മെഡിക്കൽ ഓഫീസർ  
കുടുംബാരോഗ്യകേന്ദ്രം  
പട്ടിത്തറ, പാലക്കാട് - 679534  
Signature of  
Superintendent/Medical Officer (i/c)





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :** 217615

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PfIC Pattithara</u></p> <p>Address <u>Palakkad.</u></p> <p>Ph : <u>9048548552</u></p>	<p>Call Registration Date : <u>19/06/24</u></p> <p>Caller ID : <u>117999</u></p> <p>Date of Visit : <u>20/06/24</u></p> <p>Asset No. : <u>0943347</u></p> <p>EQPT Name : <u>Biochemistry Analyser</u></p> <p>Manufacture <u>Robonik</u> Model : <u>Prebuch</u></p> <p>S. No. <u>KT0540215RBK</u> Dept. <u>LAB</u></p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** not working

**Action Taken :** checked the Machine found that not switching ON  
Problems with Main board and flow cell. Requested dem for quotation

Completed  Date : 20/06/24 Time : 11:00Am Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Tejas</u>	<u>20/06/24</u>	<u>10:00Am</u>	<u>11:00Am</u>

**Customer Remark** Completed  Pending

<p>Service Engineer Name : <u>Tejas</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/06/24</u></p> <p>Contact Number : <u>9848098422</u></p>	<p>Customer Name : <u>Suneer / LM</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/6/24</u></p> <p>Contact Number : <u>9895617551</u></p> <p>Designation : <u>Medical officer.</u></p> <p>Hospital Seal :</p>
---	--



മെഡിക്കൽ ഓഫീസർ  
കുടുംബാരോഗ്യകേന്ദ്രം  
679534

Date : 24.06.2024

Dear Valued Customer,



Thank you for choosing **Robonik (India) Pvt. Ltd.** We appreciate the opportunity to support our business. This letter is to inform you regarding the EOL (End of Life) of the instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0540315RBK.**

As required, we stated herewith that the said instrument Semi Automated Biochemistry Analyser (Model : Prietest Touch) having **Serial No. AT0540315RBK** as **End of Life (EOL) Notified.** (During from installation is 10 years).

Thank you in advance for your understanding and your cooperation. If you have any questions, please contact our Customer Care Department at Toll Free No. 18005727977.

Yours faithfully,

For **ROBONIK (INDIA) PVT. LTD.**



Authorised Signatory





+919833696980



We will mention all serial number on one letter and send to you

9:39 am

K 9:39 am ✓✓

Send Excel sheet

9:39 am

Forwarded



EOL Letter = Prietest Touch - AT0540315RBK.pdf

1 page • 89 kB • PDF

9:39 am

27 June 2024



Message





# SEMI AUTO ANALYSER

S No	Date	Received from / Issued to	Receipt	Issued	Balance	Main Store Balance
1.	26/12/16	B/F from B, Page No: 103	1		1	1
1	27/12/16	Issued for use		1	Nil	1
2.	9/4/24	Received from store	1		1	2
	9/4/24	Issued to use		1	Nil	2



*[Signature]*  
 മെമ്പർമാർക്കുവേണ്ടി  
 കൂടുതൽ വിവരങ്ങൾക്ക്  
 ടെലിഫോൺ - 678534



