



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 218059

519

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHC</u></p> <p>Address <u>Thonnakkal</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>9847337294</u></p>	<p>Call Registration Date : <u>22/05/2024</u></p> <p>Caller ID : <u>112649</u></p> <p>Date of Visit : <u>23/05/24</u></p> <p>Asset No. : <u>0145071</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>Sanso</u> Model : <u>NA</u></p> <p>S. No. <u>NA</u> Dept. <u>FWC</u></p>
--	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT WORKING -
wrong weight reading, display issue

Action Taken : Checked and found load cells and board
defective - need to replace these spares for further
checking and working condition of the equipment

Completed Date : 23/05/24 Time : 11am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Deepak</u>	<u>23/05/24</u>	<u>10:30</u>	<u>11am</u>

Customer Remark: Completed Pending
Medical Officer in charge
Family Health Centre
Thonnakkal
Kudavoor P.O. 695313



<p>Service Engineer Name : <u>Deepak C.M</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>23/05/24</u></p> <p>Contact Number : <u>7902881038</u></p>	<p>Customer Name : <u>Sobha. TS</u></p> <p>Signature : <u>Sobha</u></p> <p>Date : <u>[Blank]</u></p> <p>Contact Number : <u>9847337294</u></p> <p>Designation : <u>Jr PAM.</u></p> <p>Hospital Seal : <u>[Blank]</u></p>
--	--


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

519

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	PHC THONNAKKAL
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Weighing machine Make : Samsco Model : NA Serial No : NA
4	Equipment ID/ Barcode	0145071/112649
5	Date of purchase/ Year of manufacture/Installation Date	13/08/2019
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22/05/2024
9	Action taken	Checked and found Load cells and board defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found load cells and board defective. The apparatus is installed on 13/08/2019 and covered upto 4+years. Quotation not attached Since spares are not available in the market, so we recommend the unit for condemnation
17	Name & Signature of CYRIX Authority	DEEPAK EM 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Board defective. Since load cells and board is defective, it is not recommended to repair.


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

 22-6-2024

Signature of JC BM (NHM)



Date:

 22/6/24.

Medical Officer in charge
Family Health Centre
Thonnakkal
Kudavoor P.O. 695313

STOCK REGISTER

Health Service Departments

KARUNYA
MEDICINE DEPOT.

Name of Article SUB CENTRE EQUIPMENT KIT.

13/8/2019.

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
13/8/2019.		Brought forward			12/28		
	1.	BP Apparatus digital-			12/28		
	2.	weighing machine digital			12/28		
	3.	weighing machine adult			11/28		
	4.	Stadiometer			12/30		
	5.	Stadiometer			12/24		
	6.	Hemo meter test strips-			4/20		
	7.	Hemo meter			3/24		
	8.	Citrus steroid-			10/23		
	9.	Oil meter stop (oil plus)			12/20		
	10.	Oil meter (oil plus) -1			1/21		
	11.	Thermometer digital			1/28		
	12.	Salt testing kit for iodine			10/20		
	13.	Stethoscope (saxer deluxe)			12/28		
	14.	Hand cutter (manual syringe)			1/28		
	15.	Examination torch (LED)			1/28		
	16.	Measuring tape			12/24		
	17.	Focusscope (30mm)			4/27		
	18.	NEBULIZER			1/28		
	19.	KIT for testing residual glucose in drinking water			4/20		
	20.	LANCET 1'S (non-ally)			8/23		
	21.	DUST-BIN - BLACK 10 LTR			12/28		
	22.	DUST-BIN - RED			1/28		
	23.	DUST-BIN - YELLOW			12/28		
	24.	Urine strip (Glucose and Bilirubin)			12/20		
		Carried over					



Santhosh
28/6/24
Medical Officer in charge
Family Health Centre
Thonnakkal
Kudavoor P.O. 695313

