



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 218363

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

520

<p>Health Facility <u>THH</u></p> <p>Address <u>Vakkala</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>9633601331</u></p>	<p>Call Registration Date : <u>15-05-2024</u></p> <p>Caller ID : <u>111219</u></p> <p>Date of Visit : <u>17-05-2024</u></p> <p>Asset No. : <u>0151421</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Life care</u> Model : <u>CNB69022</u></p> <p>S. No. <u>LC 2219040011</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working


Action Taken : Checked and found the motor and piston were damaged need to change the part for further checking and working condition of the equipment.

Completed Date : 17-05-24 Time : 11:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer <u>Vishnu.K</u>	Date <u>17-05-24</u>	Start Time <u>10:15am</u>	End Time <u>11:30am</u>
-----------------------------------	-------------------------	------------------------------	----------------------------



Customer Remark Completed Pending

<p>Service Engineer Name : <u>Vishnu.K</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>17-05-24</u></p> <p>Contact Number : <u>9605884413</u></p>	<p>Customer Name : <u>Breebeechy . u</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>[Date]</u></p> <p>Contact Number : <u>1715124</u></p> <p>Designation : <u>[Designation]</u></p> <p>Hospital Seal : <u>9633601331</u></p>
--	--

TALUK HEADQUARTERS HOSPITAL
VARKALA

520

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH VARKALA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Nebulizer Make : Lifecare Model : CNB69022 Serial No : NA
4	Equipment ID/ Barcode	0151421/111219
5	Date of purchase/ Year of manufacture/Installation Date	11/5/2022
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	15/05/2024 (toll free)
9	Action taken	Checked and found Nebulizer motor defective and piston damage. Need to replace the spare for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1800/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found Nebulizer motor with piston defective. Quotation not submitted since the spares are not available in the market. As per the tender clause 5.3.14.2 recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston defective. So this cannot be repairable as per the tender conditions.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram

MS
18.06.2024

Signature of JC BM (NHM)





Date: *MS*





MS
Signature of
Superintendent/Medical Officer (i/c)
SUPERINTENDENT
TALUK HEADQUARTERS HOSPITAL
VARKALA

Nebulizer. (Electric)

	Rd	Condemned	Total	
(To in form old Register page 31)				
1/2/2010 Rd from store (LP) (Level 343)	1		1	
3/11/2012 Rd from stock (Constituted)	1		2	②
27/11/2012 Rd from stock	1		4	④
21/12/13 Rd from stock	1		5	⑤
26/2/2014 Rd from stock	1		5	⑤
27/10/2014 Rd from stock	1		5	⑤
Condemned				
As per order No. B1 - 2278 / 2015 / J.H.Q.H.V				
Dated 01/08/2015				
		3	2	⑤
			(1 cond report)	⑤
<u>2018</u>				
28/3/18 nebulizer	2		2	
9/10/18 transfer to unseviceable register page no: 1				①
12/12/18 Rd from store (page SR - 15)	2			②
11/5/18 Rd from store (page SR - 15)	1			③
23/7/23 Rd from Store (Stock register IEE page no 230)	2			




 SUPERINTENDENT
 TALUK HEADQUARTERS HOSPITAL
 VARKALA



Sew

SUPERINTENDENT
 TALUK HEADQUARTERS HOSPITAL
 VARKALA