



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223355

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>Phc</u></p> <p>Address : <u>Thonnakkal</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>9847337294</u></p>	<p>Call Registration Date : <u>22/05/24</u></p> <p>Caller ID : <u>112645</u></p> <p>Date of Visit : <u>23/05/24</u></p> <p>Asset No. : <u>0145063</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture : <u>Amul industries</u> Model : <u>Elkometer</u></p> <p>S. No. <u>130563</u> Dept. <u>FNC</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : NOT working - Mercury Spillage

Action Taken : Checked and found Mercury fully spilled out from the Mercury tank - identified Mercury tank and glass rod defective need to Replace these spares for further checking and working condition of the equipment

Completed Date : 23/05/24 Time : 10.30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Deepak. E-m</u>	<u>23/05/24</u>	<u>10am</u>	<u>10.30am</u>

Customer Remark Completed Pending



Service Engineer Name : Deepak

Signature : [Signature]

Date : 23/05/24

Contact Number : 7902881038

Customer Name : Sobha T.S

Signature : [Signature]

Date : 22/5/24

Contact Number : 9847337294

Designation : IPM

Hospital Seal :

Medical Officer In charge
 Family Health Centre
 Thonnakkal
 Kuttavoor P.O. 685313


517

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	PHC THONNAKKAL
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP Apparatus Make : Anita Industries Model : Elkometer Serial No : 430563
4	Equipment ID/ Barcode	0145063/112645
5	Date of purchase/ Year of manufacture/Installation Date	01/05/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22/05/2024
9	Action taken	Checked and found mercury fully spilled out from the Mercury tank. Identified Mercury tank and glass rod defective.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value.	Not available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found mercury spillage, identified mercury tank and glass rod defective. The unit is installed on 1/5/2021 and covered upto 3+ years. Since the mercury is not available in the market. Recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	DEEPAK E M 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Glass rod and Mercury tank defective. So this can be RBER under tender conditions.

MANEESHA MOHAN P.P.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


22-6-2024

Signature of JC BM (NHM)




21/6/24
Signature of
Sulcar
Officer in charge
Family Health Centre
Thonnadal
Kudavoor P.O. 695013



ELKOMETER

IS 13390



CM/L-0983369

**SPHYGMOMANO
METER 300**

MFD. BY :

ANITA INDUSTRIES

T-5/236, Mangol Puri Indl. Area,
Ph.-I, New Delhi-110083

S No. **430563**



CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP) 

TOLL FREE NO: 1800 - 425-7689

BAR CODE - (8004-890815225) 
0145063

ELKOMETER

