


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL
EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	GH THRISSUR
3	Name of Equipment with Make, Model and Serial Number	UPS Make - NUMERIC Model - ONFINITI SN -150100359
4	Equipment ID/ Barcode	0810312 #87310
5	Date of purchase/ Year of manufacture/Installation Date	09/12/2015
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown/ Date of registration of complaint	25/12/2023
9	Action taken	Checked the equipment and found that the control PCB,battery, transformer are defective.Enquired for spare
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	12-230 Inverter PCB – 1298/- 7 Ah battery - 1139/- 12v transformer -649/- TOTAL – 3086/-


13	Asset Value	4,500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report and quote attached
16	Reasons for recommending the equipment as BER	Checked the equipment and found that the control PCB, battery, transformer are defective, Equipment installed on 06/02/18 and aged up to 6 years, Spare cost is 68% of asset value and it is beyond economical repairable, We recommend this equipment for condemnation.
17	Signature of CYRIX Authority	Ramdas K 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spare cost is high.


Signature of JC BM (NHM)

Jr. Consultant (Biomedical)
NHM



Date _____


Signature of
Superintendent (Medical Officer) (I/c)

SUPERINTENDENT (I/C)
GENERAL HOSPITAL
THRISSUR - PIN: 680 001



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 177678

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>CAH</u>	Call Registration Date : <u>25/12/23</u>
Address : <u>Thiruvur Kerala</u>	Caller ID : <u>87310</u>
Ph : <u>8281006383</u>	Date of Visit : <u>27/12/23</u>
	Asset No. : <u>0810312</u>
	EQPT Name : <u>UPS</u>
	Manufacture <u>numeric</u> Model : <u>onfinibi 1kvAFM1</u>
	S. No. <u>15.01.00359</u> Dept. <u>Blood bank</u>

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : not working properly error showing

Action Taken : checked the equipment and found that the control PCB, battery, transformer are defective. enquired for spares locally.

Completed **Date :** 27/12/23 **Time :** 4:30 PM **Spare Required**


Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jinto</u>	<u>27/12/23</u>	<u>12:30pm</u>	<u>4:30pm</u>

Customer Remark **Completed** **Pending**

Service Engineer Name : <u>Jinto</u>	Customer Name : <u>Neenu</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>27/12/23</u>	Date : <u>27/12/23</u>
Contact Number : <u>893847125</u>	Contact Number : <u>[Number]</u>
	Designation : <u>NEENU SUPERINTENDING Lab Tech (Blood Bank)</u>
	Hospital Seal : <u>GENERAL HOSPITAL THRISSUR</u>



76 UPS

82

76

BC TSR - 76


Model : MUPS 1kVA
make : Numeric ON FINITI
Serial No : 150100859
Installed on : 9/12/2015
purchased by : From TWH chavakkad
Rs :

Entc Reg NO : 21 page 215 old no - 228

Barcode : 0810312.



[Handwritten signature]
SUPERINTENDENT (Dy. DHS)
GENERAL HOSPITAL
THRISSUR - PIN: 680 001



GENERAL HOSPITAL
THRISSUR PIN: 680 001

SUPERINTENDENT (Dy.DHS)
GENERAL HOSPITAL
THRISSUR - PIN: 680 001

AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 27/01/2024

Contact : 8921950217

Email : aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12 -230 inverter PCB	1	1100.00	1100.00	18	1,298.00
2	7Ah Better y	1	890	890	28	1,139.00
3	12v transformer	1	550	550	18	649.00
PAYABLE AMOUNT						3086.00
Commercial terms and conditions <ol style="list-style-type: none">1. Payment 80% advance along with PO balance before dispatch, 20% After installation.2. Delivery period; 2weeks.3. Tax :GST@18% for spares, 28%@Battery4. Installation; Our scope.						
Please arrange to release the PO at our above address						
Bank name :HDFC		Account no:50200082743176			IFSC Code:HDFC0004083	
We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.						
Thank you, Sincerely, BHAGATH MOHAN						





87310-0810312- gh thrissur - ups

2 messages

Bijo T Joy <bijocyrix@gmail.com>

Sat, 20 Jul, 2024 at 10:44 am

To: aionbackups12@gmail.com

Cc: Ramdas K <ramdascyrix@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

dear sir,

it is inform you that ups complaint reported at gh thrissur hospital. kindly attend and share the quotation

HOSPITAL - GH THRISSUR

equipment - ups

barcode - 0810312

caller id - 87310

contact no - 8921870010

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Bhagath Mohan <aionbackups12@gmail.com>

Sat, 20 Jul, 2024 at 10:52 am

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: Ramdas K <ramdascyrix@gmail.com>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Thanks & Regards

Bhagath Mohan

PH:-8921950217

[Quoted text hidden]