


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	PHC Vazhavatta
3	Name of Equipment with Make, Model and Serial Number	Equipment-Weighing machine MAKE: Smart care MODEL:NA SN: NA
4	Equipment ID/ Barcode	1240571 , TICKET 122132
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date - not available (As per the stock book installation date is 08/07/2021 stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	09/07/2024
9	Action taken	Checked the machine and found that load sensor and display board complaint
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	1600/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that load sensor and display board defective. As per the stock book Equipment installed on 08/07/2021 and aged up to 3 years. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.
17	Signature of CYRIX Authority	NIYAS HADI A P 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Load sensor and display board complaint
Recommended condemnation


20/7/24
CONSULTANT
Signature of Biomedical Engineer
National Health Mission
(Arogyakeralam)
Wayanad-673122


ബി.ഇ.എസ്
കുടുംബാരോഗ്യ കേന്ദ്രം
വയനാട്, വയനാട്
ഫോൺ: 673 122



Superintendent Medical Officer (i/c)

Signature of seal

Date



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 154143

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>PHC</u></p> <p>Address : <u>Vazhavatta</u> <u>Wayanad, Kerala.</u></p> <p>Ph : <u>9747 939648</u></p>	<p>Call Registration Date : <u>09/07/24</u></p> <p>Caller ID : <u>122132</u></p> <p>Date of Visit : <u>10/07/2024</u></p> <p>Asset No. : <u>1240571</u></p> <p>EQPT Name : <u>Weighing machine</u></p> <p>Manufacture <u>Smart care</u> Model : <u>Digital</u></p> <p>S. No. <u>N/A</u> Dept. <u>OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Board and load sensor error.

Action Taken : checked and found board and load sensor complaint.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ajun P</u>	<u>10/07/24</u>	<u>11:00AM</u>	<u>12:30AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Ajun P</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/07/24</u></p> <p>Contact Number : <u>759384 737</u></p>	<p>Customer Name : <u>[Signature]</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>10/7/24</u></p> <p>Contact Number : <u>9961858727</u></p> <p>Designation : <u>Nursing Officer C.I.</u></p> <p>Hospital Seal :</p>
--	--



[Signature]
 മെഡിക്കൽ ഓഫീസർ
 കുടുംബാരോഗ്യ കേന്ദ്രം
 മൂലമം, മൂലമം



ms.
ബഹിഷൽ ഓഫീസർ
കുടുംബാരോഗ്യ കേന്ദ്രം
വാഴവു, വയനാട്
ഫീൽ: 673 122

