


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Wayanad
2	Name of Hospital	UPHC Vengoor
3	Name of Equipment with Make, Model and Serial Number	Equipment-Nebuliser MAKE: Sargen MODEL:NA SN: NA
4	Equipment ID/ Barcode	1251156 , TICKET 120949
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date -24/03/2023 (As per the stock book installation date is 29/03/2023 stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	04/07/2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. It's not repairable need to replace the piston and motor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	1568
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective these parts are not repairable. As per the stock book Equipment installed on 29/03/2023 aged up to 1 years 3 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.
17	Signature of CYRIX Authority	NIYAS HADI A P 


*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and piston are defective. Can't repair or replace the parts. Recommended condemnation.


 Consultant
 Signature of JC BM NHM
 National Health Mission
 (Arogyakeralam)
 Wayanad-673122


 Superintendent/Medical Officer (i/c)


 MEDICAL OFFICER
 UPHC VENGOOR
 Signature of seal

Date 19/7/2024



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 221239

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.ki@cyrix.in

Health Facility UPHC	Call Registration Date : 04/07/2024
Address Vengoor	Caller ID : 120949
Wayanad, Kerala.	Date of Visit : 04/07/2024
Ph : 8086010834	Asset No. : 1251156
	EQPT Name : Nebulizer
	Manufacture Sargen Model : N/A
	S. No. N/A Dept. OP

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Unwanted sound from the
..... nebulizer.

Action Taken : Checked the nebulizer and found
..... compressor assembly totally damaged.

Completed Date : Time : Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
Ajun. P	04/07/24	9:30AM	10:35 AM


Customer Remark Completed

Pending

MEDICAL OFFICER
 UPHC VENGLOOR
 Date:
 673592

Service Engineer Name : Ajun. P
Signature :
Date : 04/07/2024
Contact Number : 7593847137

Customer Name : Mithu. Kudiakose
Signature :
Date : 04/07/24
Contact Number : 9539171772
Designation : Nursing Officer
Hospital Seal :



Jan

MEDICAL OFFICER
UPHC VENGOOR



