


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOZHIKODE
2	Name of Hospital	THQH KOYILANDY
3	Name of Equipment with Make, Model and Serial Number	Equipment : BP Apparatus Make: ELKO Model: EL520 SN : NA
4	Equipment ID & Barcode	118410 & 1121560
5	Date of purchase / Year of manufacture /Installation Date	Ticket master installation date -11-06-2020 (As per stock book installation date is 14-02-2021, stock book attached for reference)
6	Warranty details (Yes/No)	NO Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	NO AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	21-06-2024
9	Action taken	Checked and found Main board and display defective. Need to replace these spares for further checking of working condition of the equipment
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found that main board and display defective. In Ticket master installation date is 11-06-2020, As per the stock book Machine installed on 14 -02-2021 aged up to 03 years 4 months. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation
17	Name & Signature of CYRIX Authority	AKHILRAJ MP 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that main board & display damaged. Machine is installed on 14/2/2021. Spares are not available in the market so recommended for condemnation.



Signature of JC BM (NHM)

Consultant Bio Medical Engineer
National Health Mission
Kozhikode

Date



Seal 673305

Signature of 

Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 144026

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PHH Koyilandy</u></p> <p>Address <u>Koyilandy</u></p> <p style="padding-left: 100px;"><u>Kozhikode</u></p> <p>Ph : <u>9495308811</u></p>	<p>Call Registration Date : <u>24/6/24</u></p> <p>Caller ID : <u>118410</u></p> <p>Date of Visit : <u>24/6/24</u></p> <p>Asset No. : <u>1121560</u></p> <p>EQPT Name : <u>Bp Apparatus</u></p> <p>Manufacture <u>Elko</u> Model : <u>Digital ECG</u></p> <p>S. No. <u>Nil</u> Dept. <u>Female ward</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

not getting on

Action Taken : checked & found that main board & display defective. need to replace these spares for further checking of the working condition at the equipment.

Completed Date : 24/6/24 Time : 12:45pm Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.	←—————→		
3.			


Cyrix Engineer	Date	Start Time	End Time
<u>Akhilraj</u>	<u>24/6/24</u>	<u>12:15pm</u>	<u>12:45pm</u>

Customer Remark



Completed Pending

<p>Service Engineer Name : <u>Akhilraj</u></p> <p>Signature : </p> <p>Date : <u>24/6/24</u></p> <p>Contact Number : <u>8943677676</u></p>	<div style="text-align: center;">  </div> <p>Customer Name : <u>Dajishma</u></p> <p>Signature : </p> <p>Date : <u>24/6/24</u></p> <p>Contact Number : <u>9846914294</u></p> <p>Designation : <u>N/o</u></p> <p>Hospital Seal : </p>
---	---

SUPERINTENDENT
 Head Quarters Hospital
 Koyilandy



Superlux
Superlux Head Quarters Hospital
Koyilandy



Superlux
Superlux Head Quarters Hospital
Koyilandy

