


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

| Sl. No | Particulars | Details |
|--------|---|---|
| 1 | Name of District | KOZHIKODE |
| 2 | Name of Hospital | THQH KOYILANDY |
| 3 | Name of Equipment with Make, Model and Serial Number | Equipment: BP Apparatus Make: ELKO Model: EL 520 SN : NA |
| 4 | Equipment ID & Barcode | 118483 & 1121559 |
| 5 | Date of purchase / Year of manufacture /Installation Date | Ticket master installation date- 11-06-2020 (As per stock book installation date is 14-02-2021, stock book attached for reference) |
| 6 | Warranty details (Yes/No) | NO Warranty |
| 7 | *AMC/ CAMC Period agreed at the time of purchase | NO AMC/CAMC |
| 8 | Date of breakdown(Date of registration of complaint through email/ Toll free) | 21-06-2024 |
| 9 | Action taken | Checked and found Main board and display defective. Need to replace these spares for further checking of working condition of the equipment |
| 10 | Present status of the equipment (Fully damaged / partially damaged) | Fully damaged |
| 11 | Recommendations for repair (required service details) | Not recommending for repair |
| 12 | Cost of spares (specify parts and cost) | NA |

| | | |
|----|--|---|
| 13 | Asset Value | 2200/- |
| 14 | # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value | NA |
| 15 | Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) | Cyrix service report attached |
| 16 | Reasons for recommending the equipment as BER | Checked and found that main board and display defective. In software installation date is 11-06-2020, As per the stock book Bp Apparatus installed on 14-02-2021 aged up to 3 years 4 months. Quotation not submitted since spares are not available in the market So we recommend the equipment for condemnation |
| 17 | Name & Signature of CYRIX Authority | AKHILRAJ MP  |

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked & found that main board and display damaged. Machine installed on 14/02/21. Spare is not available in the market. So recommended for condemnation

Signature of JCBM (NHM)

Consultant Bio Medical Engineer
National Health Mission
Kozhikode



Signature of


SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy

Date

Seal

Superintendent / Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 144025

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

| | |
|--|---|
| <p>Health Facility <u>THQH Koyilandy</u></p> <p>Address <u>Koyilandy</u></p> <p style="padding-left: 40px;"><u>Kozhikode</u></p> <p>Ph : <u>9495308811</u></p> | <p>Call Registration Date : <u>21/6/24</u></p> <p>Caller ID : <u>118423</u></p> <p>Date of Visit : <u>24/6/24</u></p> <p>Asset No. : <u>1121559</u></p> <p>EQPT Name : <u>Bp Apparatus</u></p> <p>Manufacture <u>ELKO</u> Model : <u>Digital EL-520</u></p> <p>S. No. <u>nil</u> Dept. <u>Female ward</u></p> |
|--|---|

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :

not getting on

Action Taken : checked & found that main board & display defective. Need to replace these spares for further checking of working condition at the equipment.

Completed Date : 24/6/24 Time : 12:15 PM Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|------------------|----------------|-----------------|-----------------|
| <u>Akhil Raj</u> | <u>24/6/24</u> | <u>11:30 AM</u> | <u>12:15 PM</u> |

Customer Remark Completed Pending


(For Radha.K.P)
SNO

| | |
|--|---|
| <p>Service Engineer Name : <u>Akhil Raj</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/6/24</u></p> <p>Contact Number : <u>8943677676</u></p> | <p>Customer Name : <u>Rajishma</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>24/6/24</u></p> <p>Contact Number : <u>9846915111</u></p> <p>Designation : <u>NP</u></p> <p>Hospital Seal : <u>Taluk Head Quarters Hospital Koyilandy</u></p> |
|--|---|




SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy

SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy



[Signature]
SUPERINTENDENT
 Taluk Head Quarters Hospital
 Koyilandy



[Signature]
 SUPERINTENDENT
 Head Quarters
 Koyilandy

