


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC TRIPPANGOTTOOR
3	Name of Equipment with Make, Model and Serial Number	WEIGHING MACHINE Make : SAMSO Model: SKINNY SN:N/A
4	Equipment ID/ Barcode	1344587 #121057
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date- 03-06-2019 (Actual Installation Date is 15.05.2020 Stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 04-07-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found mainboard, Load cells damaged. Need to replace these spares for further checking .
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	715/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 03-06-2019. As per the stock book Installation date is 15.05.2020 aged up to 4 years 1 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that machine not starting ON. due to main board & load cells are damaged. Aged up to 4 years & above. So recommended for BER as per BER Protocol.



Signature of JC BM (NHM)

15/07/2024
CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 670002

Aswathi. A.k

~~Aswathi~~
Neg offices



Date: 16.07.24


Signature of 
Superintendent/Medical Officer (i/c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 216840

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>PNC Trippangattor</u></p> <p>Address <u>Trippangattor</u> <u>Kannur</u></p> <p>Ph : <u>8078967981</u></p>	<p>Call Registration Date : <u>4-7-24</u></p> <p>Caller ID : <u>121057</u></p> <p>Date of Visit : <u>5-7-24</u></p> <p>Asset No. : <u>1344587</u></p> <p>EQPT Name : <u>weighing machine</u></p> <p>Manufacture <u>Samsa</u> Model : <u>skinny</u></p> <p>S. No. <u>NA</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching on.

Action Taken : checked the machine and found that machine not switching on. found mainboard and load cells are damaged. Need to replace these spares for further checking.

Completed Date : 5-7-24 Time : 11:00Am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasilraj</u>	<u>5-7-24</u>	<u>10:00Am</u>	<u>11:00Am</u>


Customer Remark Completed Pending

Service Engineer Name : Rasilraj

Signature : Rasilraj

Date : 5-7-24

Contact Number : 9995521436



Customer Name : Aswath A.K

Signature : Aswath A.K

Date : 05.07.24

Contact Number : 9562123033

Designation : Neg officer

Hospital Seal :

ഹോസ്പിറ്റൽ നമ്പർ : 0490-2392350


NAME OF ITEM: WEIGHING MACHINE .


1.1

DATE	FROM WHOME RECEIVED/ TO WHOME ISSUED	RECEIPT	ISSUE	BALANCE	PHY. STOCK	COAST SIGN.
07/08/2023	Copied from SN Substock register ; E issue date 19/5/2020 Issued on 19/5/2020	02	02	02

[Signature]

കേരള ശാസ്ത്ര സാങ്കേതിക വിദ്യാലയം
കോളേജ്, കോട്ടയം
ഫോൺ: 0490-2392350





[Handwritten signature]
ഡെപ്യൂട്ടി ഓഫീസർ
പ്രാഥമിക ആരോഗ്യകേന്ദ്രം
തൃപ്രയാർ, കണ്യാർ റീഡ്
ഫോൺ: 0490 - 2392350

