

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**


**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	UDC URUVACHAL
3	Name of Equipment with Make, Model and Serial Number	DIALYZER REPROCESSING UNIT Make : TUSCAN Model: DCS PLUS SN:NA
4	Equipment ID/ Barcode	1340218 #110503
5	Date of purchase/ Year of manufacture/Installation Date	05-01-2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 10.05.2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found mainboard, smps board, Display Board, Driver board, Solenoid valves, Peristaltic pump are defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	

**MEDICAL OFFICER I/c  
Urban Dialysis Centre  
Uruvachal, Mattanur**




13	Asset Value	2,62,500/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found mainboard, smps board, Display Board, Driver board, Solenoid valves, Load cells, Peristaltic pump are defective. Equipment installed on 05.01.2018 aged up to 6 years 4 months. The OEM cura healthcare has closed its operations and completely closed down. There are no available spares. Consider the circumstances and in assurance to the tender clause 5.3.14.2 which stipulates that manufacture not existing the equipment should be evaluated for condemnation. So we recommending the equipment for condemnation
17	Name & Signature of CYRIX Authority	AVINASH T 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:  
 checked & identified that machine not switching on, main board, SMPS, Display board and other major part as defective. Aged up to 6 years above. The OEM cura not in this field now. so spare not available. recommended for BER as per BER Protocol  
  
 15/05/2024  
 Signature of JC BM (NHM)  
 KANNUR - 670002

In BEMP quarterly review on 28/6/24 decided to physical verify the equipment by Mattanur MO. That procedure still pending. So acknowledge this remark as BER consideration report.



  
 15/07/2024  
 CONSULTANT BIOMEDICAL  
 NATIONAL HEALTH MISSION  
 KANNUR - 670002

Date:

Signature of  
 Superintendent/Medical Officer (i/c)

MEDICAL OFFICER i/c  
 Urban Dialysis Centre  
 Uruvachal, Mattanur



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

No. : 215044

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : VDC unuvachal	Call Registration Date : 10-5-24
Address : Kollam	Caller ID : 110503
Kerala	Date of Visit : 11-5-24
Ph : 9496191350	Asset No. : 1340218
	EQPT Name : Dialyzer Reprocessing unit
	Manufacture : Tuscan Model : DCR plus
	S. No. : NA Dept. : Dialysis unit

Service Classification : Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :** Main board, smps board, Display board, Driver-board, solenoid valves, peristaltic pump are defective

**Action Taken :** Checked the machine and found that machine is not switching on. Found main board, smps board, display board, driver board, solenoid valves, peristaltic pump are defective

Completed  Date : 11-5-24 Time : 11:45 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
Shankrishnan	11-5-24	10:15 AM	11:45 AM

Customer Remark Completed  Pending

Service Engineer Name : Shankrishnan  
Signature :   
Date : 11-5-24  
Contact Number : 820188514





Customer Name : Theerth  
Signature :   
Date : 11-5-24  
Contact Number : 9506  
Designation : Dialysis  
Hospital/Seat :

MEDICAL OFFICER I/C  
Urban Dialysis Centre  
Unuvachal, Mattanur



**MATTANUR MUNICIPALITY  
URBAN DIALYSIS CENTRE, URUVACHAL  
STOCK BOOK OF MEDICINES AND INSTRUMENTS**

Name of item... REPROCESSING MACHINE

Date	From whom received or to whom issued	Receipt	Issue	Balance	Batch Number	Expiry Date	Price	Remarks
	CURA HEALTHCARE PRIVATE LIMITED							
	M. R. Salai, Nagkeri							
	Chennai - 600044							
	Model : DCS Plus							
	Installation date :- 05/01/2018							
	Amount :- 2,94,000/-							
	Recd code no :- 70133067							
								
								
								
								
	MEDICAL OFFICER I/c Urban Dialysis Centre Uruvachal, Mattanur							

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
HEALTHCARE PRIVATE

TOLL FREE NO: 1800 - 425-7000

BAR CODE - (8004-890615225)  
1340218



MEDICAL OFFICER I/c  
Urban Dialysis Centre  
Uruvachal, Mattannur