


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	PHC CHERUKUNNU(P)
3	Name of Equipment with Make, Model and Serial Number	BP APPARATUS Make: ELKO Model: EL 510 SN: NA
4	Equipment ID/ Barcode	1343110 #120092
5	Date of purchase/ Year of manufacture/Installation Date	27-08-2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 29-06-2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	2200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and pump Defective. Need to replace these spare for further checking. Equipment installed on 27-08-2019 aged up to 04 years 10 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that main board, Pressure sensor & Pump are defective. Aged upto 4 years & above. So service cost not feasible. - recommended for BER as per BER protocol.


12/07/2024


Signature of JC BM (NHM)
CONSULTANT BIOMEDICAL
NATIONAL HEALTH MISSION
KANNUR - 073301

Reeta. P. V

Reeta J P V

Date: 22/7/24 .




Dr. JESHNA P
for Medical Officer.
Signature of
Superintendent/Medical Officer (i/c)
PRIMARY HEALTH CENTRE
CHERUVU PUNNACHERY
KANNUR - 073301



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 227513

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERS Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bamp.k@cyrix.in

<p>Health Facility <u>PHC CHERUKUNNU P</u></p> <p>Address <u>DUNNASSERY</u> <u>KANNUR</u></p> <p>Ph : <u>9947461623</u></p>	<p>Call Registration Date : <u>29/6/24</u></p> <p>Caller ID : <u>120092</u></p> <p>Date of Visit : <u>30/6/24</u></p> <p>Asset No. : <u>1343110</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>ELKO</u> Model : <u>EL510</u></p> <p>S. No. <u>na</u> Dept. <u>Sub centre.</u></p>
---	---

Service Classification : Breakdown Call **PMS** **Calibration** **Cust.Training**

Problem Identified : Not switching on.

Action Taken : checked the equipment found equipment is not switching on. found that main board, pressure sensor and pump defective. Need to replace these spares for further checking.

Completed **Date :** 30/6/24 **Time :** 12:50 pm **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
<u>Aswis</u>	<u>30/6/24</u>	<u>12:00 pm</u>	<u>12:50 pm</u>

Customer Remark **Completed** **Pending**

Service Engineer Name : Aswis

Signature :

Date : 30/6/24

Contact Number : 8301012759



Customer Name : Reetha N Dr. JETHNA P

Signature :

Date : 30/6/24

Contact Number : 9774480000

Designation : ASSISTANT SURGEON

Hospital Seal : J & P H N PRIMARY HEALTH CENTRE CHERUKUNNU (PUNNACHEI) KANNUR - 870601

B.P. APPARATUS


21.05.17	B/E P.No. 59 Vol II	2 Nos damaged 1 Net in use	3 Nos
22.05.17	Transferred to unseizable Reg. P.No. 5	2 Nos	1 Net
21/8/2019	Received from store Make: ELKO	1 Nos	

~~J. John~~

Dr. JESHNA P



ASSISTANT SURGEON
PRIMARY HEALTH CENTRE
CHERUKUNNU (PUNNACHERY)
KANNUR - 670301



Jeshna
D. JESHNA.P

ASSISTANT SURGEON
PRIMARY HEALTH CENTRE
CHERUKUNNU (PUNNACHERY)
KANNUR - 670301