


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	PHC PADIYUR
3	Name of Equipment with Make, Model and Serial Number	OXYGEN CONCENTRATOR Make: Medoxy Model: jay -5 SN: MZJS11689
4	Equipment ID/ Barcode	0844162 # 116649
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date – NA (stock details available – 01-01-2019 stock book attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 12-06-2024
9	Action taken	Checked the machine and found that compressor motor, sieve bed, pressure regulator and solenoid valve defective, need to replace these spares for further checking
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Compressor - 16000/- Sieve be - 14000/- Pressure regulator - 3000/- Solenoid valve - 7000/- IGST - 7200/- Total - 47200/-


13	Asset Value	33400/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	141.31%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report Attached and oem quotation attached
16	Reasons for recommending the equipment as BER	Upon inspecting the machine, it was discovered that the compressor motor, sieve bed, pressure regulator and solenoid valve are faulty. As a Stock book Installation date is 01-01-2019 aged up to 5 years and 6 months. service charge is 141.31% of asset value. Both criteria for RBER met. As per tender clause 5.3.14.1 we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Spall cost is very high


Sr. Consultant (Biomedical)
NHM, Thrissur

Mess
Signature of JC BM (NHM)



Date:


Signature of
Superintendent Medical Officer (i/c)

Medical Officer in Charge
FHC, 1st floor
Thrissur DL - 600 088



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 228045

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC padiyun</u>	Call Registration Date : <u>12/6/24</u>
Address <u>Thriassur</u>	Caller ID : <u>116649</u>
Ph :	Date of Visit : <u>13/6/24</u>
	Asset No. : <u>0844162</u>
	EQPT Name : <u>O₂ concentration</u>
	Manufacture <u>Medoxy</u> Model : <u>Jay-5</u>
	S. No. <u>MSJS11689</u> Dept. <u>OP</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Equipment turning off automatically

Action Taken : Checked the equipment and found that compressor, sieve bed, pressure regulator, and solenoid valve are defective. Need to replace these spares for further checking.

Completed Date : 13/6/24 Time : Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Randdas</u>	<u>13/6/24</u>		

Customer Remark Completed Pending

Service Engineer Name : <u>Randdas</u>	Customer Name : <u>Jincy FRANCES</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>13/6/24</u>	Date : <u>13/06/2024</u>
Contact Number : <u>9745520163</u>	Contact Number : <u>98956033</u>
	Designation : <u>MLSP</u>
	Hospital Seal : <u>[Seal]</u>



Medical Officer in Charge
Thriassur Dt. - 680 008



MODITECH INTERNATIONAL

ADMIN OFFICE : PLOT NO. 1045-B, 1ST FLOOR, JLPL INDUSTRIAL AREA,
SECTOR-82, MOHALI, SAS NAGAR, PUNJAB-140306
CONTACT NOS.: 92164-64081, 92161-74081
E-MAIL : moditech1@gmail.com / info@medoxy.com

GST No. 03ABFFM2524D1ZW

Ref: 023/2024-25

Date 27/06/2024

To,
CYRIX HEATHCARE PRIVATE LIMITED,
30/641 B, Petta Junction, Poonithura
Kochi, Kerala

Subject- Quotation

Dear Sir/Madam,

We are pleased to quote below our Lowest Price for the following for your kind consideration

S.NO	Nomenclature of the item	HSN	Qty	Unit Price	Amount
1.	COMPRESSOR	90330000	01	Rs. 16,000/-	Rs. 16,000/-
2.	SIEVE BED	90330000	01	Rs. 14,000/-	Rs. 14,000/-
3.	PRESSURE REGULATOR	90330000	01	Rs. 3,000/-	Rs. 3,000/-
4.	SOLENOID VALVE	90330000	01	Rs. 7,000/-	Rs. 7,000/-
				SUB TOTAL	Rs. 40,000/-
				IGST @18%	Rs. 7,200/-
				TOTAL	Rs. 47,200/-

TERMS AND CONDITIONS

1. Rates are FOR destination.
2. Delivery within 10 days.
3. Payment 100% advance against order.

4. Bank Details: Karnataka Bank Ltd
Acc No.- 1762000100017101

Branch: Sector 38 D Chandigarh
IFSC: KARB0000176

Thanking you and assuring you of our Best quality products at all times,

Yours truly,

For MODITECH INTERNATIONAL

For MODITECH INTERNATIONAL

Prithi Aggarwal

Auth Signatory

Authorized Signatory

spare quotation - phc padiyur - 08441689 - 116649

Bijo T Joy <bijocyrix@gmail.com>
To: medoxy@india.com, info@medoxy.com <info@medoxy.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Wed, 19 Jun, 2024 at 9:10 pm

Dear sir,
kindly provide the quotation for oxygen concentrator spares.
quotation for - compressor,sieve bed,pressure regulator and solenoid valve
equipment - oxygen concentrator
make - medoxy
model - jay -5
sr no - MZJS11689
hospital name - phc padiyur
barcode - 0844162
ticket id - 116649

With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Mail Delivery Subsystem <mailer-daemon@googlemail.com>
To: bijocyrix@gmail.com

Wed, 19 Jun, 2024 at 9:10 pm



Message blocked

Your message to **medoxy@india.com** has been blocked. See technical details below for more information.

The response from the remote server was:

550 5.4.1 Recipient address rejected: Access denied. [HK2PEPF00006FB5.apcprd02.prod.outlook.com 2024-06-19T15:40:56.932Z 08DC8C4F7068ABDA]

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>
To: medoxy@india.com, info@medoxy.com <info@medoxy.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Tue, 25 Jun, 2024 at 3:30 pm

reminder

With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

Mail Delivery Subsystem <mailer-daemon@googlemail.com>
To: bijocyrix@gmail.com

Tue, 25 Jun, 2024 at 3:31 pm



Message blocked

Your message to **medoxy@india.com** has been blocked. See technical details below for more information.

The response from the remote server was:

550 5.4.1 Recipient address rejected: Access denied. [HK2PEPF00006FAE.apcprd02.prod.outlook.com 2024-06-25T10:01:14.867Z 08DC8C646233AECA]

----- Forwarded message -----

From: Bijo T Joy <bijocyrix@gmail.com>
To: medoxy@india.com, "info@medoxy.com" <info@medoxy.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>
Bcc:
Date: Tue, 25 Jun 2024 15:30:58 +0530
Subject: Re: spare quotation - phc padiyur - 08441689 - 116649
----- Message truncated -----

Bijo T Joy <bijocyrix@gmail.com>
Draft

Thu, 27 Jun, 2024 at 12:26 pm

reminder
With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

On Tue, 25 Jun 2024 at 15:31, Mail Delivery Subsystem <mailer-daemon@googlemail.com> wrote:



Message blocked

Your message to **medoxy@india.com** has been blocked. See technical details below for more information.

The response from the remote server was:

550 5.4.1 Recipient address rejected: Access denied. [25T10:01:14.867Z 08DC8C646233AECA]

2024-06-

----- Forwarded message -----

From: Bijo T Joy <[redacted]>
To: "Medoxy India Pvt. Ltd." <info@medoxy.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>
Bcc:
Date: Tue, 25 Jun 2024 15:30:58 +0530
Subject: Re: spare quotation - phc padiyur - 08441689 - 116649
----- Message truncated -----

Bijo T Joy <bijocyrix@gmail.com>
To: medoxy@india.com, info@medoxy.com <info@medoxy.com>
Cc: Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 27 Jun, 2024 at 12:26 pm

reminder

With Regards,
BIJO T JOY
DISTRICT INCHARGE THRISSUR,
CYRIX HEATHCARE PRIVATE LIMITED
30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia
M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

On Wed, 19 Jun 2024 at 21:10, Bijo T Joy <bijocyrix@gmail.com> wrote:
[Quoted text hidden]

Mail Delivery Subsystem <mailer-daemon@googlemail.com>
To: bijocyrix@gmail.com

Thu, 27 Jun, 2024 at 12:27 pm



Message blocked

Your message to **medoxy@india.com** has been blocked. See technical details below for more information.

The response from the remote server was:

550 5.4.1 Recipient address rejected: Access denied. [SG1PEPF000082E4.apcprd02.prod.outlook.com 2024-06-27T06:57:03.622Z 08DC89A84528AAF0]

[Quoted text hidden]

Moditech International <info@medoxy.com>
To: Bijo T Joy <bijocyrix@gmail.com>
Cc: medoxy@india.com, Sarang KM <sarangcyrix@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>

Thu, 27 Jun, 2024 at 5:17 pm

Dear Sir/Madam,
Kindly find the attached quotation.

Thanks & Regards,

Invoice No. 4353418-49
 Dated:- 19.12.2018

INVOICE

GST NO. 01ABFFM2524DZL
 PAN NO. ABFFM2524D

MODITECH INTERNATIONAL 1/1/19

Regd. Off. Metro Plaza, SCO-54-55, Sector-9-D, Chandigarh (UT)160017

Phones : 9216474081, 9216464081 and 9216174081 Email ID- moditech1@gmail.com

To,
 AMERICARES INDIA FOUNDATION
 SEAGULL HOUSE, 1ST FLOOR, P.NO.-6
 SHIVJI COLONY, OFF ANDHERI KULA ROAD, CHAKALA,
 ANDHERI EAST, MUMBAI 400099, INDIA
 GSTIN NO.- 27AABTA6931D120,

CONSIGNEE/ DELIEVERY ADDRESS,
 To,
 GOVERNMENT PRIMARY HEALTHCARE CENTRE PADIYOOR,
 VALAVANANGADI-EDATHIRINJI RD, PADIYUR, THIRUSSAR
 KERALA 680688

PO No. PO/F.Y 18-19/6225

DATED 26.11.18

S.No.	DESCRIPTION OF GOODS	HSN CODE	QTY	UNIT PRICE	AMOUNT Rs.
1	Oxygen Concentrator	90192010	01	26700	26,700 -
	Model: JAY 5 Make: Longfan Scitech. Co. Ltd.				
Electronic Reference No. _____ Date _____				Total	26,700 -
Amount of Tax Subject to Reversal				FREIGHT CHARGES	-
				PACKING CHARGES	-
CGST	UTGST	IGST	SUB TOTAL		26,700 -
			CGST @		-
			UTGST @		-
Rupees in Words <u>Twenty Nine Thousand Nine</u> <u>Hundred Four</u> only.				IGST @ 12%	3,204 -
				Invoice Value	29,904 -

BANK DETAIL:
 PUNJAB NATIONAL BANK
 SCO-44-45, SECTOR-9/D, CHANDIGARH (UT) 160017
 C/C ACCOUNT NO.: 1488008700001412
 IFSC CODE: PUNB0148800

*Shade Ra.
 Equipments
 Chassis related
 P.N. 17 -*






For MODITECH INTERNATIONAL

For MODITECH INTERNATIONAL

Auth. Signatory

Medical Officer in Charge
 FHC Padiyoor
 Thirussar Dt.-680 688

Medical Oxygen Concentrator 

VOLTAGE AC 230V ± 10% POWER 550 VA
FREQUENCY 50 ± 1HZ
OXYGEN PURITY ≥ 90% SOUND LEVEL ≤ 50 dB
FLOW RATE 0-5 L/min
OUTLET PRESSURE: 0.04-0.07Mpa  
MODEL JAY-5
SN

MZJ5S 111689

