


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THRISSUR
2	Name of Hospital	CHC PUTHENCHIRA
3	Name of Equipment with Make, Model and Serial Number	Nebuliser MAKE: Easycare MODEL: EC-7011 SN: NA
4	Equipment ID/ Barcode	0830757 # 118732
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date- 05-10-2018 (As per Stock book installation date - 04-10-2012, stock book attached for reference)
6	Warranty details	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown/ Date of registration of complaint	22-06-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. need to replace the piston and motor for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Details of service required (Recommendations for repair)	Not recommending
12	Cost of spares (specify parts and cost)	NA



13	Asset Value	1,353/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX.	Cyrix Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. In ticket master installation is 05.10.2018 As per stock book installation date is 04-10-2012 aged up to 11 years 6 month. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation
17	Signature of CYRIX Authority	BIJO T JOY 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

No spare available



 Signature of JC BM (NHM)



 Superintendent/Medical Officer (6010) കേന്ദ്രം, പുത്തൻചിറ
 Puthenchira Health Centre
 Pin- 680 682, Puthenchira@emil.com


 Date

Signature of seal



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 228043

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Patta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>CHC Puthenchira</u>	Call Registration Date : <u>22/6/24</u>
Address <u>Thissur</u>	Caller ID : <u>118732</u>
Ph : <u>9446925570</u>	Date of Visit : <u>24/6/24</u>
	Asset No. : <u>0830757</u>
	EQPT Name : <u>Nebuliser</u>
	Manufacture <u>Gaycan</u> Model : <u>GC-7011</u>
	S. No. <u>NA</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : No fumes

Action Taken : Checked the equipment and found that Motor, Piston are defective, need to replace the piston, Motor for further checking

Completed Date : 24/6/24 Time : Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Randax</u>	<u>24/6/24</u>	<u>10AM</u>	<u>12:30 PM</u>

Customer Remark Completed Pending

Service Engineer Name : <u>Randax</u>	Customer Name : <u>SARITHA - A</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>24/6/24</u>	Date : <u>24.6.24</u>
Contact Number : <u>9745520163</u>	Contact Number : <u>9446925570</u>
	Designation : <u>SNO</u>
	Hospital Seal : <u>[Seal]</u>



Sl No	Date of Repair	Remarks	Segn	Date of Compl.	Name of Per Code	Date of Repair	Remarks	Segn
-------	----------------	---------	------	----------------	------------------	----------------	---------	------

JANUARY

02	20/1/24	coil connected replaced	P2	21/6/24	NEELADRI CSLOCK (Book 4/10/2012)	18/1/24		
65	-				Not working	18/1/24		

754	24/1/24	Battery replaced	3		Battery Code (0830757)	18/1/24		
85		Battery changed	3	23/1/24		18/1/24		

operator ad.	10/2/24	complaint attended	HTN					
30795								

gild	14/3/24	coil replaced	D					
For preventive maintenance								

ulti para 30754	5/4/24	Bp cuff	4					
ulti para 830754		tube changed						

ulti para 830755	}	Pasehatri						
ulti para 830754		Bp cuff damaged						

May 2024

control chair		for repair done						

ulti para 0830754		Battery change not maintained						
-------------------	--	-------------------------------	--	--	--	--	--	--



Nov 2018

CYRIX
HEALTH SERVICES
C.N: CHC 2018
P.O: Arakkalpet
No: 0830287
Call: 91711723
Fax: 91711724
180-801-2007 Certified by PHC



EC-7011
COMPRESSOR NEBULIZER
AC 230V - 50Hz
OPERATING TIME: 30 min. ON - 30 min. OFF
CE
ME20
LOT 1610

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
BAR CODE - (8004-890615225) 0830757



